

**Massachusetts Institute of Technology
Woods Hole Oceanographic Institution**



**Joint Program
in Oceanography/
Applied Ocean Science
and Engineering**



DOCTORAL DISSERTATION

Laboratory Evaluation of Laser-Induced Breakdown
Spectroscopy (LIBS) as a New *in situ* Chemical Sensing
Technique for the Deep Ocean

by

Anna Pauline Miranda Michel

September 2007

20080122054

MIT/WHOI

2007-19

**Laboratory Evaluation of Laser-Induced Breakdown Spectroscopy (LIBS) as a
New *in situ* Chemical Sensing Technique for the Deep Ocean**

by

Anna Pauline Miranda Michel

Massachusetts Institute of Technology
Cambridge, Massachusetts 02139

and

Woods Hole Oceanographic Institution
Woods Hole, Massachusetts 02543

September 2007


DOCTORAL DISSERTATION

Funding was provided by the National Science Foundation under grants OCE0352278 and OCE0352242. Additional support was received from WHOI's Deep Ocean Exploration Institute, the WHOI Ocean Ventures Fund, and the Department of Defense for a National Defense Science and Engineering Graduate Fellowship.

Reproduction in whole or in part is permitted for any purpose of the United States Government. This thesis should be cited as: Anna Pauline Miranda Michel, 2007. Laboratory Evaluation of Laser-Induced Breakdown Spectroscopy (LIBS) as a New *in situ* Chemical Sensing Technique for the Deep Ocean. Ph.D. Thesis. MIT/WHOI, 2007-19.

Approved for publication; distribution unlimited.

Approved for Distribution:




James F. Lynch, Chair

Department of Applied Ocean Physics and Engineering



Paola Malanotte-Rizzoli
MIT Director of Joint Program



James A. Yoder
WHOI Dean of Graduate Studies

Laboratory evaluation of laser-induced breakdown spectroscopy (LIBS) as a new *in situ* chemical sensing technique for the deep ocean

by
Anna Pauline Miranda Michel
B.S. Chemical Engineering, B.S. Biology
Massachusetts Institute of Technology (1998)
M.S. Ocean Engineering
Massachusetts Institute of Technology (2002)

Submitted to the MIT/WHOI Joint Program in Oceanography/Applied Ocean
Science and Engineering

in partial fulfillment of the requirements for the degree of
Doctor of Philosophy

at the
MASSACHUSETTS INSTITUTE OF TECHNOLOGY
and the

WOODS HOLE OCEANOGRAPHIC INSTITUTION

September 2007

© Anna Pauline Miranda Michel, MMVII. All rights reserved.

The author hereby grants to MIT and WHOI permission to reproduce and to
distribute publicly paper and electronic copies of this thesis document in whole or in
part in any medium now known or hereafter created.

Author
MIT/WHOI Joint Program in Oceanography/Applied Ocean Science and
Engineering

Certified by
August 3, 2007
Alan D. Chave

Certified by
Thesis Supervisor

Alexandra H. Techet
Associate Professor, MIT
Thesis Supervisor

Accepted by
Henrik Schmidt

Chair, Joint Committee for Applied Ocean Science
and Engineering

Accepted by
Lallit Anand

Chairman, Committee on Graduate Students

**Laboratory evaluation of laser-induced breakdown
spectroscopy (LIBS) as a new *in situ* chemical sensing
technique for the deep ocean**

by

Anna Pauline Miranda Michel

Submitted to the MIT/WHOI Joint Program in Oceanography/Applied Ocean
Science and Engineering
on August 3, 2007, in partial fulfillment of the
requirements for the degree of
Doctor of Philosophy

Abstract

Present-day expeditionary oceanography is beginning to shift from a focus on short-term ship and submersible deployments to an ocean observatory mode where long-term temporally-focused studies are feasible. As a result, a critical need for *in situ* chemical sensors is evolving. New sensors take a significant amount of time to develop; thus, the evaluation of techniques in the laboratory for use in the ocean environment is becoming increasingly important. Laser-induced breakdown spectroscopy (LIBS) possesses many of the characteristics required for such *in situ* chemical sensing, and is a promising technique for field measurements in extreme environments. Although many LIBS researchers have focused their work on liquid jets or surfaces, little attention has been paid to bulk liquid analysis, and especially to the effect of oceanic pressures on LIBS signals. In this work, laboratory experiments validate the LIBS technique in a simulated deep ocean environment to pressures up to 2.76×10^7 Pa. A key focus of this work is the validation that select elements important for understanding hydrothermal vent fluid chemistry (Na, Ca, Mn, Mg, K, and Li) are detectable using LIBS. A data processing scheme that accurately deals with the extreme nature of laser-induced plasma formation was developed that allows for statistically accurate comparisons of spectra. The use of both single and double pulse LIBS for high pressure bulk aqueous solutions is explored and the system parameters needed for the detection of the key analytes are optimized. Using both single and double pulse LIBS, the limits of detection were found to be higher than expected as a result of the spectrometer used in this experimentation. However, the results of this validation show that LIBS possesses the characteristics to be a viable chemical sensing method for *in situ* analyte detection in high pressure environments like the deep ocean.

Thesis Supervisors:

Alan D. Chave, Senior Scientist, WHOI

Alexandra H. Techet, Associate Professor, MIT

Acknowledgments

I would like to first thank my advisor, Dr. Alan Chave (WHOI), for challenging me to this project. When I was looking for a PhD project and to enter the MIT-WHOI Joint Program, Alan had just been inspired by a former colleague to investigate an exciting technology that had the possibility of becoming a new oceanographic sensor. Although through the years we learned that an actual oceanographic sensor was further off than we had anticipated, Alan supported the laboratory work to show that this sensor will one day work in the ocean. In addition, his statistics knowledge has been invaluable to this work. Aside from this project, Alan also gave me the opportunity to experience sea-going oceanographic research by involving me in two cruises as part of the H20 project, encouraged me to travel to numerous conferences to present my work, and allowed me to be an independent researcher. I am very grateful to have had such diverse experiences as an oceanographic engineer.

I would like to thank the rest of my committee for their invaluable contributions, conversations, insight, knowledge, and patience. Professor S. Michael Angel allowed me to spend six months working in his lab in collaboration with his graduate students (Marion Lawrence-Snyder and Jon Scaffidi) during two research stints at the University of South Carolina - Columbia learning first about the LIBS technique and then gathering the data needed to publish my first paper on LIBS. In addition, when my spectrometer broke and had to be shipped to Germany for an extended time, the Angel lab willingly lent me their spectrometer which I am very grateful for. Professor Alexandra Techet (MIT) has been my sounding board for research and life too many times to count over the years. Alex has been an amazing role model and mentor. Norman Farr (WHOI) helped me to take a box full of optics and optomechanics, two lasers, a spectrometer, and a breadboard and turn the small black walled lab into a working LIBS lab. Norm's optics knowledge was invaluable and without his alignment skills, we probably would have burned even more lenses than we did. Meg Tivey (WHOI) was crucial for my knowledge of hydrothermal vent chemistry and for reminding me of the need for new sensors for vent systems. She was also the one that answered all of my vent chemistry questions. I would also like to thank Sheri White (WHOI) for chairing my defense and for sharing her lab space with me. Sheri kindly allowed me to use her sink and lab area so I would not have to mix solutions and wash glassware in the bathroom. Sheri also helped me on an almost daily basis with discussions about spectroscopy, replacement parts when things broke, the lending of tools, and of course our almost daily tea, coffee, and chocolate breaks.

There are numerous people at WHOI that helped me survive my PhD. Jeff Seewald was extremely helpful in designing the experiments for studying the matrix effect (Chapter 6). Hanu Singh kindly lent me a camera which I used for plasma imaging. Everyone at DSL helped make my experience at WHOI wonderful. I would like to thank Stace Beaulieu for sharing a bunk and standing Jason watch with me for two months, for all the chats, for drinking about 1000 cups of coffee with me, for eating about 300 bowls of tomato bisque with me, for all the swims and dives in the ocean, and for being an amazingly supportive friend. I would like to thank Karen Schwamb for being my “grad school mom”. Whenever I had a question or needed anything, Karen had the answer, found the answer, and got me what I needed in an amazingly efficient manner. Karen has made me laugh and cry and kept me amused and given me advice so many times. John Bailey, Alan Gardner, and Fritz Sonnichsen helped me numerous times when I needed tools, a soldering iron, a random part, or just help lifting heavy breadboards.

I would like to thank the Academic Programs office, Marsha Gomes, Julia Westwater, Laishona Vitelli, Jim Yoder, Jim Price, Judy McDowell, Valerie Caron, Christine Charette and John Farrington for all of their help and support personally, professionally, and financially.

I owe many thanks to the girls of the Joint Program (Cara, Jessica, Clare, Petra, Diane, Rachel, and Emily (and Carolyn - who we adopted as a JP student)) for keeping me sane with all the walks, the coffee, the lunches and dinners, the chats on the elliptical machines, the swims, the engagements, the gelato, the X-gals articles, the girls’ nights and the trips. As we go our separate ways, I will miss each of you. You are amazing friends and amazing scientists. I’d also like to thank all of the other wonderful friends that I made while at WHOI: Greg for being an awesome friend and for helping me in countless ways, including fixing my pump, even if he only wishes that he is an engineer; Brian for all the fun times we had both at WHOI and on the many trips; Matt for making my climb all the way to the top of Mt. St. Helens, and his wonderful wife Anna Skuladottir; and Lara who, although she moved away before the girls finished, has been a great cheerleader for all of us! Katy Croff has kept me amused since the day I started my masters degree at MIT, whether it was by making me a birthday cake that actually erupted and came with 20 lbs of dry ice, by making me rice krispies jelly fish, by sending me photos of Mr. T’s latest travels, or by decorating the trailer we shared our first summer at DSL (which I later watched get demolished after it was condemned). My non-WHOI friends have also been very supportive of me during grad school. Lindsey, Sarah, and Heather have

been supportive friends from afar always checking in on me through email and always being available to take a break to root for the Red Sox! Nao Teshima has been a wonderful friend since we were twelve and she wasn't about to miss hearing about the ups and downs of grad school and I thank her tremendously for that. Joshua Davis has been a great friend for many years and has shared my love of the ocean for even longer. He has been invaluable when I needed help with Matlab or needed a second and sometimes first opinion about an engineering problem. His Matlab coding skills are brilliant and I thank him for all the times that I called him for his brain power. Luckily he never charged me the consulting rate that he charges his real clients!

I would like to thank my family for believing in me and for encouraging me for all these years in science and engineering and especially in my love for the ocean. Probably not many parents would let their fifteen year old child fly to the Galapagos Islands to go on a research cruise, but luckily mine did and I thank them for that!

Last but definitely not least I would like to thank Masaya who, when I started grad school, was my boyfriend, and by the end was my husband. Somehow through it all he endured my living in another state from him so that I could pursue my lifelong goal of getting my PhD. I thank him for not only the devotion and the support he showed me but also for the sacrifices he made for me all during my graduate school. During the last three years of grad school, he put 50,000 miles on his car driving from Manhattan every weekend so that I could work to finish my degree and I thank him immensely for that. Finally, I thank him for helping me with all the day-to-day chores that just would not have gotten done otherwise.

I would like to acknowledge the National Science Foundation for support of this research under grants OCE0352278 and OCE0352242. Additional support was received from WHOI's Deep Ocean Exploration Institute who awarded this research with two grants. I am also very personally thankful to the WHOI Ocean Ventures Fund who awarded me a student grant which allowed me to build the titanium pressure cell and to buy many of the optical components that I needed. This funding came at a time when I was in need of laboratory funding and it was much appreciated. I also thank the Department of Defense for awarding me a National Defense Science and Engineering Graduate Fellowship which gave me the flexibility to choose a project that interested me without having to worry about funding.

Contents

1	Introduction	19
1.1	Laser-Induced Breakdown Spectroscopy	19
1.1.1	How LIBS Works	20
1.1.2	Advantages and Disadvantages of the LIBS Technique	21
1.1.3	LIBS in Liquids	23
1.2	Scientific Application: Hydrothermal Vent Chemistry	24
1.3	Thesis Overview	27
	References	28
2	Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: evaluation of key measurement parameters	33
2.1	Abstract	33
2.2	Introduction	34
2.3	Experiment	36
2.4	Results and Discussion	39
2.4.1	The Effect of Pulse Energy on LIBS Emission	39
2.4.2	Interrelationship of pulse energy, gate delay, and pressure for Lithium	47
2.4.3	Effect of NaCl Concentration on LIBS Spectra	47
2.4.4	Detection of Calcium at Varying Concentrations	50
2.4.5	Solution Temperature Effects on Calcium Spectra	50
2.5	Conclusions	51
2.6	Acknowledgments	51
	References	52
3	Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics	57

3.1	Abstract	57
3.2	Introduction	58
3.3	Experimental	61
3.3.1	Echelle Spectrometer Set-up	61
3.3.2	Czerny-Turner Spectrometer Set-up	62
3.4	Results and Discussion	63
3.4.1	The Generalized Extreme Value Distribution	63
3.4.2	Applicability of Extreme Value Statistics	65
3.4.3	Extreme Value Statistical Parameters	70
3.4.4	Variability	75
3.5	Conclusions	77
3.6	Acknowledgements	77
	References	78
4	Single pulse laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: Interrelationship of gate delay and pulse energy	81
4.1	Abstract	81
4.2	Introduction	82
4.3	Experimental	83
4.4	Results and Discussion	87
4.4.1	Sodium	87
4.4.2	Manganese	95
4.4.3	Calcium	101
4.5	Conclusions	110
4.6	Acknowledgments	110
	References	111
5	Double pulse laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: Interrelationship of gate delay, pulse energies, interpulse delay, and pressure	115
5.1	Abstract	115
5.2	Introduction	116
5.3	Experimental	118
5.4	Results and Discussion	121
5.4.1	Magnesium	122
5.4.2	Potassium	125

5.4.3	Calcium	128
5.4.4	Manganese	132
5.4.5	Sodium	136
5.5	Conclusions	143
5.6	Acknowledgments	144
	References	145
6	Preliminary investigations on matrix effects of Na, K, and Ca for bulk liquids at oceanic pressures	149
6.1	Abstract	149
6.2	Introduction	149
6.3	Experimental	150
6.4	Results and Discussion	151
6.4.1	Matrix Effects of K and Na	151
6.4.2	Matrix Effects of Ca and Na	155
6.4.3	Detection of Na and K in a Chloride Versus Sulfate Matrix . .	161
6.5	Conclusions	163
6.6	Acknowledgments	163
	References	164
7	Conclusions and Future Directions	165
7.1	Conclusions	165
7.1.1	Development of a New Data Processing Scheme	165
7.1.2	Single Pulse LIBS	166
7.1.3	Double Pulse LIBS	167
7.1.4	Matrix	168
7.2	Future Work	168
7.2.1	Laboratory Work	168
7.2.2	Design of an Ocean-Going Sensor	170
	References	173

List of Figures

1-1	Chemistry of Hydrothermal Vents	26
2-1	Schematic of the laboratory LIBS apparatus	37
2-2	Optical arrangements	38
2-3	Effect of laser pulse energy on the LIBS signal intensity of Na(I) . . .	40
2-4	Effect of laser pulse energy on the LIBS emission intensity of Mn (I) .	42
2-5	Effect of pressure on LIBS emission intensity	43
2-6	Spectra of Ca under four dual pulse conditions	45
2-7	Ca emission at different interpulse delays	46
2-8	Spectra of Mn under four dual pulse conditions	47
2-9	Effect of gate delay on the LIBS signal for Li	48
2-10	Effect of the addition of NaCl on spectra of Ca	49
2-11	Detection of Ca in a simulated vent fluid	50
3-1	Laboratory set-up using an Echelle spectrometer	61
3-2	Laboratory set-up using a Czerny-Turner spectrometer	63
3-3	q-q plots for the 588.9953 nm Na I peak for halite using the Echelle set-up	67
3-4	q-q plots for the 588.9953 nm Na I peak for bulk aqueous solution using the Echelle set-up	68
3-5	q-q plots for the 588.9953 nm Na I peak for bulk aqueous solution using the Czerny-Turner set-up	69
3-6	Shape parameter, location parameter, and sample mean for halite using the Echelle set-up	72
3-7	Shape parameter, location parameter, and sample mean for bulk aque- ous solution using the Echelle set-up	73
3-8	Shape parameter, location parameter, and sample mean for bulk aque- ous solution using the Czerny-Turner set-up	74
3-9	Images of plasmas formed in bulk aqueous solution	76

3-10	Comparison of laser energies	76
4-1	Laboratory set-up	84
4-2	Timing parameters	85
4-3	Optical Configuration	85
4-4	Interrelationship of gate delay, laser pulse energy, and peak intensity for Na (588.995 nm)	88
4-5	Interrelationship of Pressure, Gate Delay, Energy, and Intensity for Na (589.6 nm)	89
4-6	Interrelationship of pressure, gate delay, E, and SBR for Na (588.995 nm)	90
4-7	Interrelationship of pressure, gate delay, energy, and SBR for Na (589.6 nm)	91
4-8	Na spectra taken with a pulse energy of 40 mJ and a gate delay of 50 ns	92
4-9	Sodium calibration curves	93
4-10	Spectra of sodium made over a range of NaCl concentrations	94
4-11	Interrelationship of pressure, t_d , E, and intensity for Mn (403.076 nm)	96
4-12	Interrelationship of pressure, t_d , E, and SBR for Mn (403.076 nm) . .	97
4-13	Manganese spectra using a 30 mJ energy pulse and a gate delay of 50 ns	98
4-14	Mn calibration curve	99
4-15	Spectra of manganese made at a range of concentrations	100
4-16	Interrelationship of pressure, t_d , E, and intensity for Ca (393 nm) . .	101
4-17	Interrelationship of pressure, t_d , energy, and intensity for Ca (396 nm)	102
4-18	Interrelationship of pressure, t_d , energy, and intensity for Ca (422 nm)	103
4-19	Interrelationship of pressure, t_d , E, and SBR for Ca (393 nm)	104
4-20	Interrelationship of pressure, t_d , E, and SBR for Ca (396 nm)	105
4-21	Interrelationship of pressure, t_d , E, and SBR for Ca (422 nm)	106
4-22	Calcium spectra using 30 mJ and a 50 ns gate delay	107
4-23	Calcium calibration curves	108
4-24	Calcium spectra at 2.76×10^7 Pa	109
5-1	Laboratory set-up	119
5-2	Timing parameters	120
5-3	Mg (I) optimization at 1×10^5 Pa	122
5-4	Mg (I) optimization at 1.38×10^7 Pa	123
5-5	Mg (I) optimization at 2.76×10^7 Pa	123
5-6	Mg (I) calibration curve	124

5-7	Mg (I) spectra	124
5-8	K (I) optimization at 1×10^5 Pa	125
5-9	K (I) optimization at 1.38×10^7 Pa	126
5-10	K (I) optimization at 2.76×10^7 Pa	126
5-11	K (I) calibration	127
5-12	K (I) spectra	127
5-13	Ca (I) optimization at 1×10^5 Pa	128
5-14	Ca (I) optimization at 1.38×10^7 Pa	129
5-15	Ca (I) optimization at 2.76×10^7 Pa	129
5-16	Ca calibration	130
5-17	Ca spectra	131
5-18	Mn (I) optimization at 1×10^5 Pa	133
5-19	Mn (I) optimization at 1.38×10^7 Pa	133
5-20	Mn (I) optimization at 2.76×10^7 Pa	134
5-21	Mn (I) calibration curve	134
5-22	Mn (I) spectra	135
5-23	Na (I) optimization at 1×10^5 Pa	136
5-24	Na (I) optimization at 1.38×10^7 Pa	137
5-25	Na (I) optimization at 2.76×10^7 Pa	137
5-26	Na (I) calibration	138
5-27	Spectra of Na (I) peaks	139
5-28	Effect of interpulse delay on intensity for Na at 1×10^5 Pa	141
5-29	Na calibration curve at 1×10^5 Pa	141
5-30	Na (I) spectra	142
6-1	Effect of presence of Na on peak intensity on K	152
6-2	Effect of presence of K on peak intensity on Na	154
6-3	Effect of presence of Na on peak intensity on Ca (II)	156
6-4	Effect of presence of Na on peak intensity on Ca (I)	157
6-5	Ratio of 393 nm Ca (II) peak to 422 nm Ca (I) peak	158
6-6	Effect of presence of Ca on peak intensity on Na	160
6-7	Detection of potassium in a chloride versus sulfate matrix	161
6-8	Detection of sodium in a chloride versus sulfate matrix.	162
7-1	Future oceanic LIBS system	172

List of Tables

2.1	Conditions used to study the effect of dual pulse energies	44
2.2	Dual pulse emission intensity	45
4.1	Calibration curve conditions	86
6.1	Conditions used to study the K-Na matrix effect	151
6.2	Conditions used to study the Na-K matrix effect	153
6.3	Conditions used to study the Ca-Na Matrix Effect	155
6.4	Conditions used to study the effect of $\text{CaCl}_2 \cdot 2\text{H}_2\text{O}$ on Na	159

Chapter 1

Introduction

The development of *in situ* chemical sensors is needed to explore and understand the ocean and its processes. In present day oceanographic work, new sensors are required for expeditionary science with underwater vehicles such as ROVs, AUVs, and submersibles. A new paradigm for ocean study has begun with the implementation of ocean observatories. As these permanent observatories become the new mode of oceanography, there will be a critical need for chemical sensors capable of long-term deployment for ocean observatories to reach their full potential. The time required to transform a bench-top laboratory technique into a full ocean-going system is significant. The development phase initially requires validation that an analytical technique will work under *in situ* conditions. In this thesis, laser-induced breakdown spectroscopy (LIBS) is evaluated in the laboratory to determine if this bench-top analytical technique is viable for development into a field-going oceanographic chemical sensor.

1.1 Laser-Induced Breakdown Spectroscopy

Laser-induced breakdown spectroscopy is a type of atomic emission spectroscopy that was first reported in the literature in 1962 and has since evolved into a technique for laboratory chemical analysis [1]. LIBS technology is currently undergoing transformation from a benchtop analytical technique into a viable tool for field measurements, and is emerging as a tool for chemical, geochemical, and environmental analysis in extreme and hostile environments [2]. LIBS has been used in the analysis of a wide variety of sample types including soils [3], archaeological artifacts [4], metal alloys [5], bacterial spores, molds, pollens, and proteins [6], pharmaceuticals [7], glass [8], nuclear power station steam generator tubes [9], and pigments in artwork [10]. A

mobile instrument has been designed for studying polluted soils [11, 12]. LIBS is also capable of stand-off, non-contact measurements, and a field deployable system has been proposed by Palanco *et al.* with a detection range on the order of hundreds of meters [13]. A field-portable LIBS system has been developed for landmine detection [14]. Along with land-based environmental applications, LIBS is finding applicability to space exploration [15–20]. LIBS will be part of the ChemCam instrument package for the Mars Science Laboratory Rover that will be launched in 2009 for geological analysis.

Although there are many proposed land and space based applications of LIBS field-going sensors, it also holds promise for *in situ* ocean use. There are numerous possible applications in the areas of chemical, geological, and biological oceanography ranging from laboratory experimentation to fieldwork in environments from estuaries to the coastal zone to the deep ocean. The development of an oceanographic LIBS sensor could allow scientists to determine the chemical composition of sediments, rocks, or ocean fluids in an *in situ*, real-time mode. Several researchers have already applied the LIBS technology to marine-related applications. For example, Niu *et al.* successfully used LIBS to determine strontium levels in marine algae [21] and Barbini *et al.* used LIBS shipboard to analyze marine sediments [22]. In the laboratory, De Giacomo *et al.* have explored the use of LIBS for the detection of submerged solid targets [23–26].

1.1.1 How LIBS Works

The LIBS technique is based on the analysis of the spectral emission from laser-induced plasmas, and is a type of atomic emission spectroscopy. Atomic emission spectroscopy uses an external energy source to excite ground state atoms. The atoms spontaneously emit radiation as they revert back to a lower energy state, with the emission intensity being proportional to the concentration of atoms in the ground state [27]. In the LIBS technique, one or two high-power, pulsed lasers, typically Q-switched Nd:YAG units, are used as the excitation source. The lasers are usually operated at the fundamental wavelength of 1064 nm, although other wavelengths can be used; for example, the lasers can be frequency doubled to deliver 532 nm. Additional components of a typical LIBS system include focusing and collection optics, a spectrometer, and a data acquisition computer. Commercially available broadband spectrometers span the 200 - 1000 nm wavelength region, allowing for the simultaneous detection of multiple elements [28].

To generate a plasma spark, a short duration, high power laser beam is focused onto or into a sample. An optically-induced plasma or spark is formed on the surface (of a solid or on a liquid) or in the sample (in bulk liquid or in a gaseous medium) when the laser power density or irradiance exceeds the breakdown threshold of the sample [27]. Breakdown threshold irradiances using a 7-ns pulse width, 1064 nm Nd:YAG laser have been reported by Kennedy *et al.* to be 5.6×10^9 W/cm² for tap water and 8.31×10^9 W/cm² for saline solutions [29]. The plasma radiates both a continuum component due to inverse Bremsstrahlung radiation from electron-ion collisions, which decays rapidly, and an emission line component that decays more slowly. Therefore, the plasma emission can be analyzed by spectroscopic methods by time gating. The plasma light is initially dominated by a white light continuum which contains little intensity variation as a function of wavelength. After breakdown, the plasma expands outwards and back towards the focusing lens. The expansion occurs at 10^5 m/s and creates an audible shock wave [30]. Plasma temperatures in bulk liquids are in the range of 7,000 - 12,000 K [30]. Plasma decay occurs by radiative, quenching, and electron-ion recombination processes that result in the formation of neutral species [31].

A gated spectrometer covering part or all of the ultraviolet through near infrared range (200 nm - 1000 nm) is used to capture the plasma spectrum. For simultaneous multi-element analysis, an Echelle spectrometer can be used which contains an Echelle diffraction grating with coarse grooves and large blaze angles. The grooves have steep sides to cover the full range of wavelengths, and a prism is needed to separate overlapping orders of the grating [27, 32]. CCDs (charge coupled devices) or ICCDs (intensified charge coupled devices) serve as the detector devices within the spectrometers. The spectral line wavelengths and intensities obtained from plasma ablation can be compared with a standard atomic line reference and/or calibrated against samples of known makeup to determine the elemental composition of the sample. The intensity of the spectral lines provides a quantitative description of the elemental concentration [28, 33].

1.1.2 Advantages and Disadvantages of the LIBS Technique

LIBS as a spectrochemical technique possesses many advantages which make it especially attractive for development into an *in situ* sensor for oceanographic research. However, the technique also possesses several disadvantages that must be addressed.

LIBS can provide rapid multi-element analysis and has the capability to detect elements in the parts-per-million (ppm) range. All elements emit in the 200 - 940 nm wavelength region and every element has its own unique spectral signature; therefore, all elements can theoretically be detected with LIBS. The emission lines obtained from plasma ablation can be compared with a standard reference of atomic lines to determine the chemical composition of the sample by comparison of wavelengths. For example, the National Institute of Standards and Technology (NIST) maintains an on-line atomic spectrum database (<http://physics.nist.gov/PhysRefData/ASD/index.html>). Unlike other techniques that are useful for analyzing one form of material, LIBS can chemically analyze all three states of matter (solid, liquid, and gas). The sample size for LIBS is minimal with typically, 0.1 μg - 0.1 mg of material ablated if a solid sample is analyzed. The sample size required minimizes the destructiveness and invasiveness of the technique. LIBS does not require any preparation of a sample before analysis and unlike many analytical techniques, does not require chemical reagents to dissolve the sample [27]. This avoids contamination problems and reduces the time for analysis. These characteristics makes LIBS suitable for real-time, *in situ* analysis. Most analytical methods (e.g., wet chemistry techniques) cannot produce instant results in a field environment. Other methods can be time-intensive, with several days needed before results are available. Rapid analysis is possible with LIBS as it only requires one laser pulse to generate a plasma, although typically numerous accumulations of plasmas are obtained. A large number of measurements can be taken quickly, making the elemental composition of the sample identifiable on a nearly real-time basis.

LIBS has been identified as a viable technique for use in extreme environments because direct contact with the sample is not necessary; only optical access is required. Fiber optics can be used to reach distances far from the laser. The laser can also be focused on a sample at a distance, making LIBS useful for remote measurements. The stand-off analysis capabilities of LIBS make it a viable *in situ* analytical technique. The components of the system can also be made small and rugged. Unlike many traditional analytical techniques that require collection of a sample followed by the transport of the sample to a laboratory, LIBS measurements can be made directly in the field. This eliminates the need to store the sample in conditions that differ from the natural environment. This is an important advantage for oceanography due to pressure and temperature effects. LIBS can also be used for surface cleaning of a sample. Repetitive sampling at the same spot can be used for ablation through weathered surfaces to reach underlying material. This could be useful for rocks which could be “drilled” through to areas beneath fouling and to determine chemical makeup

of the fouling. These benefits make LIBS useful for chemical analysis in extreme environments suggesting that it is a viable technique for use in the deep ocean [28, 31, 34–36].

Although there are many advantages to the LIBS technique, there are several disadvantages that must be acknowledged. LIBS is a semi-quantitative technique as it is often difficult to obtain suitable standards. LIBS does not have the sensitivity and precision of many state-of-the-art laboratory analytical techniques. Furthermore, measurements are highly variable. There are also chemical interference (matrix) effects between analytes that must be addressed [27].

1.1.3 LIBS in Liquids

Several researchers have been successful in causing plasma ablation to occur on materials submerged in water and on liquid surfaces. However, only a few researchers have focused their work on LIBS of dissolved analytes within bulk aqueous solutions [23, 37–42] due to the inherent difficulty of detection. If the liquid is transparent at the laser wavelength, a plasma can be formed in the bulk liquid below the surface [37]. The plasma formed in a bulk liquid compared to that formed on a solid or in air displays reduced light intensity and emission lifetime due to quenching [37, 38, 43, 44]. This emission time interval is on the order of 1 μ s or less in liquid, which is significantly lower than at the air-liquid interface where the time interval averages 5 - 20 μ s. In addition, spectral lines are broadened by the Stark effect which results from ion and electron collisions [37, 45]. Furthermore, “moving breakdown” occurs that randomly changes the distance between the plasma and the collection optics, a phenomenon that is not relevant to solids in air. The plasma expands along the beam path, resulting in an elongated plasma that cavitates cylindrically [46]. For many aqueous applications, these issues can be avoided by analysis on a liquid surface, jet, or film; however, for the development of an *in situ* oceanic system, it is necessary to work directly with bulk liquids.

As early as 1984, Cremers *et al.* [37] were able to identify many elements in an aqueous solution. This work showed that the use of two laser sparks, (double or dual pulse LIBS), with a short interpulse delay, improved the detection limit for dissolved species. Using two laser pulses allowed lower concentrations of elements to be detected. The first spark creates a cavitation bubble. The second pulse is fired into the bubble, forming a plasma within it, allowing the measurement to be made in a gaseous environment.

Few LIBS researchers have focused on liquids and no work has been done at high pressures, and therefore little is known about the pressure effect. Relatively few studies have focused on dissolved species in liquids and therefore much work on the experimental parameters for measurements of such species is necessary.

1.2 Scientific Application: Hydrothermal Vent Chemistry

Study of *in situ* hydrothermal vent chemistry could benefit greatly from the development of a LIBS sensor. At mid-ocean ridges, seawater circulates through the fractured and permeable crust. Interactions with the surrounding rock induces major chemical changes to the fluid. At vent orifices, exit temperatures reach 200 - 405°C at ambient pressures of 8.1×10^6 Pa to 3.6×10^7 Pa corresponding to ocean depths of 800 m to 3600 m [47]. The circulation is driven by the direct or indirect thermal effects of magma at sub-seafloor depths of up to a few km; additionally, substantial changes in fluid composition occur due to interaction with the host rock, phase separation, and possibly magma degassing [48]. For example, many alkalis (e.g., Li, Na, and Ca) and transition metals (e.g., Fe, Mn, Cu, and Zn) are leached from the host rock and concentrated to varying degrees in the fluid, while Mg and SO_4 are largely removed from the fluid [47].

As vent fluids exit the seafloor, the very hot fluids mix with ambient seawater which rapidly changes the chemistry (Figure 1-1). Measuring the fluid properties *in situ* is very difficult. Collecting samples for measurement shipboard or back in a shore laboratory is usually done by using non-reactive titanium samplers to extract water, which is then brought to the surface. Some elements remain in solution as the temperature and pressure changes, however, others precipitate out [49]. In addition, the chemistry of some vents are known to change over short (days to years) time scales [50]. The use of a method like LIBS, therefore, is attractive for obtaining an understanding of the chemistry of vents that has thus far been impossible to achieve.

Six elements (sodium, calcium, manganese, magnesium, potassium, and lithium) were selected as the primary focus of this work because of their key importance at hydrothermal vents. Sodium (Na) is the most abundant cation in vent fluids and its study is important for gaining an understanding of phase separation processes [48]. Manganese (Mn) exists as a trace metal in seawater but has a higher concentration in vent fluids due to leaching from the host rock [48]. Mn can also be studied si-

multaneously with Fe as an indication of subsurface deposition as Fe precipitates out while Mn stays in solution. Calcium (Ca) is the second most abundant cation in vent fluids, and is typically enriched in vent fluids compared to seawater [50]. Ca is released into vent fluids when sodium is taken up during albitization reactions with the host rock [50]. Magnesium (Mg) is very low to nonexistent in hydrothermal vent fluids; however, it is present in seawater [48]. If Mg is detected in vent fluid samples, it is evidence for contamination; thus, a sensor that can detect Mg is desirable. Potassium (K) and Lithium (Li) are typically highly enriched in vent fluids due to leaching from basalts [48]. In vent fluids, concentrations range from approximately 250 - 23,163 ppm for Na, 0.6 - 399 ppm for Mn, -54 - 4477 ppm for Ca, -47 - 3166 ppm for K, 0 ppm for Mg, and 0.7 - 1073 ppm for Li [47]. In seawater, concentrations are approximately 10933 ppm Na, <0.001 Mn, 419 ppm Ca, 405 ppm K, 1300 ppm Mg, and 0.2 ppm Li [47].




Figure 1-1: At hydrothermal vents, the cold seawater seeps down through the permeable seafloor where it undergoes chemical changes. The hot vent fluid finally vents at the seafloor. Illustration by E. Paul Oberlander. Reprinted from *Oceanus*, Dec. 1, 1998 with permission from Susan Humphris, WHOI.

1.3 Thesis Overview

New sensors take a significant amount of time to develop; thus, the evaluation of techniques in the laboratory for use in the ocean environment is becoming increasingly important. This thesis focuses on this proof-of-concept phase, in which the LIBS analytical technique is evaluated in the laboratory under *in situ* conditions. It is divided into five chapters that cover single and double pulse LIBS and delves into the parameters that must be optimized for the detection of elements in high pressure aqueous solutions. A new data processing scheme for dealing with the inherent variability of laser-induced plasmas is developed in this thesis. This processing scheme is applied to all data presented in Chapters 4 - 6.

Chapter Two, "Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: Evaluation of key measurement parameters," is a manuscript that appeared in the 1 May 2007 issue of *Applied Optics* [40]. It presents preliminary investigations on the feasibility of using LIBS to detect analytes in bulk liquids at oceanic pressures. This work was completed as part of two extensive research visits to the University of South Carolina.

Chapter Three, "Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics," is a manuscript that has been accepted for publication by *Spectrochimica Acta: Part B* [51]. It presents a new data processing approach for LIBS spectra.

Chapters Four and Five are complementary chapters that look at the detection of analytes in bulk aqueous solutions at oceanic pressures using single pulse (Chapter 4) and double pulse (Chapter 5) LIBS. These two chapters focus on the optimization of the key experimental parameters for the detection of analytes. Chapter Four deals with the detection of three elements: sodium, calcium, and manganese and the interrelationship of pressure, gate delay, and pulse energy. Chapter Five concentrates on the detection of sodium, calcium, manganese, potassium, and magnesium and the interrelationship of pressure, gate delay, pulse energies, and interpulse delay. In both chapters, calibration curves and limits of detection are presented.

Chapter Six presents preliminary investigations into matrix effects for three elements: sodium, calcium, and potassium. This chapter also explores the effect of a chloride versus sulfate background matrix on the detection of sodium and potassium.

Bibliography

- [1] F. Brech and L. Cross. Optical microemission stimulated by a ruby maser. *Applied Spectroscopy*, 16:59, 1962.
- [2] R. S. Harmon, F. C. DeLucia, C. E. McManus, N. J. McMillan, T. F. Jenkins, M. E. Walsh, and A. Miziolek. Laser-induced breakdown spectroscopy - an emerging chemical sensor technology for real-time field-portable, geochemical, mineralogical, and environmental applications. *Applied Geochemistry*, 21(5):730–747, May 2006.
- [3] A. S. Eppler, D. A. Cremers, D. D. Hickmott, M. J. Ferris, and A. C. Koskelo. Matrix effects in the detection of Pb and Ba in soils using laser-induced breakdown spectroscopy. *Applied Spectroscopy*, 50:1175–1181, 1996.
- [4] D. Anglos. Laser-induced breakdown spectroscopy in art and archaeology. *Applied Spectroscopy*, 55:186A–205A, 2001.
- [5] G. Galbacs, I. B. Gornushkin, B. W. Smith, and J. D. Winefordner. Semi-quantitative analysis of binary alloys using laser-induced breakdown spectroscopy and a new calibration approach based on linear correlation. *Spectrochimica Acta Part B*, 56:1159–1173, 2001.
- [6] A. C. Samuels, F. C. DeLucia, Jr., K. L. McNesby, and A. W. Miziolek. Laser-induced breakdown spectroscopy of bacterial spores, molds, pollens, and protein: initial studies of discrimination potential. *Applied Optics*, 42(30):6205–6209, 2003.
- [7] L. St-Onge, E. Kwong, M. Sabsabi, and E. B. Vadas. Quantitative analysis of pharmaceutical products by laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 57(7):1131–1140, 2002.
- [8] N. Carmona, M. Oujjab, E. Rebollar, H. Romich, and M. Castillejo. Analysis of corroded glasses by laser induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 2005.
- [9] A. I. Whitehouse, J. Young, I. M. Botheroyd, S. Lawson, C. P. Evans, and J. Wright. Remote material analysis of nuclear power station steam generator tubes by laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 56:821–830, 2001.
- [10] D. Anglos, S. Couris, and C. Fotakis. Artworks: laser-induced breakdown spectroscopy in pigment identification. *Applied Spectroscopy*, 51:1025–1030, 1997.
- [11] A. Bertolini, G. Carelli, F. Francesconi, M. Francesconi, L. Marchesini, P. Marsili, F. Sorrentino, G. Cristoforetti, S. Legnaioli, V. Palleschi, L. Pardini, and A. Salvetti. Modi: a new mobile instrument for *in situ* double-pulse LIBS analysis. *Analytical and Bioanalytical Chemistry*, V385(2):240–247, May 2006.

- [12] M. Corsi, G. Cristoforetti, M. Hidalgo, S. Legnaioli, V. Palleschi, A. Salvetti, E. Tognoni, and C. Vallebona. Double pulse, calibration-free laser-induced breakdown spectroscopy: A new technique for *in situ* standard-less analysis of polluted soils. *Applied Geochemistry*, 21(5):748–755, 2006.
- [13] S. Palanco, C. Lopez-Moreno, and J. J. Laserna. Design, construction and assessment of a field-deployable laser-induced breakdown spectrometer for remote elemental sensing. *Spectrochimica Acta Part B*, 61(1):88–95, January 2006.
- [14] R. S. Harmon, F. C. DeLucia, A. LaPointe, R. J. Winkel, and A. W. Miziolek. LIBS for landmine detection and discrimination. *Analytical and Bioanalytical Chemistry*, V385(6):1140–1148, 2006.
- [15] Z. A. Arp, D. A. Cremers, R. C. Wiens, D. M. Wayne, B. Sallé, and S. Maurice. Analysis of water ice and water ice/soil mixtures using laser-induced breakdown spectroscopy: Application to Mars polar exploration. *Applied Spectroscopy*, 58:897–909, 2004.
- [16] Z. A. Arp, D. A. Cremers, R. D. Harris, D. M. Oschwald, G. R. Parker Jr., and D. M. Wayne. Feasibility of generating a useful laser-induced breakdown spectroscopy plasma on rocks at high pressure: preliminary study for a Venus mission. *Spectrochimica Acta Part B*, 59:987–999, 2004.
- [17] G. B. Courrèges-Lacoste, B. Ahlers, and F. R. Pérez. Combined Raman spectrometer/laser-induced breakdown spectrometer for the next ESA mission to Mars. *Spectrochimica Acta Part A*, In Press, 2007.
- [18] R. Brennetot, J. L. Lacour, E. Vors, A. Rivoallan, D. Vailhen, and S. Maurice. Mars analysis by laser-induced breakdown spectroscopy (MALIS): Influence of Mars atmosphere on plasma emission and study of factors influencing plasma emission with the use of Doehlert designs. *Applied Spectroscopy*, 57(7):744–752, 2003.
- [19] A. Knight, N. Scherbarth, D. Cremers, and M. Ferris. Characterization of laser-induced breakdown spectroscopy (LIBS) for application to space exploration. *Applied Spectroscopy*, 54:331–340, 2000.
- [20] B. Sallé, J.-L. Lacour, P. Mauchien, P. Fichet, S. Maurice, and G. Manhes. Comparative study of different methodologies for quantitative rock analysis by laser-induced breakdown spectroscopy in a simulated Martian atmosphere. *Spectrochimica Acta Part B*, 61:301–313, 2006.
- [21] L. Niu, H. Cho, K. Song, H. Cha, Y. Kim, and Y. Lee. Direct determination of strontium in marine algae samples by laser-induced breakdown spectrometry. *Applied Spectroscopy*, 56:1511–1514, 2002.
- [22] R. Barbini, F. Colao, V. Lazic, R. Fantoni, A. Palucci, and M. Angelone. On board LIBS analysis of marine sediments collected during the XVI Italian campaign in Antarctica. *Spectrochimica Acta B*, 57:1203–1218, 2002.

- [23] A. De Giacomo, M. Dell’Aglio, F. Colao, R. Fantoni, and V. Lazic. Double-pulse LIBS in bulk water and on submerged bronze samples. *Applied Surface Science*, 247:157–162, 2005.
- [24] A. De Giacomo, M. Dell’Aglio, F. Colao, and R. Fantoni. Double pulse laser produced plasma on metallic target in seawater: basic aspects and analytical approach. *Spectrochimica Acta B*, 59:1431–1438, 2004.
- [25] A. De Giacomo, M. Dell’Aglio, A. Casavola, G. Colonna, O. De Pascale, and M. Capitelli. Elemental chemical analysis of submerged targets by double-pulse laser-induced breakdown spectroscopy. *Analytical and Bioanalytical Chemistry*, V385(2):303–311, 2006.
- [26] A. De Giacomo, M. Dell’Aglio, O. De Pascale, and M. Capitelli. From single pulse to double pulse ns-laser induced breakdown spectroscopy under water: elemental analysis of aqueous solutions and submerged solid samples. *Spectrochimica Acta Part B*, in press, 2007.
- [27] Y. Lee, K. Song, and J. Sneddon. *Laser-induced breakdown spectrometry*. Nova Science Publishers, Inc, Huntington, NY, 2000.
- [28] X. Hou and B. T. Jones. Field instrumentation in atomic spectroscopy. *Microchemical Journal*, 66:115–145, 2000.
- [29] P. K. Kennedy, S. A. Boppart, D. X. Hammer, B. A. Rockwell, G. D. Noojin, and W. P. Roach. A first-order model for computation of laser-induced breakdown thresholds in ocular and aqueous media: Part II - comparison to experiment. *IEEE Journal of Quantum Electronics*, 31:2250–2257, 1995.
- [30] D. A. Cremers and L. J. Radziemski. *Laser-Induced Breakdown Spectroscopy (LIBS): Fundamentals and Applications*, chapter History and Fundamentals of LIBS, pages 1–39. Cambridge University Press, 2006.
- [31] D. A. Rusak, B. C. Castle, B. W. Smith, and J. D. Winefordner. Recent trends and the future of laser-induced plasma spectroscopy. *Trends in Analytical Chemistry*, 17:453–461, 1998.
- [32] J. Ingle Jr. and S. Crouch. *Spectrochemical Analysis*. Prentice Hall, New Jersey, 1988.
- [33] R. Nyga and W. Neu. Double-pulse technique for optical emission spectroscopy of ablation plasmas of samples in liquids. *Optics Letters*, 18:747–749, 1993.
- [34] R. T. Wainner, R. S. Harmon, A. W. Miziolek, K. L. McNesby, and P. D. French. Analysis of environmental lead contamination: Comparison of LIBS field and laboratory instruments. *Spectrochimica Acta Part B*, 56:777–793, 2001.

- [35] D. A. Rusak, B. C. Castle, B. W. Smith, and J. D. Winefordner. Fundamentals and applications of laser-induced breakdown spectroscopy. *Critical Reviews in Analytical Chemistry*, 27:257–290, 1997.
- [36] E. Tognoni, V. Palleschi, M. Corsi, and G. Cristoforetti. Quantitative micro-analysis by laser-induced breakdown spectroscopy: A review of the experimental approaches. *Spectrochimica Acta Part B*, 57:1115–1130, 2002.
- [37] D. A. Cremers, L. J. Radziemski, and T. R. Loree. Spectrochemical analysis of liquids using the laser spark. *Applied Spectroscopy*, 38:721–729, 1984.
- [38] R. Knopp, F. J. Scherbaum, and J. I. Kim. Laser induced breakdown spectroscopy (LIBS) as an analytical tool for the detection of metal ions in aqueous solutions. *Fresenius' Journal of Analytical Chemistry*, 355:16–20, 1996.
- [39] W. Pearman, J. Scaffidi, and S. M. Angel. Dual-pulse laser-induced breakdown spectroscopy in bulk aqueous solution with an orthogonal beam geometry. *Applied Optics*, 42:6085–6093, 2003.
- [40] A. P. M. Michel, M. Lawrence-Snyder, S. M. Angel, and A. D. Chave. Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: evaluation of key measurement parameters. *Applied Optics*, 46, 2007.
- [41] M. Lawrence-Snyder, J. Scaffidi, S. M. Angel, A. P.M. Michel, and A. D. Chave. Sequential-pulse laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, 61:171–176, 2007.
- [42] M. Lawrence-Snyder, J. Scaffidi, S. M. Angel, A. P. M. Michel, and A. D. Chave. Laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, 60:786–790, 2006.
- [43] A. E. Pichahchy, D. A. Cremers, and M. J. Ferris. Elemental analysis of metals under water using laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 52:25–39, 1997.
- [44] C. Haisch, J. Liermann, U. Panne, and R. Niessner. Characterization of colloidal particles by laser-induced plasma spectroscopy (LIPS). *Analytica Chimica Acta*, 346:23–25, 1997.
- [45] D. A. Cremers and L. J. Radziemski. *Handbook of Laser-Induced Breakdown Spectroscopy*. John Wiley and Sons, Ltd., 2006.
- [46] P. K. Kennedy, D. X. Hammer, and B. A. Rockwell. Laser-induced breakdown in aqueous media. *Progress in Quantum Electronics*, 21:155–248, 1997.
- [47] C. R. German and K. L. Von Damm. *Treatise on Geochemistry*, chapter Hydrothermal Processes, pages 181–222. Elsevier, 2003.

- [48] K. L. Von Damm. Controls on the chemistry and temporal variability of seafloor hydrothermal fluids. *Seafloor hydrothermal systems: physical, chemical, biological, and geological interactions: Geophysical Monograph 91*, pages 222–247, 1995.
- [49] J. H. Trefry, D. B. Butterfield, S. Metz, G. J. Massoth, R. P. Trocine, and R. A. Feely. Trace metals in hydrothermal solutions from Cleft segment on the southern Juan de Fuca Ridge. *Journal of Geophysical Research*, 99:4925–4935, 1994.
- [50] K. L. Von Damm. Chemistry of hydrothermal vent fluids from 9° - 10° N, East Pacific Rise: ‘Time zero,’ The immediate post-eruptive period. *Journal of Geophysical Research*, 105:11203–11222, 2000.
- [51] A. P. M. Michel and A. D. Chave. Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics. *Spectrochimica Acta Part B*, In Press.

Chapter 2

Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: evaluation of key measurement parameters

The work in this chapter is published in the 1 May 2007 issue of *Applied Optics* as A. P. M. Michel, M. Lawrence-Snyder, S. M. Angel, and A. D. Chave, “Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: evaluation of key measurement parameters.”

2.1 Abstract

The development of *in situ* chemical sensors is critical for present day expeditionary oceanography and the new mode of ocean observing systems that we are entering. New sensors take a significant amount of time to develop; therefore, validation of techniques in the laboratory for use in the ocean environment is necessary. Laser-induced breakdown spectroscopy (LIBS) is a promising *in situ* technique for oceanography. Laboratory investigations on the feasibility of using LIBS to detect analytes in bulk liquids at oceanic pressures were carried out. LIBS was successfully used to detect dissolved Na, Mn, Ca, K, and Li at pressures up to 2.76×10^7 Pa. The effects of pressure, laser pulse energy, interpulse delay, gate delay, temperature, and NaCl

concentration on the LIBS signal were examined. An optimal range of laser pulse energies was found to exist for analyte detection in bulk aqueous solutions at both low and high pressures. No pressure effect was seen on the emission intensity for Ca and Na and an increase in emission intensity with increased pressure was seen for Mn. Using the dual pulse technique for several analytes, a very short interpulse delay resulted in the greatest emission intensity. The presence of NaCl enhanced the emission intensity for Ca, but had no effect on peak intensity of Mn or K. Overall, increased pressure, the addition of NaCl to a solution, and temperature did not inhibit detection of analytes in solution and sometimes even enhanced the ability to detect the analytes. The results suggest that LIBS is a viable chemical sensing method for *in situ* analyte detection in high pressure environments like the deep ocean.

2.2 Introduction

Since laser-induced breakdown spectroscopy (LIBS) was first reported in 1962, the technique has evolved into a widely used method for laboratory analytical chemistry [1–8]. Due to several advantages over other methods, LIBS has been identified as a viable tool for *in situ* measurements, especially in extreme environments [9, 10]. The technique yields simultaneous sensitivity to virtually all elements in the parts-per-million (ppm) or better range in solids, gases, aerosols, and at the gas-liquid interface. LIBS is effectively non-invasive, requiring only a small sample (typically, pg to ng of material are ablated). Unlike for many analysis techniques, the sample does not need to be prepared. LIBS requires only optical access to a sample and therefore can be used in a stand-off mode without perturbing the sample environment. LIBS measurements are essentially real-time, with typical sampling rates of less than one second per cycle. These characteristics are all required for *in situ* chemical sensing in the ocean [11–15].

Although researchers have been successful at inducing plasma ablation on submerged materials [16], on a water surface or film [17–22], and in liquid jets, droplets, and flowing solutions, [23–29] only limited LIBS work has focused on analyte detection within bulk aqueous solutions [30–32]. Furthermore, the work within bulk aqueous solutions has been at atmospheric pressure. Pioneering work by Cremers *et al.* [30] showed that LIBS could identify Li, Na, K, Rb, Cs, Be, Ca, B, and Al in aqueous solutions with varying detection limits, but typically at the ppm level. Several studies in bulk liquids have displayed a reduction in the time during which plasma emission can be observed as compared to that in air [16, 30, 31, 33]. The

plasma lifetime is typically $\leq 1 \mu\text{s}$ in bulk liquids, whereas at an air-liquid interface it averages 5-20 μs . Laser-induced plasmas formed in solution are also characterized by a reduction in plasma light intensity.

The effects of elevated pressure and temperature on LIBS spectra have received limited attention. Although a few researchers report on LIBS at super-atmospheric pressures, they do not extend beyond $1 \times 10^7 \text{ Pa}$ (note: $1 \text{ Pa} = 1 \times 10^{-5} \text{ bars}$), which is well below the ambient pressure in the deep ocean [9, 34]; yet, none of these studies were for liquids. Although, we have previously reported preliminary findings that show the ability to detect analytes in high pressure bulk aqueous solutions [35], we now focus on the key measurement parameters that are needed for analyte detection. The influence of *in situ* temperature is anticipated to be weak because of the high plasma temperature ($\approx 8000\text{K}$ at early times) [36–39].

For many years, oceanography has been in an expeditionary mode where research vessels are used for short term instrument deployments with limited resolution in time. Although oceanographers will continue to study the ocean in this way, a new paradigm using ocean observatories for long term *in situ* observing is upon us. As this shift towards long term ocean observing systems becomes recognized, we must also acknowledge the need for *in situ* sensors; especially those capable of temporal studies. A major need is for chemical sensors. The development of new sensors for oceanography takes a significant amount of time, and hence laboratory validation of techniques such as LIBS is necessary now to identify techniques that are viable for chemical detection in high pressure, high salinity, aqueous environments.

Although LIBS has the potential for use in numerous ocean environments, and has applicability to solids and liquids, we have focused on the feasibility of detecting elements at one extreme ocean environment, hydrothermal vents. Hydrothermal venting occurs on mid-ocean ridges where seawater circulates through the fractured and permeable oceanic crust. Exit temperatures at discrete (orifice diameters of a few centimeters) high-temperature vents range from 200-405°C at ambient pressures of 1.5×10^7 to $3.7 \times 10^7 \text{ Pa}$. Low-temperature (usually $< 35^\circ\text{C}$) diffuse flow seeping from porous surfaces or cracks is frequently observed [40]. The circulation is driven by the direct or indirect thermal effects of magma at sub-seafloor depths of up to a few km. Substantial changes in fluid composition occur due to interaction with the host rock, phase separation into a mixed liquid-vapor form, and possibly magma degassing. Many alkalis (e.g., Li, Na, and Ca) and transition metals (e.g., Fe, Mn, Cu, and Zn) are leached from the host rock and concentrated to varying degrees in the hydrothermal fluid, while Mg and SO_4 are largely removed from the fluid by pre-

precipitation into Mg-OH-Si minerals and anhydrite, respectively [40]. Von Damm and Butterfield *et al.* provide comprehensive reviews of the chemistry of hydrothermal vent fluids [40, 41].

In this paper, we explore the effect of vent system environmental factors such as pressure, temperature, and NaCl concentration on the LIBS signal to assess the feasibility of developing LIBS for *in situ* chemical sensing in the ocean. In addition, several system parameters (laser energy per pulse, interpulse spacing, and gate delay) are optimized for high pressures for the first time.

2.3 Experiment

A laboratory LIBS system was designed to operate with a high pressure cell (Figure 2-1). For single pulse experiments, a Continuum Surelite III laser (5-ns pulse width) was utilized. For dual pulse experiments, a Quantel Nd 580 (9-ns pulse width) was used for the first laser pulse followed by a second pulse from the Surelite laser. Both lasers were Q-switched Nd:YAG types operated at the fundamental wavelength with a repetition rate of 5 Hz. For dual pulse experiments, a variable clock (Stanford Instruments Model SR250) and a delay generator (Stanford Instruments Model DG535) controlled laser triggering.

The laser pulses were focused into a high pressure cell, designed to reach pressures of 3.45×10^7 Pa and constructed of stainless steel Swagelok fittings with six 1"-ID (1 in. = 2.54 cm) and 1- $\frac{1}{4}$ "-OD ports. Stainless steel tubing ($\frac{1}{8}$ ") connected one port to a pump (Isco Syringe Pump Model 260D, Teledyne Technologies Incorporated) that allowed aqueous solutions to flow into the cell and the cell to be pressurized. A second port was equipped with the same tubing and a regulating valve for cell drainage. Two ports were fitted with 1" diameter, $\frac{1}{8}$ " thick circular sapphire windows (MSW100/125, Meller Optics Incorporated) held in place by hex nuts and sealed with rubber washers, allowing $\frac{3}{4}$ " of each window to be visible outside the cell. The remaining two ports were sealed with Swagelok plugs (SS-1610-P).

Three different optical arrangements for focusing the laser pulses into the cell and for collection of the plasma emission were used in these experiments, as detailed in Figure 2-2. Because the purpose of these experiments was initial investigation into the feasibility of using LIBS for ocean applications, one of the goals was determining the best optical set-up. For dual pulse operation, the lasers were collinear. In some single pulse configurations, light collection was collinear to the laser pulse, while in others it was orthogonal for ease of alignment. All optics were mounted on micrometer stages,




Figure 2-1: Schematic of the laboratory LIBS apparatus. Note that in the drawing, the laser pulses are simply represented by arrows as their optical paths are described in Figure 2-2.




Figure 2-2: Optical arrangements used in experiments showing the high pressure cell with respect to incoming laser pulses (signified by a dashed line). FO = Optical Fiber. (a) L_1 , L_2 , and $L_3 = f/4$ lenses; M_1 = dielectric coated mirror, (b) $L_1=f/4$, To study the effect of NaCl concentration on spectra: $L_2 = f/3$ lens, $L_3 = f/2$ lens. To study the detection of Ca at varying concentrations: $L_2 = f/4$ lens, $L_3 = f/3$ lens, (c) $L_1=f/4$, M_1 and M_2 = parabolic off-axis mirrors

enabling precise control of beam overlap and collection field of view within the high pressure cell. All lenses were made of fused silica. In all optical configurations, the plasma emission was focused onto a 2-mm-core-diameter, 0.51-N.A. light guide (Edmund Scientific Co. Model 02551). The light guide was connected to a 0.25-m, $f/4$ spectrograph (Chromex model 250is/RF) with a 1200-groove/mm grating blazed at 500 nm. The slit width (W) ranged from 25 - 250 μm . Data were collected on an intensified CCD detector (Princeton Instruments, I-Max 1024E) and acquired with a computer running WinSpec/32 software. All spectra were accumulations of 250 shots at the maximum gain setting of 255. All error bars represent $\pm 1\sigma$. A similar apparatus and set-up was previously used to demonstrate the feasibility of high pressure LIBS [35].

The key LIBS timing parameters have been previously described [16, 32]. The first and second laser pulse energies are referred to as E_1 and E_2 . For dual pulse experiments, the time interval between the two pulses or interpulse delay, is referred to as ΔT . The gate delay, t_d , is the time between the last laser pulse and turn-on of the detector. The plasma emission is recorded by the detector for the length of time set by the gate width, t_b , which was set at 1 μs for all experiments reported here.

Laser beam waist width d_{σ_o} can be estimated by:

$$d_{\sigma_o} = \frac{4f\lambda M^2}{\pi D} \quad (2.1)$$

where f is the focal length of the focusing lens (100 mm), λ is the laser wavelength (1064 nm), M^2 is the beam propagation ratio which is typically 2 - 10 for Nd:YAG lasers (we use a value of 6), and D is the diameter of the illuminated aperture of the focusing lens (≈ 25 mm) [42]. The beam waist width for the system is approximately 0.03 mm. The average irradiance (I_f) at the beam waist can be estimated using:

$$I_f = \frac{\pi E_L D^2}{4\tau_L f^2 \lambda^2 M^4} \quad (2.2)$$

where E_L is the laser pulse energy and τ_L is the pulse duration at the full peak width at half of the maximum intensity (FWHM)[42] (for the Continuum laser, $\tau_L = 5$ ns, and for the Quantel laser, $\tau_L = 9$ ns). The pulse energies of the Continuum laser used vary between ≈ 10 - 100 mJ. The irradiance of the beam at the beam waist thus varies from $\approx 2.4 \times 10^{11}$ to $\approx 2.4 \times 10^{12}$ W/cm². The pulse energies of the Quantel laser used vary between ≈ 10 - 125 mJ with the irradiance thus varying from 1.3×10^{11} to 1.7×10^{12} W/cm².


Sample solutions were made by dissolving NaCl, CaCl₂, LiCl, and MnSO₄·H₂O in deionized water. Where noted, NaCl was added to the solutions to simulate a seawater environment. All concentrations are listed in parts per million (ppm wt./vol.).

2.4 Results and Discussion


2.4.1 The Effect of Pulse Energy on LIBS Emission

Single-Pulse LIBS

Two key constraints on the design of an oceanographic sensor system are instrument power consumption and form factor, both of which must be minimized. LIBS operation with a small, low power laser would simplify the design of an oceanographic LIBS instrument. The effect of pulse energy on signal intensity for analytes in solution at elevated pressure was investigated with the goal of minimizing power consumption. The peak signal intensity for four analytes (Li, Ca, Na, and Mn) was measured at laser pulse energies ranging from 11 to 91 mJ at both low (7×10^5 Pa) and high (2.76×10^7 Pa) pressures using the collinear optical configuration shown in Figure 2-2(a)



(a) Data taken at 7×10^5 Pa (\bullet) and 2.76×10^7 Pa (\triangle).



(b) Na(I) spectra taken at 2.76×10^7 Pa. Spectra offset for clarity.

Figure 2-3: Effect of laser pulse energy on the LIBS signal intensity of 100 ppm Na(I) (588.995 nm).

($t_d = 350$ ns, $W = 75$ μ m for Na, Mn, and Ca studies, and $W = 250$ nm for Li). Ten spectra were recorded and averaged for each condition.


Figure 2-3 shows the dependence of the Na(I) (588.995 nm) emission line on laser pulse energy for 100 ppm Na. In both low and high pressure experiments, as pulse energy increases, a corresponding increase in peak intensity occurs until a maximum intensity is reached at 22 mJ (Figure 2-3(a)). Above this value, emission intensity decreases sharply up to ≈ 50 mJ, above which a more gradual decrease with energy is observed. These data suggest that, independent of pressure, a low laser pulse energy yields greater emission intensity providing the energy exceeds a threshold value. Figure 2-3(b) compares spectra taken at laser pulse energies below, above, and in the optimal energy range for Na. The top trace (22 mJ) shows a significantly greater intensity than at either a very low (middle trace, 11 mJ) or a high (bottom trace, 88 mJ) pulse energy.

The effect of laser pulse energy on Ca (422.673 nm) and Li (670.776 nm and 670.791 nm, unresolved doublet) emission displayed similar trends. When less than 14 mJ was used, Ca was virtually undetectable. As the pulse energy was increased above this level, emission intensified until a maximum was achieved at 36 mJ for low pressure (7×10^5 Pa) and at 29 mJ for high pressure (2.76×10^7 Pa). This range for both the low and high pressure environments was ≈ 25 to 50 mJ. At energy levels beyond the optimal range, intensity decreased slowly with increasing pulse energy, possibly due to plasma shielding. Plasma shielding occurs when the plasma itself reduces the transmission of the laser pulse energy along the beam path. Calcium displayed a more gradual increase and then decrease in intensity and a wider range of optimal energy compared to Na. Similar trends were observed for Li. At both low and high pressures, plasma emission was not detectable below 11 mJ. At higher pulse energies and both pressures, the emission maximum was recorded at 27 mJ, above which a sharp decrease in intensity to 46 mJ was observed, followed by flattening to 72 mJ.


The relationship between emission intensity and laser pulse energy for the unresolved 403 nm Mn(I) triplet was slightly different than for the other three analytes. Figure 2-4(a) shows that the lowest laser pulse energy (11 mJ) resulted in the highest emission intensity. At pulse energies greater than 11 mJ, the emission intensity gradually decreased until it was no longer detectable above ≈ 40 mJ and ≈ 70 mJ for low and high pressures, respectively. The peak intensity was greater at high than at low pressure. Figure 2-4(b) compares spectra taken at 11, 22, and 88 mJ at 2.76×10^7 Pa.

The data for Na, Ca, Li, and Mn suggest that the pulse energy required to optimize the LIBS signal is analyte-dependent due to different ionization energies, but is minimally pressure dependent. A pulse energy threshold is also observed. For the four analytes studied, a relatively low laser pulse energy (less than 50 mJ) produced the greatest signal intensity. A low energy optimal range may exist due to effects from plasma shielding or moving breakdown. Plasmas can expand back along the laser beam path towards the laser resulting in elongated plasmas [43]. A higher energy pulse may form a more elongate plasma or a series of plasmas as the breakdown threshold of the liquid is exceeded before the pulse reaches the focal point. This may result in non-optimal collection of the plasma emission. Further studies using imaging techniques are needed to elucidate the effect of pulse energy on the plasma.

Figure 2-5 shows the effect of pressure on the LIBS signal for Na (588.995 nm), Mn (403 nm unresolvable triplet) and Ca (422.673 nm) using a low energy single pulse.



(a) Data taken at 7×10^5 Pa (●) and 2.76×10^7 Pa (△)



(b) Mn(I) spectra taken at 2.76×10^7 Pa. Spectra offset for clarity

Figure 2-4: Effect of laser pulse energy on the LIBS emission intensity of the unresolvable Mn (I) triplet (403 nm) (5,000 ppm Mn in 2,540 ppm NaCl).




Figure 2-5: Effect of pressure on LIBS emission intensity. \square = 100 ppm Na (588.995 nm) with $E = 22$ mJ; \bullet = 5,000 ppm Mn (403 nm unresolvable triplet) with 2,540 ppm NaCl, $E = 14$ mJ; \triangle = 500 ppm Ca (422.673 nm) with 2,540 ppm NaCl, $E = 20$ mJ.

The gate delay was fixed at 350 ns and the slit width was fixed at $75\mu\text{m}$. Na and Ca display no change in signal intensity with increasing pressure, but Mn shows an increase. For all analytes examined, the FWHM did not change with pressure. Pressure under oceanic conditions does not induce a deleterious effect on signal intensity or on FWHM.

In these single pulse energy experiments, the same gate delay and gate width were used for all energy levels and pressures. As discussed later in this paper, optimal gate delay may be energy dependent. Optimal gate width was not investigated and may be pressure and/or energy level dependent. As a result, the selected gate width and gate delay may influence the measured emission intensity. Optimal gate delay could also be analyte dependent, and hence a different gate delay could yield another trend with pulse energy. However, the selected conditions demonstrate that low energy single laser pulses at high pressures are viable for measuring analytes in bulk aqueous liquids. This is promising toward the development of an ocean-going instrument where a small, low power laser will be critical.

Dual Pulse LIBS at High Pressure

An evaluation of dual pulse LIBS for high pressure bulk solutions shows that analyte detection is highly dependent on the interpulse delay. If the interpulse delay is short ($\ll 1\mu\text{s}$), signal intensity is greatly enhanced when compared to that measured using longer delay times. However, such a small interpulse delay may not be sufficient for a cavitation bubble to fully form before the second laser pulse creates a spark.

Dual pulse LIBS has been shown to enhance the signal intensity for some analytes in bulk aqueous solutions at atmospheric pressure [30, 32]. However, such enhancements using longer interpulse delay times do not occur for high pressure liquids.

To demonstrate the coupled effect of interpulse delay and pulse energy on emission intensity, four energy level conditions were compared for four analytes at high pressure (2.76×10^7 Pa) over a range of interpulse delay times. The four conditions were: 1) low E_1 , low E_2 (low-low), 2) high E_1 , high E_2 (high-high), 3) low E_1 , high E_2 (low-high), and 4) high E_1 , low E_2 (high-low), and are detailed in Table 2.1, ($t_d = 350$ ns, $t_b = 1$ μ s). These experiments were completed using the optical configuration shown in Figure 2-2(a).

For Ca ($W = 100\mu\text{m}$), using a low E_1 followed by a low E_2 resulted in the highest peak intensity, possibly because when summed they give a low total energy (Figure 2-6). The greatest emission is observed for $E_1 = 13$ mJ and $E_2 = 6$ mJ and yields the ionic Ca peaks (393.366 nm and 396.847 nm in addition to the atomic peak 422.673 nm). For this low-low condition, Figure 2-7(a) shows the emission intensity change with ΔT . For ΔT greater than 1 μ s, the intensity remained stable at a value of 1.5×10^4 a.u.

For ΔT less than 1 μ s, the low-low configuration yielded intensities between 2.5×10^4 and 8.7×10^4 a.u. Figure 2-7(b) compares spectra at very short (30 ns - upper trace) to long (30 μ s - lower trace) ΔT values. When a short ΔT is used, three Ca peaks (Ca(II) 393.366 nm, Ca(II) 396.847 nm, and Ca(I) 422.673 nm) are visible, while for long ΔT , only the Ca(I) peak is present with a much lower intensity. When ΔT is 30 ns, the low-low configuration yields significantly greater emission intensity

Table 2.1: Conditions used to study the effect of dual pulse energies on LIBS emission

	Low E_1 Low E_2		High E_1 High E_2		Low E_1 High E_2		High E_1 Low E_2	
	E_1 (mJ)	E_2 (mJ)	E_1 (mJ)	E_2 (mJ)	E_1 (mJ)	E_2 (mJ)	E_1 (mJ)	E_2 (mJ)
1,000 ppm Ca, 2,540 ppm NaCl	13	6	105	84	13	84	105	6
100 ppm Li	31	20	105	84	31	84	105	20
100 ppm Na	13	6	105	84	13	84	105	6
5,000 ppm Mn, 2,540 ppm NaCl	13	6	105	84	13	84	105	6




Figure 2-6: Spectra of 1,000 ppm Ca with 2,540 ppm NaCl at 2.76×10^7 Pa under four dual pulse conditions.


than the other pulse energy configurations. For Li ($W = 250 \mu\text{m}$), a low E_1 followed by a low E_2 resulted in the greatest emission intensity. Table 2.2 shows peak emission for Li for four different dual pulse conditions for ΔT between 50 ns and 1 μs .

A small delay time ($< 1 \mu\text{s}$) enhanced the emission as compared to a longer delay time when the low-high and low-low energy levels were used. For Na(I), the low-high and low-low conditions yielded similar intensities at all delay times, with maximum values of 9.3×10^5 a.u. and 8.3×10^5 a.u., respectively, ($W = 75 \mu\text{m}$). After these four runs were compared, an additional configuration consisting of a 13 mJ first pulse followed by a 22 mJ second pulse was tested as a low-low dual pulse condition with a slightly increased second pulse energy. This resulted in peaks with intensities of 2.4×10^6 - 2.9×10^6 a.u. for all ΔT values between 10 ns and 100 μs , suggesting again that a low-low energy condition produces the greatest emission intensity. For Mn ($W = 250 \mu\text{m}$), at all interpulse delay times between 20 ns and 100 μs , a low E_1 followed by a high E_2 , resulted in the highest emission intensity (Figure 2-8).


These results show that the best dual pulse conditions vary by analyte. However,

Table 2.2: Dual pulse emission intensity (a.u.)

	Condition 1	Condition 2	Condition 3	Condition 4
	Low E_1 Low E_2	High E_1 High E_2	Low E_1 High E_2	High E_1 Low E_2
100 ppm Li	2.5×10^5 - 3.7×10^5	2×10^4 - 7×10^4	5×10^4 - 1.5×10^5	1.5×10^4 - 5.3×10^4
5,000 ppm Mn, 2,540 ppm NaCl	4×10^3 - 4.2×10^5	2.1×10^5 - 5×10^5	7×10^5 - 8.3×10^5	1.6×10^3 - 2.6×10^5



(a) Effect of dual laser pulse energies on emission intensity at 2.76×10^7 Pa for 1,000 ppm Ca in 2,540 ppm NaCl at various interpulse delays. Each data point is the average of 5 spectra.



(b) Spectra of Ca showing the enhancement in signal for $\Delta T = 30$ ns over $\Delta T = 30$ μ s.

Figure 2-7: Ca emission at different interpulse delays. ($E_1 = 13$ mJ, $E_2 = 6$ mJ.)




Figure 2-8: Spectra of 5,000 ppm Mn with 2,540 ppm NaCl at 2.76×10^7 Pa under four dual pulse conditions. The highest emission intensity is observed for a low-high pulse combination.


it is important to note that at high pressure, very short interpulse spacing results in a higher signal intensity than when dual pulses are separated by a more significant delay in time. The need for such rapid firing of the two pulses is only accomplished using two independent lasers instead of firing one laser rapidly. Two pulses separated by a short ΔT approaches single pulse conditions, suggesting that dual pulse LIBS may not be advantageous at elevated pressure.

2.4.2 Interrelationship of pulse energy, gate delay, and pressure for Lithium


Emission intensity was recorded for the unresolved Li(I) doublet (670.776 nm and 670.791 nm) at two single pulse energies (27 mJ and 68 mJ) at both low (7×10^5 Pa) and high pressure (2.76×10^7 Pa) over a range of gate delays (0.1 - 3.7 μ s), ($t_b = 1 \mu$ s, $W = 25 \mu$ m) using the optical configuration of Figure 2-2(a). Comparing the two curves in Figure 2-9, it is clear that a short gate delay should be used to enhance emission intensity. These results also suggest that the optimal gate delay may be pulse energy but not pressure dependent.

2.4.3 Effect of NaCl Concentration on LIBS Spectra

Understanding how pervasive Na and Cl ions in solution affect the detection of other analytes is important for assessing the feasibility of using LIBS in the ocean, where



(a) Data taken with a single low energy pulse (27 mJ)



(b) Data taken with a single high energy pulse (68 mJ).

Figure 2-9: Effect of gate delay on the LIBS signal for 1,000 ppm Li (670 nm unresolved doublet). $\bullet = 7 \times 10^{-5}$ Pa, $\triangle = 2.57 \times 10^7$ Pa.




Figure 2-10: Effect of the addition of NaCl in solution on spectra of 1,000 ppm Ca at 2.57×10^7 Pa.

the nominal concentrations of Na and Cl are 1.08×10^4 ppm and 1.95×10^4 ppm, respectively [44]. Cremers et al. previously reported a decrease in the intensity ratio of Ca II/Ca I with the addition of NaCl [30]. The peak signal intensity for three analytes (1,000 ppm Ca, 100 ppm Mn, and 1,000 ppm K) was compared in three solutes: 1) deionized water, 2) 2,540 ppm NaCl dissolved in deionized water, and 3) 25,400 ppm NaCl dissolved in deionized water using the optical configuration of Figure 2-2(b) and for a range of pressures (3×10^5 , 7×10^5 , 1.7×10^6 , 3.4×10^6 , 6.9×10^6 , 1.38×10^7 , 2.07×10^7 , 2.76×10^7 Pa). These studies were carried out with $E_1 = 40$ mJ, $E_2 = 125$ mJ, $\Delta T = 46$ ns, $W = 35$ μ m, and $t_d = 100$ ns for Ca and K and $t_d = 200$ ns for Mn. The addition of NaCl significantly increased the emission intensity of the 422.673 nm Ca(I) atomic line whereas no significant effect was seen on the 393.366 nm and 396.847 nm Ca(II) ionic lines (Figure 2-10). The signal:noise ratio for the Ca (II) ionic lines showed no significant change with the addition of NaCl, whereas the signal:noise ratio of Ca(I) increased from 22 to 30 with the addition of NaCl. The same increase was seen with the addition of 254 ppm NaCl and 25,400 ppm NaCl. In atomic emission, adding an easily ionizable element, for example Na, can suppress ionization of other elements. This suggests that ionization suppression by Na increases the Ca(I) emission relative to the Ca(II) lines. No intensity change was seen for Mn(I) (403 nm unresolvable triplet) or K(I) (766.491 nm and 769.897 nm). However, since only atomic lines were detectable for Mn and K, the relative increase of atomic to ionic lines could not be compared.

These two outcomes (enhancement of the signal or no change to the signal) suggest that the high NaCl concentration in the ocean will not have a deleterious effect on the ability to detect Ca, Mn, and K analytes. It also suggests that further work is




Figure 2-11: Detection of Ca (422.673 nm) in a simulated vent fluid at varying pressures and concentrations. ● = 7×10^5 Pa, △ = 7×10^6 Pa, □ = 2.76×10^7 Pa.

needed to elucidate the effect NaCl has on other analytes.

2.4.4 Detection of Calcium at Varying Concentrations

Ca was used to determine whether increased pressure affects the limit of detection. Five pressures ranging from 7×10^5 Pa to 2.76×10^7 Pa were investigated at concentrations ranging from 10 to 5,000 ppm in a solution containing several analytes (69 ppm Br, 10,828 ppm Na, 89 ppm Fe, 958 ppm K, 46 ppm Mn, 18,932 ppm Cl in DI water) found in hydrothermal vent fluids at representative concentrations (10 ppm, 25 ppm, 50 ppm, 100 ppm, 500 ppm, 1000 ppm, 5000 ppm), using the optical configuration of Figure 2-2(b) ($E_1 = 31$ mJ; $E_2 = 15$ mJ; $\Delta T = 72$ ns; $t_d = 700$ ns; $t_d = 1$ μ s, = 35 μ m). Figure 2-11 shows that varying concentrations of Ca are detectable at pressure and with a minimal change in intensity, suggesting that detection limits to the ppm level will be obtainable at high pressure.

2.4.5 Solution Temperature Effects on Calcium Spectra

To characterize the temperature effect for Ca spectra, the sample cell was placed in a sand bath heated by a hot plate. The drainage port was removed and a thermocouple was inserted to record the temperature of the aqueous solution. We investigated the effect of temperature on the peak intensity of Ca(I) (422.673 nm) over the range 27 - 99°C. Once the solution reached 99°C, the hot plate was turned off and allowed to cool. Spectra were taken repeatedly using single-pulse LIBS as the temperature dropped. Ca line intensities were measured for a solution of 1,000 ppm Ca and 2,540

ppm NaCl at atmospheric pressure using a single laser pulse of 37 mJ, ($t_d = 100$ ns, $t_b = 1$ μ s, $W = 35$ μ m), and the optical configuration shown in Figure 2-2(c). Over this range, temperature had no effect on peak intensity.

2.5 Conclusions

An optimal range of low laser pulse energies exists for the detection of Li, Ca, Mn, K, and Na in bulk aqueous solutions at both low and high pressures. No pressure effect was seen on the emission intensity for Ca and Na, and an increase in intensity with increased pressure was seen for Mn. No line broadening due to pressure was observed for Ca, Na, or Mn emission. A low energy pulse may create a smaller, more tightly focused plasma that forms only at the focal spot. However, for a high energy pulse, the high energy density may cause breakdown even before the pulse reaches the focal spot. This may allow breakdown to occur over a longer distance. In addition, plasma shielding may occur. Further studies using imaging techniques will help to elucidate the relationship between the laser pulse energy and the subsequent plasma that is formed. Using the dual pulse technique for several analytes, a very short interpulse delay resulted in the greatest emission intensity. Since this condition approaches single pulse conditions, dual pulse LIBS may not be advantageous for some elements at high pressure. For different gate delays at fixed pressure, laser pulse energy affects peak intensity. The addition of NaCl enhanced the emission intensity for Ca, but had no effect on the intensity of Mn or K peaks. Ca was detectable over a wide range of concentrations and pressures. In addition, temperature changes below 99°C had no noticeable effect on the emission intensity of Ca. Overall, increased pressure, the addition of NaCl to a solution, and temperature did not inhibit detection of analytes in solution. The results presented here suggest that LIBS is a viable technique for *in situ* chemical analysis in the deep ocean and further work should be carried out to develop LIBS into an *in situ* oceanographic sensor.

2.6 Acknowledgments

We acknowledge the National Science Foundation for support of this research under grants OCE0352278 and OCE0352242. Additional support was received from the Deep Ocean Exploration Institute of the Woods Hole Oceanographic Institution.

Bibliography

- [1] F. Brech and L. Cross. Optical microemission stimulated by a a ruby maser. *Applied Spectroscopy*, 16:59, 1962.
- [2] V. Majidi and M. R. Joseph. Spectroscopic applications of laser-induced plasmas. *Critical Reviews in Analytical Chemistry*, 23:143–162, 1992.
- [3] L. J. Radziemski. Review of analytical applications of laser plasmas and laser ablation, 1987-1994. *Microchemical Journal*, 50:218–234, 1994.
- [4] D. A. Rusak, B. C. Castle, B. W. Smith, and J. D. Winefordner. Fundamentals and applications of laser-induced breakdown spectroscopy. *Critical Reviews in Analytical Chemistry*, 27:257–290, 1997.
- [5] K. Song, Y. Lee, and J. Sneddon. Applications of laser-induced breakdown spectrometry. *Applied Spectroscopy Reviews*, 32:183–235, 1997.
- [6] J. Sneddon and Y. Lee. Novel and recent applications of elemental determination by laser-induced breakdown spectroscopy. *Analytical Letters*, 32:2143–2162, 1999.
- [7] D. A. Rusak, B. C. Castle, B. W. Smith, and J. D. Winefordner. Recent trends and the future of laser-induced plasma spectroscopy. *Trends in Analytical Chemistry*, 17:453–461, 1998.
- [8] W. Lee, J. Wu, Y. Lee, and J. Sneddon. Recent applications of laser-induced breakdown spectrometry: a review of material approaches. *Applied Spectroscopy Reviews*, 39:27–97, 2004.
- [9] Z. A. Arp, D. A. Cremers, R. D. Harris, D. M. Oschwald, G. R. Parker Jr., and D. M. Wayne. Feasibility of generating a useful laser-induced breakdown spectroscopy plasma on rocks at high pressure: preliminary study for a Venus mission. *Spectrochimica Acta Part B*, 59:987–999, 2004.
- [10] B. Salle, J. Lacour, E. Vors, P. Fichet, S. Maurice, D. A. Cremers, and R. C. Wiens. Laser-induced breakdown spectroscopy for Mars surface analysis: capabilities at stand-off distances and detection of chlorine and sulfur elements. *Spectrochim. Acta*, 59:1413–1422, 2004.
- [11] W. E. Seyfried, Jr., K. S. Johnson, and M. K. Tivey, editors. *In-Situ Sensors: Their Development and Application for the Study of Chemical, Physical and Biological Systems at Mid-Ocean Ridges NSF/RIDGE-Sponsored Workshop*, October 2000.
- [12] The next generation of *in situ* biological and chemical sensors in the ocean: a workshop report. 2004.

- [13] K. L. Daly, R. H. Byrne, A. G. Dickson, S. M. Gallagher, M. J. Perry, and M. K. Tivey. Chemical and biological sensors for time-series research: current status and new directions. *Journal of the Marine Technology Society*, 38:121–143, 2004.
- [14] T. Dickey. The role of new technology in advancing ocean biogeochemical studies. *Oceanography*, 14:108–120, 2001.
- [15] M. S. Varney, editor. *Chemical Sensors in Oceanography*. Gordon and Breach, 2000.
- [16] A. E. Pichahchy, D. A. Cremers, and M. J. Ferris. Elemental analysis of metals under water using laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 52:25–39, 1997.
- [17] P. Fichet, D. Menut, R. Brennetot, E. Vors, and A. Rivoallan. Analysis by laser-induced breakdown spectroscopy of complex solids, liquids, and powders with an Echelle spectrometer. *Applied Optics*, 42:6029–6039, 2003.
- [18] P. Fichet, P. Mauchien, J.-F. Wagner, and C. Moulin. Quantitative elemental determination in water and oil by laser induced breakdown spectroscopy. *Analytica Chimica Acta*, 429:269–278, 2001.
- [19] G. Arca, A. Ciucci, V. Palleschi, S. Rastelli, and E. Tognoni. Trace element analysis in water by the laser-induced breakdown spectroscopy technique. *Applied Spectroscopy*, 51:1102–1105, 1997.
- [20] O. Samek, D. C. S. Beddows, J. Kaiser, S. V. Kukhlevsky, M. Liska, H. H. Telle, and J. Young. Application of laser-induced breakdown spectroscopy to *in situ* analysis of liquid samples. *Optical Engineering*, 39:2248–2262, 2000.
- [21] J. R. Wachter and D. A. Cremers. Determination of uranium in solution using laser-induced breakdown spectroscopy. *Applied Spectroscopy*, 41:1042–1048, 1987.
- [22] A. Kuwako, Y. Uchida, and K. Maeda. Supersensitive detection of sodium in water with use of dual-pulse laser-induced breakdown spectroscopy. *Applied Optics*, 42:6052–6056, 2003.
- [23] V. N. Rai, F. Y. Yueh, and J. P. Singh. Study of laser-induced breakdown emission from liquid under double pulse excitation. *Applied Optics*, 42:2094–2101, 2003.
- [24] X. Y. Pu and N. H. Cheung. ArF laser induced plasma spectroscopy of lead ions in aqueous solutions: Plume reheating with a second Nd:YAG laser pulse. *Applied Spectroscopy*, 57:588–590, 1997.
- [25] W. F. Ho, C. W. Ng, and N. H. Cheung. Spectrochemical analysis of liquids using laser-induced plasma emissions: effect of laser wavelength. *Applied Spectroscopy*, 51:87–91, 1997.

- [26] S. Nakamura, Y. Ito, and K. Sone. Determination of an iron suspension in water by laser-induced breakdown spectroscopy with two sequential laser pulses. *Analytical Chemistry*, 68:2981–2986, 1996.
- [27] K. M. Lo and N. H. Cheung. ArF laser-induced plasma spectroscopy for part-per-billion analysis of metal ions in aqueous solutions. *Applied Spectroscopy*, 56:682–688, 2002.
- [28] L. St-Onge, E. Kwong, M. Sabsabi, and E. B. Vadas. Rapid analysis of liquid formulations containing sodium chloride using laser-induced breakdown spectroscopy. *Journal of Pharmaceutical and Biomedical Analysis*, 36:277–284, 2004.
- [29] J. Huang, C. Ke, and K. Lin. Matrix effect on emission/current correlated analysis in laser-induced breakdown spectroscopy of liquid droplets. *Spectrochimica Acta Part B*, 59:321–326, 2004.
- [30] D. A. Cremers, L. J. Radziemski, and T. R. Loree. Spectrochemical analysis of liquids using the laser spark. *Applied Spectroscopy*, 38:721–729, 1984.
- [31] R. Knopp, F. J. Scherbaum, and J. I. Kim. Laser induced breakdown spectroscopy (LIBS) as an analytical tool for the detection of metal ions in aqueous solutions. *Fresenius’ Journal of Analytical Chemistry*, 355:16–20, 1996.
- [32] W. Pearman, J. Scaffidi, and S. M. Angel. Dual-pulse laser-induced breakdown spectroscopy in bulk aqueous solution with an orthogonal beam geometry. *Applied Optics*, 42:6085–6093, 2003.
- [33] C. Haisch, J. Liermann, U. Panne, and R. Niessner. Characterization of colloidal particles by laser-induced plasma spectroscopy (LIPS). *Analytica Chimica Acta*, 346:23–25, 1997.
- [34] M. Noda, Y. Deguchi, S. Iwasaki, and N. Yoshikawa. Detection of carbon in a high-temperature and high-pressure environment using laser-induced breakdown spectroscopy. *Spectrochimica Acta*, 57:701–709, 2002.
- [35] M. Lawrence-Snyder, J. Scaffidi, S. M. Angel, A. P. M. Michel, and A. D. Chave. Laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, 60:786–790, 2006.
- [36] C. Aragon, J. A. Aguilers, and J. Campos. Determination of carbon content in molten steel using laser-induced breakdown spectroscopy. *Applied Spectroscopy*, 47:606–608, 1993.
- [37] J. I. Yun, R. Klenze, and J. I. Kim. Laser-induced breakdown spectroscopy for the on-line multielement analysis of highly radioactive glass melt simulants: Part II analyses of molten glass samples. *Applied Spectroscopy*, 56:852–858, 2002.
- [38] A. K. Rai, F. Y. Yueh, and J. P. Singh. Laser-induced breakdown spectroscopy of molten aluminum alloy. *Applied Optics*, 42:2078–2084, 2003.

- [39] L. G. Blevins, C. R. Shaddix, S. M. Sickafoose, and P. M. Walsh. Laser-induced breakdown spectroscopy at high temperatures in industrial boilers and furnaces. *Applied Optics*, 42:6107–6118, 2003.
- [40] K. L. Von Damm. Chemistry of hydrothermal vent fluids from 9° - 10° N, East Pacific Rise: ‘Time zero,’ The immediate posteruptive period. *Journal of Geophysical Research*, 105:11203–11222, 2000.
- [41] D. A. Butterfield, I. R. Jonasson, G. J. Massoth, R. A. Feely, K. K. Roe, R. E. Embley, J. F. Holden, R. E McDuff, M. D. Lilley, and J. R. Delaney. Seafloor eruptions and evolution of hydrothermal fluid chemistry. *Philosophical Transactions: Mathematical, Physical and Engineering Sciences*, 355, 1997.
- [42] R. Noll. Terms and notations for laser-induced breakdown spectroscopy. *Analytical and Bioanalytical Chemistry*, V385(2):214–218, 2006.
- [43] P. K. Kennedy, D. X. Hammer, and B. A. Rockwell. Laser-induced breakdown in aqueous media. *Progress in Quantum Electronics*, 21:155–248, 1997.
- [44] J. Brown, A. Colling, D. Park, J. Phillips, D. Rothery, and J. Wright. *Seawater: its composition, properties and behaviour*. The Open University, Milton Keynes, England, 1989.

Chapter 3

Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics

The work in this chapter is currently in press in *Spectrochimica Acta Part B* as, A. P. M. Michel and A. D. Chave, “Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics.”

3.1 Abstract

In most instances, laser-induced breakdown spectroscopy (LIBS) spectra are obtained through analog accumulation of multiple shots in the spectrometer CCD. The average acquired in the CCD at a given wavelength is assumed to be a good representation of the population mean, which in turn is implicitly regarded to be the best estimator for the central value of the distribution of the spectrum at the same wavelength. Multiple analog accumulated spectra are taken and then in turn averaged wavelength-by-wavelength to represent the final spectrum. In this paper, the statistics of single-shot and analog accumulated LIBS spectra of both solids and liquids were examined to evaluate whether the typically used spectrum averaging approach is statistically defensible. At a given wavelength, LIBS spectra are typically drawn from a Frechet extreme value distribution, and hence the mean of an ensemble of LIBS spectra is not necessarily an optimal summary statistic. Under circumstances

that are broadly general, the sample mean for LIBS data is statistically inconsistent and the central limit theorem does not apply. This result appears to be due to very high shot-to-shot plasma variability in which a very small number of spectra are high in intensity while the majority are very weak, yielding the extreme value form of the distribution. The extreme value behavior persists when individual shots are analog accumulated. An optimal estimator in a well-defined sense for the spectral average at a given wavelength follows from the maximum likelihood method for the extreme value distribution. Example spectra taken with both an Echelle and a Czerny-Turner spectrometer are processed with this scheme to create smooth, high signal-to-noise summary spectra. Plasma imaging was used in an attempt to visually understand the observed variability and to validate the use of extreme value statistics. The data processing approach presented in this paper is statistically reliable and should be used for accurate comparisons of LIBS spectra instead of arithmetic averaging on either complete or censored data sets.

3.2 Introduction

Laser-induced breakdown spectroscopy (LIBS) is a spectrochemical technique that has been successfully used for elemental analysis of solids, liquids, gases, and aerosols, and is finding increasing application in basic and applied research. However, one of the major problems that precludes more quantitative use is a lack of reproducibility of spectra at a given wavelength on a shot-to-shot basis.

Dramatic peak intensity fluctuations at the shot-to-shot level, suggesting the presence of high random variability, has been noted by many LIBS researchers [1–15]. There are numerous potential causes for this, including repetitive laser pulse instability, unstable laser pulse characteristics, laser pulse-plasma interactions, lens-to-sample distance variation (which in turn changes the distance from the plasma to the collection fiber), laser-material coupling, variable sample ablation, plasma position instability, matrix effects, perturbations of the plasma due to physical and chemical characteristics of the sample (i.e., composition, homogeneity, roughness, color, and moisture content), scattering of light, atmospheric conditions, weak ionization of the plasma, and non-optimal collection of plasma emission [1, 3–11, 16]. Optical instability affects the ablation process, the plasma profile, and the plasma volume [7]. Carranza and Hahn [16] suggest that above a threshold value, absorption of pulse energy by the plasma saturates, reducing variability at higher levels. A given plasma is not completely homogeneous, and there may be property gradients due to boundary

effects and its transient nature. Spatial variation in the position of the plasma changes the coupling of the plasma light into the collection optics. Carranza and Hahn further suggest that shot-to-shot variation may be reduced by using sufficient laser pulse energy to achieve saturation and a suitable collection geometry (backscatter mode) to minimize spatial variability. The intensity of the laser itself can fluctuate by 1-5% [8]. However, Castle *et al.* simultaneously measured the analyte signal and the laser pulse energy, and found no significant correlation, suggesting laser pulse variance has only a minor influence on overall variability [13].

The sample type also influences variability. For aqueous samples, additional fluctuation can be caused by “moving breakdown” that changes the distance between the spark and the collection fiber as the plasma moves in the solution. The plasma typically expands along the beam path toward the laser, inducing elongate plasmas that cavitate radially [17]. Variability in aqueous solutions can also be caused by suspended ablated particles [2, 18] and bubbles [19] both by reducing the energy delivered to a sample and the light transmitted to the collection optics. Bubbles, formed when breakdown occurs, and dissolved gases can scatter or absorb incident laser radiation [17, 19].

Lazic *et al.* [2] report high variability of the plasma intensity for both aqueous solutions (including bulk water) and solid samples immersed in water. Significant variability was not observed when high laser pulse energies were used to measure the elemental composition of flat homogeneous solid samples. In aqueous solution, LIBS emission was sometimes not detectable even when the maximum laser energy was used. The lack of emission was also observed for rough inhomogeneous solids. For flat samples, the only time no breakdown occurred was when low laser energies were used; yet, shot-to-shot signal variability was always present. Lazic *et al.* [10] reported peak intensity histograms. The distributions of these data sets are clearly not Gaussian, and show that a very high intensity peak is a rare event, with very low intensity occurring for the bulk of the trials.

For a solid sample, inhomogeneity, porosity, or surface roughness can change the distance between the focusing optics and the sample, either from prior crater formation or by changing the location of ablation. Panne *et al.* [12] report significant pulse-to-pulse variation of the plasma electronic excitation temperature and electron density from material-laser interaction for homogeneous glass samples.

Laser ablation is highly nonlinear, and even more so in aerosol samples as the plasma may form at different positions along the beam [8]. For aerosols, the location of the particles within the plasma volume and the focal volume of the optics con-

tribute to variability [14]. Schechter’s [15] analysis of spectral fluctuations of aerosols showed large shot-to-shot variability possibly caused by laser pulses hitting different numbers of particles, particle characteristic variation (size, mass, and location), and location variation of the plasma. Whatever the cause of observed LIBS intensity variability, analog averaging of multiple plasma emissions, where light from numerous laser shots is accumulated on a CCD to create a single spectrum, is an often used experimental approach in order to increase the signal and the signal-to-noise ratio in the presence of shot-to-shot variability [20]. Analog averaged spectra are replicated and the ensemble of replicates are in turn averaged to create a representative spectrum. However, this implicitly assumes that the sample mean is a reasonable estimator for the statistical average, and this condition may not hold for some non-Gaussian distributions.

LIBS researchers have recognized the potential impact of intensity variability, and have devised a variety of methods to reduce the effect of spectral variability. Schechter used a rejection algorithm to eliminate anomalous spectra (e.g., spectra with no elemental lines, spectra with a too weak or too intense baseline due to laser fluctuations, and weak spectra) from the ensemble. This typically removed 75% of measured spectra [15]. Carranza and Hahn [16] used a sorting algorithm to remove irregular spectra, eliminating 60-70% of single shot data. Lazic *et al.* [2] removed spectra below a threshold value to increase the signal-to-noise ratio and make emission lines more readily visible.

In the present work, rather than using an *ad hoc* approach, the statistical variability of LIBS spectral intensity has been quantified and a data processing scheme based on the observed statistics has been devised. It will be demonstrated that LIBS intensity (whether single shot or analog averaged) typically has a Frechet extreme value distribution, and that for the characteristic range of statistical parameters, the distribution may not possess a variance. As a consequence, the sample mean is not an appropriate estimator for the average intensity, the central limit theorem does not apply, and Gaussian-based inference will be in error. Instead, a maximum likelihood estimator for the extreme value distribution is advocated as an alternative. The result is illustrated using single shot and analog averaged LIBS spectra for a solid target using one experimental set-up and for bulk aqueous solutions using two experimental set-ups.




Figure 3-1: Laboratory set-up of LIBS using an Echelle spectrometer.

3.3 Experimental

3.3.1 Echelle Spectrometer Set-up

To examine the variability of LIBS at shot-to-shot and analog averaged levels, the peak intensity was examined over the 580 - 600 nm range for Na using both solid NaCl (halite) obtained from Fisher Scientific and an aqueous NaCl solution. Solutions were made using de-ionized water and NaCl to yield a Na concentration of 100 parts per million (ppm, wt./vol.). The variability of intensity was measured for the solid sample using an Echelle spectrometer and for aqueous solutions using both Echelle and Czerny-Turner units. A dark background spectrum was initially subtracted from all raw spectra. For the halite specimen, 100 single shot and 10 shot analog averaged spectra were obtained. For the aqueous specimen, 100 single shot and 100 shot analog averaged spectra were collected.

The first experimental set-up utilizing an Echelle spectrometer (LLA Echelle ESA 3000) is shown in Figure 3-1. The spectrometer is capable of detecting elements over the 200 - 780 nm range with a spectral resolution of 10 to 50 pm. A Big Sky CFR-200 Nd:YAG laser (7.5-ns pulse width) operated at the fundamental wavelength of 1064 nm with a repetition rate of 5 Hz was used for plasma excitation. The laser is equipped with a variable attenuator controlled by a computer that allows laser pulse energy to range from 0 to 200 mJ in increments of <1 mJ. A timing box (Berkeley Nucleonics Corporation Model 565) was used to accurately control firing of the laser in relation to turn-on of the spectrometer.

For liquid samples, a cubic titanium sample chamber (8.89 cm × 8.89 cm × 8.89 cm) equipped with two sapphire windows (Meller Optics - 2.54 cm diameter x 0.64 cm thick, AR coated at 1064 nm, custom part) that allows laser pulses to enter the cell and the plasma to be imaged from the side of the cell (orthogonal to the entering laser beam) was used. AR-coated optics focus the laser beam into the chamber. For

solid samples, the chamber was removed and the final focusing lens was placed in front of the sample. Additional optics were used to focus the plasma light onto an optical fiber that delivers it to the spectrometer. Data were collected using ESAWIN software. All spectra were taken with a pulse energy of 80 mJ and the maximum MCP amplification of 4000. For aqueous NaCl solutions, the gate delay = 75 ns and gate width = 200 ns. For halite samples, the gate delay = 10,000 ns and gate width = 100 ns.

Plasma images were taken using a Pixelfly camera with a microscope lens and an iris diaphragm. The images were taken through the sapphire window on the pressure chamber, orthogonal to the incoming laser pulses (80 mJ/pulse). The shutter remained open for 5 μ s and was externally synched to the Q-switch of the laser.

3.3.2 Czerny-Turner Spectrometer Set-up

The second set-up used a Czerny-Turner spectrometer and is shown in Figure 3-2. A Continuum Surelite III laser (5-ns pulse width, 1064 nm, 1 Hz repetition rate) was used for plasma excitation with a pulse energy of 81 mJ. Laser pulses were focused into a chamber constructed of stainless steel Swagelok fittings with six 2.54 cm-ID and 3.18 cm-OD ports. Two ports were fitted with 2.54 cm diameter, 0.32 cm thick circular sapphire windows (MSW100/125, Meller Optics Incorporated) held in place by hex nuts and sealed with rubber washers, allowing 1.91 cm of each window to be visible outside the cell. The plasma emission was focused onto a 2-mm-core-diameter, 0.51-N.A. light guide (Edmund Scientific Co. Model 02551). The light guide was connected to a 0.25-m, $f/4$ spectrograph (Chromex model 250is/RF) with a 1200-groove/mm grating blazed at 500 nm. Data were collected on an intensified CCD detector (Princeton Instruments, I-Max 1024E) and acquired with a computer running WinSpec/32 software.

All spectra were taken at the maximum gain setting of 255, with gate delay = 175 ns and gate width = 200 ns. Solutions were made using de-ionized water, MnSO_4 , ZnBr_2 , and NaSO_4 . The solution contained 5000 parts per million (ppm, wt./vol.) Mn, 5000 ppm Zn, and 2000 ppm Na.




Figure 3-2: Laboratory set-up of LIBS using a Czerny-Turner spectrometer.

3.4 Results and Discussion

3.4.1 The Generalized Extreme Value Distribution

Extreme value distributions describe the stochastic behavior of the maximum or minimum of independent and identically distributed random variables drawn from some parent distribution. There are three types of extreme value distributions: Weibull, Gumbel, and Frechet. The von Mises-Jenkinson or generalized extreme value distribution (GEVD) combines the three into a single functional form [21, 22]. The probability density function for the GEVD is given by

$$f(x|k, \mu, \sigma) = \left(\frac{1}{\sigma}\right) \exp\left(-\left(1 + k \frac{(x - \mu)}{\sigma}\right)^{-\frac{1}{k}}\right) \left(1 + k \frac{(x - \mu)}{\sigma}\right)^{\left(-1 - \frac{1}{k}\right)} \quad (3.1)$$

where the distribution is Weibull, Gumbel and Frechet for the shape parameter $k < 0$, $k = 0$, and $k > 0$, respectively. The remaining parameters in Equation 3.1 are the location parameter μ (analogous to the mean for the Gaussian distribution) and the scale parameter σ (analogous to the standard deviation for the Gaussian distribution). The Gumbel distribution is obtained through an appropriate limiting process, but is not of interest in the present work. For the Weibull and Frechet distributions, the range of the variate is respectively $-\infty < x \leq \mu - \sigma/k$ and $\mu - \sigma/k \leq x < \infty$. The Weibull distribution has a finite upper endpoint and hence corresponds to a short-tailed parent. The

Frechet distribution has a polynomially decreasing upper tail and corresponds to a long-tailed parent.

As will subsequently be shown, LIBS intensities are typically distributed as Frechet extreme value. It is important to note that the second and higher order moments (and hence the variance) do not exist for $k > 1/2$, and the first moment (the mean) does not exist for $k > 1$. This implies that the standard estimator for the ensemble average, the sample mean, will either be inconsistent (i.e., will not display a reduced variance as the size of the sample increases) for $k > 1/2$ or will not exist at all for $k > 1$. A different formulation is required to obtain the three parameters in Equation 3.1 so that defensible statistical inferences about LIBS intensities can be made. A standard approach is the maximum likelihood method that seeks the solutions for k , μ and σ that maximize the joint distribution of a given set of data, or the likelihood function. For independent samples, the joint distribution of N data is the product of Equation 3.1 for each datum with common shape, location and scale parameters

$$L(k, \mu, \sigma | x_i) = \prod_{i=1}^N \left(\frac{1}{\sigma} \right) \exp \left(- \left(1 + k \frac{x_i - \mu}{\sigma} \right)^{-\frac{1}{k}} \right) \left(1 + k \frac{x_i - \mu}{\sigma} \right)^{-1 - \frac{1}{k}} \quad (3.2)$$

This is maximized by setting the first derivative of Equation 3.2, or its logarithm, for each parameter to zero, yielding a set of three equations

$$\frac{1}{\sigma} \sum_{i=1}^N \left[- \left(1 + \frac{k(x_i - \mu)}{\sigma} \right)^{-\left(\frac{1}{k} + 1\right)} + \frac{k + 1}{1 + k \frac{x_i - \mu}{\sigma}} \right] = 0 \quad (3.3)$$

$$\begin{aligned} \frac{1}{k^2} \left[\sum_{i=1}^N \log \left(1 + k \frac{x_i - \mu}{\sigma} \right) - k \left(\left(1 + k \right) \sum_{i=1}^N \frac{x_i - \mu}{\sigma \left(1 + k \frac{x_i - \mu}{\sigma} \right)} \right) + \right. \\ \left. k \sum_{i=1}^N \left(1 + \frac{k(x_i - \mu)}{\sigma} \right)^{-\frac{1}{k}} \left(\frac{\log \left(1 + \frac{k(x_i - \mu)}{\sigma} \right)}{k^2} - \frac{(x_i - \mu)}{k \sigma \left(1 + \frac{k(x_i - \mu)}{\sigma} \right)} \right) \right] = 0 \end{aligned} \quad (3.4)$$

$$-\frac{N}{\sigma} + \sum_{i=1}^N \left[- \left(\frac{\mu - x_i}{\sigma^2} \right) \left(1 + k \frac{x_i - \mu}{\sigma} \right)^{-\left(\frac{1}{k} + 1\right)} - \left(1 + \frac{1}{k} \right) \left(\frac{1}{1 + k \frac{x_i - \mu}{\sigma}} \right) \left(\frac{k}{\sigma^2} \right) (\mu - x_i) \right] = 0 \quad (3.5)$$

These are coupled and nonlinear, and must be solved iteratively for the maximum likelihood estimators (mles) \hat{k} , $\hat{\mu}$ and $\hat{\sigma}$ using an appropriate algorithm. For the GEVD, the mles are asymptotically efficient (loosely speaking, highly concentrated about the true value for large numbers of data), normal and unbiased, but are nei-

ther unbiased nor fully efficient for finite samples. However, neither is the sample mean when the second and higher moments do not exist. The mles have the distinct advantage of being defined when the moments of the distribution do not exist, and are relatively easy to compute. Presuming that the extreme value distribution is a good fit to LIBS intensity data, the maximum likelihood estimate for the location parameter $\hat{\mu}$ is a good representation of the peak intensity at a given wavelength, and the scale parameter $\hat{\sigma}$ and asymptotic normality can be used to compute approximate confidence intervals on $\hat{\mu}$.

3.4.2 Applicability of Extreme Value Statistics

Quantile-quantile (q-q) plots will be used to demonstrate that LIBS intensities are typically distributed as the Frechet extreme value distribution. The N quantiles of a target distribution are the abscissa values that divide the area under the pdf into $N+1$ equal probability intervals. They are easily obtained from the pdf by solving for Q_j in


$$\int_{-\infty}^{Q_j} f(x)dx = \frac{j - \frac{1}{2}}{N} \quad (3.6)$$

where $j=1,\dots,N$. The order statistics of the intensity data are obtained by sorting them into ascending order. The order statistics divide the area under the target pdf into intervals that will correspond to equal probability if the data are drawn from it, and hence a plot of the quantiles against the order statistics will be a straight line. Systematic departures of the data from the distribution are visible as changes in slope, and anomalous values or outliers are manifest at the extremes, and hence a q-q plot is a useful qualitative tool to assess the suitability of the target distribution as a statistical model.


The fit may be quantified by testing the null hypothesis that the data are drawn from the target distribution against the alternate hypothesis that they are not using the nonparametric Kolmogorov-Smirnov statistic [23] that compares the empirical and target cumulative distribution functions. The Komogorov-Smirnov test statistic may be assessed at the standard 0.95 level for which the critical value is 0.134 for 100 realizations, as were used throughout this work. The null hypothesis is rejected if the test statistic exceeds the critical value.

Figure 3-3 shows q-q plots for the 588.9953 nm Na I peak for single shot and 10 shot analog accumulations on halite using the Echelle set-up. Both are approximately straight, and both accept the null hypothesis that the extreme value distribution is correct (Kolmogorov-Smirnov test statistics of 0.067 and 0.066, respectively). Figure

3-4 shows q-q plots at the same wavelength for single shot and 100 shot analog accumulations for Na in bulk aqueous solution using the Echelle set-up. The single shot data exhibit a shift in slope that reflects the fact that the vast majority of the data correspond to no signal, and hence are only instrument noise. The 100 shot analog accumulations are slightly short-tailed at the upper end. Nevertheless, both pass the Kolmogorov-Smirnov test with test statistics of 0.113 and 0.098, respectively. Figure 3-5 compares q-q plots at the same peak for single shot and 100 shot analog accumulations using the Czerny-Turner set up with a bulk aqueous target. Both are slightly short tailed at the top of the distribution, but both accept the null hypothesis that the data are extreme value (Kolmogorov-Smirnov test statistics of 0.084 and 0.058, respectively). Similar results are observed for the 589.5923 nm Na peak (not shown), or at other wavelengths where signal is present. Further, the shape parameter persistently lies in the region corresponding to the Frechet extreme value distribution, and in many instances exceeds 0.5 so that the variance does not exist.




(a) Single Shot




(b) 10 Shot Accumulations

Figure 3-3: q-q plots for the 588.9953 nm Na I peak for halite using the Echelle set-up




(a) Single Shot




(b) 100 Shot Accumulations

Figure 3-4: q-q plots for the 588.9953 nm Na I peak for bulk aqueous solution using the Echelle set-up



(a) Single Shot



(b) 100 Shot Accumulations

Figure 3-5: q-q plots for the 588.9953 nm Na I peak for bulk aqueous solution using the Czerny-Turner set-up

3.4.3 Extreme Value Statistical Parameters

As a demonstration of the importance of using an appropriate set of statistical estimators for LIBS intensity data, the extreme value mles and the sample mean were computed over the wavelength band 578.5 - 605.9 nm for the halite and aqueous Na samples using both the Echelle and Czerny-Turner set-ups. At each wavelength, 100 realizations of 10 (halite) or 100 (aqueous samples) analog-accumulated shots serve as the data from which the mles (Equations 3.3 - 3.5) and the sample mean are estimated.


Figure 3-6 compares the shape parameter, the location parameter, and the sample mean for halite using the Echelle set-up. While there is wavelength-by-wavelength statistical variability, the shape parameter is persistently above the 0.5 threshold beyond which the variance does not exist, and frequently exceeds the 1.0 threshold beyond which the mean does not exist, especially in the vicinity of the 589 nm Na doublet and over two bands slightly above 600 nm. The extreme value location parameter produces a smooth representation of the LIBS spectrum with limited statistical variability that is consistent with the number of samples. By contrast, the sample mean displays substantial statistical variability with two anomalous nulls amid the Na doublet band. These correspond to the wavelengths where the shape parameter dips below 1.0, so that the mean becomes defined. For data that are long-tailed, such as those drawn from a Frechet extreme value distribution, the sample mean will be dominated by a few large values. This results both in the large wavelength-to-wavelength variability that is apparent in the sample mean and the substantial difference in amplitude between the sample mean and the location parameter. In addition, two large peaks are observed above 600 nm that are barely visible in the extreme value location parameter. These correspond to peaks well above 1.0 in the shape parameter where the extreme value distribution is very long-tailed, and serve as graphic illustration of the sort of erroneous conclusions that can be derived through use of inappropriate statistical estimators.

Figure 3-7 compares the shape parameter, the location parameter, and the sample mean for Na in bulk aqueous solution using the Echelle set-up. The shape parameter is much more uniform with wavelength than for halite (Figure 3-6), but persistently lies around 0.75 where the variance does not exist. As a consequence, the extreme value location parameter displays much less variability than the sample mean, as in Figure 3-6. Figure 3-7 is another example where the use of a standard sample mean estimator may lead to incorrect inferences.


Figure 3-8 compares the shape parameter, the location parameter, and the sample

mean for Na in bulk aqueous solution using the Czerny-Turner set-up. In contrast to the results with the Echelle set-up, the shape parameter is persistently below 0.5, and displays occasional excursions below 0 where the distribution is Weibull. Since both the mean and variance exist throughout the wavelength domain, the extreme value location parameter and the sample mean yield qualitatively similar results. However, the distribution remains extreme value rather than simple Gaussian, and the uncertainty inferred for the sample mean will be inaccurate.


While LIBS intensity data (whether single shot or analog accumulated) empirically appear to persistently be drawn from an extreme value distribution, systematic differences are observed between different experimental set-ups. Whether this is due to the spectrometer or laser design, the experimental geometry, the element under study or some other factor remains unknown. Since incorrect conclusions might be drawn from the use of an inappropriate statistical model, it is strongly urged that LIBS practitioners examine their data to determine the correct approach for each set-up and sample.



(a) Shape Parameter




(b) Location Parameter




(c) Sample Mean


Figure 3-6: Shape parameter, location parameter, and sample mean for halite using the Echelle set-up



(a) Shape Parameter




(b) Location Parameter




(c) Sample Mean


Figure 3-7: Shape parameter, location parameter, and sample mean for bulk aqueous solution using the Echelle set-up



(a) Shape Parameter



(b) Location Parameter




(c) Sample Mean

Figure 3-8: Shape parameter, location parameter, and sample mean for bulk aqueous solution using the Czerny-Turner set-up


3.4.4 Variability

In an attempt to understand the source of data variability, images of plasmas in liquids were taken and the energies of numerous single shot laser pulses were compared. Figure 3-9 shows three sample plasma images taken under the same conditions that visually demonstrates the significant variability of plasma formation. The images are time averaged over 5 μ s. Although hundreds of plasma images were taken, these images were selected to illustrate the variation in intensity, size, and location of formation of the plasma within a bulk liquid. The plasma shown in Figure 3-9(c) displays significantly greater emission intensity and size than those in (a) and (b), illustrating the extreme nature of some plasmas. In contrast, the plasma in Figure 3-9(b) is very weak, with very little emission produced. The distinct differences shown between these three plasmas is a clear indication that on a shot-to-shot basis that plasma variability exists and that extreme plasmas are formed which could account for the extreme intensities recorded.


The energy of 500 laser pulses was measured to examine the contribution of laser pulse energy fluctuations to plasma variability. Figure 3-10 shows that shot-to-shot pulse energy fluctuations do exist; however, the variability is not extreme in nature, suggesting that laser pulse energy variation is not the dominant cause of plasma and peak intensity variations. This suggests that the variability of the plasma formation is due to other effects.



(a) Plasma Image 1



(b) Plasma Image 2



(c) Plasma Image 3

Figure 3-9: Images of plasmas formed in bulk aqueous solution that illustrate the shot-to-shot variability of formation. Images were taken orthogonal to the incoming laser beam. In the images shown, the beam enters from the left. Plasmas were formed using 80 mJ of laser pulse energy.




Figure 3-10: Comparison of laser energies measured for 500 individual laser pulses.

3.5 Conclusions

Examination of the variability of peak intensity for both single shot and analog accumulated LIBS spectra reveals that such data are drawn from an extreme value distribution. In many instances, the distribution has no variance, and in some cases the mean is also undefined. Under either circumstance, the use of the sample mean or variants that include censoring will be statistically inconsistent and the central limit theorem will not apply. A maximum likelihood estimator data processing scheme is presented that accurately deals with the extreme value nature of laser-induced plasma formation. It is strongly urged that this approach be used to ensure accurate scientific inference from LIBS data, and that use of estimators based on the sample mean be discontinued. Plasma images reveal large spatial and intensity differences on a shot-to-shot basis. Laser pulse energy fluctuations are shown to contribute to the variability but are not the primary source.

3.6 Acknowledgements

We acknowledge the National Science Foundation for support of this research under grant OCE0352278. Additional support was received from the Deep Ocean Exploration Institute and the Ocean Ventures Fund of the Woods Hole Oceanographic Institution (WHOI). We thank Norman Farr (WHOI) for his assistance in setting up the imaging system and Hanumant Singh (WHOI) for lending us the camera. We thank the laboratory of S. Michael Angel at the University of South Carolina - Columbia for use of their Czerny-Turner spectrometer.

Bibliography

- [1] E. Tognoni, V. Palleschi, M. Corsi, G. Cristoforetti, N. Omenetto, I. Gornushkin, B. W. Smith, and J. D. Winefordner. *Laser-Induced Breakdown Spectroscopy (LIBS): Fundamentals and Applications*, chapter From Sample to signal in LIBS, pages 122–170. Cambridge University Press, 2006.
- [2] V. Lazic, F. Colao, R. Fantoni, and V. Spizzicchino. Laser-induced breakdown spectroscopy in water: Improvement of the detection threshold by signal processing. *Spectrochimica Acta Part B*, 60:1002–1013, 2005.
- [3] D. A. Cremers and L. J. Radziemski. *Handbook of Laser-Induced Breakdown Spectroscopy*. John Wiley and Sons, Ltd., 2006.
- [4] D. A. Rusak, B. C. Castle, B. W. Smith, and J. D. Winefordner. Recent trends and the future of laser-induced plasma spectroscopy. *Trends in Analytical Chemistry*, 17:453–461, 1998.
- [5] V. Hohreiter, J. E. Carranza, and D. W. Hahn. Temporal analysis of laser-induced plasma properties as related to laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 59(3):327–333, March 2004.
- [6] C. V. Bindhu, S. S. Harilal, M. S. Tillack, F. Najmabadi, and A. C. Gaeris. Energy absorption and propagation in laser-created sparks. *Applied Spectroscopy*, 58(6):719–726, June 2004.
- [7] R. Wisbrun, I. Schechter, R. Niessner, H. Schroder, and K. L. Kompa. Detector for trace elemental analysis of solid environmental samples by laser plasma spectroscopy. *Analytical Chemistry*, 66:2964–2975, 1994.
- [8] L. Xu, V. Bulatov, V. V. Gridin, and I. Schechter. Absolute analysis of particulate materials by laser-induced breakdown spectroscopy. *Analytical Chemistry*, 69:2103–2108, 1997.
- [9] C. V. Bindhu, S. S. Harilal, M. S. Tillack, F. Najmabadi, and A. C. Gaeris. Laser propagation and energy absorption by an argon spark. *Journal of Applied Physics*, 94(12):7402–7407, December 2003.
- [10] V. Lazic, F. Colao, R. Fantoni, and V. Spizzicchino. Recognition of archeological materials underwater by laser induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 60:1014–1024, 2005.
- [11] J. Huang, C. Ke, and K. Lin. Matrix effect on emission/current correlated analysis in laser-induced breakdown spectroscopy of liquid droplets. *Spectrochimica Acta Part B*, 59:321–326, 2004.
- [12] U. Panne, C. Haisch, M. Clara, and R. Niessner. Analysis of glass and glass melts during the vitrification process of fly and bottom ashes by laser-induced plasma

- spectroscopy. Part I: Normalization and plasma diagnostics. *Spectrochimica Acta Part B*, 53:1957–1968, 1998.
- [13] B. C. Castle, K. Talabardon, B. W. Smith, and J. D. Winefordner. Variables influencing the precision of laser-induced breakdown spectroscopy measurements. *Applied Spectroscopy*, 52(5):649–657, 1998.
 - [14] G. A. Lithgow and S. G. Buckley. Effects of focal volume and spatial inhomogeneity on uncertainty in single-aerosol laser-induced breakdown spectroscopy measurements. *Applied Physics Letters*, 87:011501–1 –011501–3, 2005.
 - [15] I. Schechter. Direct aerosol analysis by time resolved laser plasma spectroscopy-improvement by single shot measurements. *Analytical Science and Technology*, 8(4):779–786, 1995.
 - [16] J. E. Carranza and D. W. Hahn. Sampling statistics and considerations for single-shot analysis using laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 57:779–790, 2002.
 - [17] P. K. Kennedy, D. X. Hammer, and B. A. Rockwell. Laser-induced breakdown in aqueous media. *Progress in Quantum Electronics*, 21:155–248, 1997.
 - [18] Y. Ito, O. Ueki, and S. Nakamura. Determination of colloidal iron in water by laser-induced breakdown spectroscopy. *Analytica chimica acta*, 299:401–405, 1995.
 - [19] N. F. Bunkin and A. V. Lobeyev. Influence of dissolved gas on optical breakdown and small-angle scattering of light in liquids. *Physics Letters A*, 229:327–333, 1997.
 - [20] U. Panne and D. Hahn. *Laser-Induced Breakdown Spectroscopy (LIBS) Fundamentals and Applications*, chapter Analysis of aerosols by LIBS, pages 194–253. Cambridge University Press, 2006.
 - [21] S. Kotz and S. Nadarajah. *Extreme Value Distributions: Theory and Applications*. Imperial College Press, 2000.
 - [22] N. L. Johnson, S. Kotz, and N. Balakrishnan. *Continuous Univariate Distributions: Volume 2*. John Wiley & Sons, 2nd edition, 1995.
 - [23] E. L. Lehmann and J. P. Romano. *Testing Statistical Hypotheses: 3rd Edition*. Springer, 2005. Section 14.2.

Chapter 4

Single pulse laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: Interrelationship of gate delay and pulse energy

4.1 Abstract

The ability to make sustained measurements of ocean processes is limited by the number of sensors that are usable for long-term *in situ* analysis. In recent years, laser-induced breakdown spectroscopy (LIBS) has been identified as a viable technique for development into an oceanic chemical sensor. In this paper, single pulsed laser-induced breakdown spectroscopy of high pressure bulk aqueous solutions is used to detect three analytes (sodium, manganese, and calcium) which are of key importance in hydrothermal vent fluids, an ocean environment that would greatly benefit from the development of an oceanic LIBS sensor. The interrelationship of the key experimental parameters, pulse energy and gate delay, for a range of pressures up to 2.76×10^7 Pa, are studied. A minimal effect of pressure on the peak intensity is observed. A short gate delay (less than 200 ns) must be used at all pressures. The need for a relatively low laser pulse energy (less than ≈ 60 mJ) for optimal detection of analytes at high pressure is also established. Na, Mn, and Ca are detectable at pressures up to 2.76×10^7 Pa at 50 ppm, 500 ppm, and 50 ppm, respectively, using an Echelle spectrometer.

4.2 Introduction

New chemical sensors are needed for use in process studies, and are of critical importance as oceanography shifts to a new operational mode using permanent ocean observatories. New sensors take significant time to develop and transform from bench-top laboratory prototypes to ocean-going systems. The development phase requires validation that an analytical technique will work under *in situ* conditions.

LIBS is a type of atomic emission spectroscopy that has been identified as a viable technique for use as a field-going sensor for geochemical and environmental sensing [1]. For example, a new mobile instrument has been developed for evaluating polluted soils [2, 3]. Palanco *et al.* have proposed a field deployable laser-induced breakdown spectrometer system for stand-off measurements at hundreds of meters range [4]. Several groups have investigated LIBS for space exploration [5–10]. Courrèges-Lacoste *et al.* [7] are developing a combined Raman/LIBS instrument for investigating past and present life on Mars. Arp *et al.* [6] have investigated the use of LIBS in the high temperature ($>700\text{K}$), high pressure (order of 9×10^6 Pa) environment of Venus. Another proposed *in situ* application of LIBS is its development into an oceanic chemical sensor.

One ocean environment that would benefit greatly from the development of such a sensor is hydrothermal vents that occur at mid-ocean ridges where seawater circulates through the permeable ocean crust. As seawater moves through the crust, the fluid interacts with the surrounding rock, inducing major chemical changes to the rock and the fluid. At vent orifices, exit temperatures reach $200 - 405^\circ\text{C}$ at ambient pressures of 8.1×10^6 Pa to 3.6×10^7 Pa corresponding to ocean depths of 800 m to 3600 m [11]. *In situ* chemical measurement of the fluids is difficult due to the corrosive nature of the vent environment and irreversible changes in composition that occur when they are removed to the surface. Three elements of importance in vent fluids are sodium, calcium, and manganese. Sodium is the most abundant cation in vent fluids and can be studied to understand phase separation processes [12]. Manganese exists as a trace metal in seawater but is leached from the host rock making it present at higher concentrations in vent fluids [12]. When measured simultaneously with Fe, Mn can be used as an indication of subsurface deposition as Fe precipitates out while Mn stays in solution. Calcium is the second most abundant cation in vent fluids, and is typically enriched in vent fluids, when compared to seawater [13]. Ca is released into vent fluids when sodium is taken up during albitization reactions with the host rock. In vent fluids, concentrations range from approximately 250 - 23,163 ppm for

Na, 0.6 - 399 ppm for Mn, and -54 - 4477 ppm for Ca [11]. In seawater, concentrations are approximately 10,933 ppm Na, <0.001 ppm Mn, and 419 ppm Ca [11].

There is limited prior work on the study of dissolved analytes within bulk aqueous solutions [14–20] due to inherent difficulties in detection. The plasma formed in a bulk liquid displays reduced light intensity and a reduction in emission lifetime due to quenching [14, 15, 21, 22]. In addition, spectral lines are broadened through the Stark effect [14]. Furthermore, “moving breakdown” occurs that randomly changes the distance between the plasma and the collection optics, a phenomenon that is not important for solids in air. The plasma expands along the beam path of the laser, resulting in an elongated plasma that cavitates cylindrically [23]. For many aqueous applications, these issues can be avoided by analysis on a liquid surface, jet, or film; however, for the development of an *in situ* oceanic system, it is necessary to work directly with bulk liquids.

Although we have previously reported on the successful use of LIBS for detection of bulk aqueous analytes at high pressure (up to 2.76×10^7 Pa) [17–19], the development of LIBS into an oceanic chemical sensor requires the optimization of the experimental system to maximize the signal-to-background ratio (SBR) of the spectra and improve the limit of detection. In this work, a comprehensive study of the effect of the two key parameters for single pulse LIBS on SBR was completed. Peak intensities were measured to determine optimal conditions for the detection of Na, Ca, and Mn at high pressure. Subsequently, calibration curves were constructed to estimate the limits of detection using an Echelle spectrometer.

4.3 Experimental

The laboratory set-up for simulating a LIBS sensor in the deep ocean is depicted in Figure 4-1. Plasma formation is induced with a Big Sky CFR-200 Nd:YAG laser operated at 1064 nm with a 5 Hz repetition rate. The laser is equipped with a motorized variable attenuator, serially controlled by a computer, enabling the laser pulse energy (E) to be varied from 0 mJ to 200 mJ in increments of approximately 1 mJ. Plasma emission is collected with an Echelle spectrometer (LLA Echelle ESA 3000) capable of detecting wavelengths of 200 - 780 nm with a spectral resolution of 10 - 50 pm. Timing control of the laser and turn-on of the spectrometer is managed by a timing box (Berkeley Nucleonics Corporation Model 565). The LIBS timing parameters are gate delay (t_d , the time between the laser pulse and turn-on of the spectrometer) and gate width (t_b , the integration time of the spectrometer), both




Figure 4-1: LIBS laboratory set-up

shown in Figure 4-2.

An 8.89 cm \times 8.89 cm \times 8.89 cm titanium sample chamber that holds 27 ml of liquid is connected to a high pressure metering pump (Eldex Model A-30-S) that is used to pressurize samples to 4.1×10^7 Pa. The sample chamber is equipped with a sapphire window (Meller Optics, 2.54 cm diameter \times 0.64 cm thickness, AR-coated at 532 nm/1064 nm, custom part) that enables laser pulses to enter the sample chamber. A series of AR-coated optics are used to focus the laser beam into the sample chamber (Figure 4-3), with the final focusing lens fit into the sample chamber. An additional lens is used to focus the plasma light onto an optical fiber and delivers it to the Echelle spectrometer (Figure 4-3). The plasma light is collected collinear to the incoming laser beam. This optical geometry was selected because it is the only practical configuration for an ocean-going LIBS system. Data were collected using ESAWIN software. Laser energy is measured using a laser energy sensor (Coherent J25LP-MB) combined with an energy meter (Coherent FieldMaxII-Top).

To determine optimal conditions for the detection of the three analytes, spectral intensities were measured over a range of LIBS system parameters. For Na and Mn, detailed studies were conducted at five pressures (1×10^5 Pa, 6.89×10^6 Pa, 1.38×10^7 Pa, 2.07×10^7 Pa, and 2.76×10^7 Pa). For Ca, the studies were conducted at three pressures (1×10^5 Pa, 1.38×10^7 Pa, and 2.76×10^7 Pa). The gate delay and laser pulse energy were systematically varied to determine their effect on both plasma intensity and SBR. Five spectra were taken for each parameter pair, each composed of 100 accumulated shots. The laser pulse energy ranged from 10 to 170 mJ in 10 mJ increments. Laser beam waist width d_{σ_o} can be estimated from

$$d_{\sigma_o} = \frac{4f\lambda M^2}{\pi D} \quad (4.1)$$

where f is the focal length of the focusing lens (35 mm), λ is the laser wavelength




Figure 4-2: Timing parameters




Figure 4-3: Optical Configuration. M_1 = 25 mm diameter 1064 nm Nd:YAG mirror; L_1 = 12 mm x 12 mm lens; L_2 = 25 mm x 50 mm lens; M_2 = 50 mm diameter 1064 nm Nd:YAG mirror; L_3 = 25 mm x 35 mm lens; FO = Fiber Optic

(1064 nm), M^2 is the beam propagation ratio which is typically 2 - 10 for Nd:YAG lasers (we therefore choose a value of 6), and D is the diameter of the illuminated aperture of the focusing lens (≈ 25 mm) [24]. The beam waist width for the system is approximately 0.07 mm. The average irradiance (I_f) at the beam waist is

$$I_f = \frac{\pi E_L D^2}{4\tau_L f^2 \lambda^2 M^4} \quad (4.2)$$

where E_L is the laser pulse energy and τ_L ($= 7.5$ ns) is the pulse duration at the full peak width at half of the maximum intensity (FWHM) [24]. The pulse energy was varied between 10 - 170 mJ resulting in variation of irradiance at the beam waist from $\approx 1.31 \times 10^{12}$ to 2.23×10^{13} W/cm². The gate delay was 50, 75, 100, 125, 150, 200, 300, and 500 ns. Each combination of energy and gate delay was tested, resulting in 136 different conditions for the optimization studies.

The signal-to-background ratio is

$$SBR = 20 \log_{10} \frac{Amplitude_{peak}}{Amplitude_{background}} \quad (4.3)$$

where the amplitude of the background is defined as the spectral average over a region where no peaks are expected. The background was calculated for Na by using the spectral region 200 nm to 500 nm and for Mn and Ca by using the spectral region 430 nm to 530 nm.

Calibration curves were made for Na, Ca, and Mn with ten spectra being taken at each concentration, each composed of 100 accumulated shots. The experimentation conditions used for the calibration curves are detailed in Table 4.1.

Table 4.1: Calibration curve conditions

Analyte	Concentrations Tested (ppm)	E (mJ)	t_d (ns)
Na	0.05, 0.1, 0.5, 1, 5, 10, 50, 100, 500, 1000	40	50
Mn	1, 5, 10, 50, 100, 500, 1000	30	50
Ca	1, 5, 10, 50, 100, 500, 1000	30	50

For all experiments, the gate width was held constant at 200 ns. In addition, the amplification of the Echelle spectrometer was set to the maximum value of 4000.

All raw spectra were processed using extreme value distribution statistics detailed in a paper by Michel and Chave [25]. Data from 9 wavelengths were grouped for processing. Where shown, error bars represent the double sided 95% confidence limits [25].

Solutions were made from NaCl, $\text{MnSO}_4 \cdot \text{H}_2\text{O}$, and $\text{CaCl}_2 \cdot 2\text{H}_2\text{O}$ dissolved in de-ionized water for the Na, Mn, and Ca studies, respectively. All concentrations are given in parts per million (ppm, wt./vol.).

4.4 Results and Discussion

4.4.1 Sodium

The interrelationship of gate delay and laser pulse energy for sodium was studied using a concentration of 100 ppm. The intensities of the 588.995 nm and 589.6 nm Na peaks were measured, and are shown in Figures 4-4 and 4-5. As pressure rises, an increase in signal intensity is observed with the maximum peak intensity present at 2.76×10^7 Pa. An examination of these figures shows that the greatest peak intensity exists at the shortest gate delay. As t_d increases, peak intensity decreases, independent of both the laser pulse energy and ambient pressure. The data variability present in the plots is indicative of significant plasma variability. Overall, there appears to be little effect of energy on intensity. Examination of the SBR provides important information for selecting optimal parameters. Figures 4-6 and 4-7 detail the interrelationship of t_d , energy, and SBR. A smaller t_d tends to exhibit a higher SBR due to a stronger signal, as seen in Figures 4-4 and 4-5. Furthermore, Figure 4-4 suggests that a lower energy pulse consistently provides a higher SBR. As pressure increases, the SBR again tends to increase. The data suggest that the highest SBR exists when a low energy pulse (20 - 60 mJ) and a relatively small t_d are used (50 - 150 ns). As first reported in Michel et al. [17], this suggests that an optimal range of laser energies exists that tend to be relatively low.

For the best SBR, a pulse energy of 40 mJ and a gate delay of 50 ns were identified for detection of Na over a range of pressures and spectra of this condition are plotted in Figure 4-8. Calibration curves for sodium were constructed to determine the limit of detection (Figure 4-9). These suggest that sodium can be detected at a concentration of approximately 50 ppm using the present apparatus. Spectra of the Na calibration data are shown in Figure 4-10 that illustrate that 50 ppm is the lowest concentration where Na is detectable.




Figure 4-4: Interrelationship of gate delay, laser pulse energy, and peak intensity for Na (588.995 nm) over a range of pressures (a) 1×10^5 Pa (b) 6.89×10^6 Pa (c) 1.38×10^7 Pa (d) 2.07×10^7 Pa (e) 2.76×10^7 Pa




Figure 4-5: Interrelationship of gate delay, laser pulse energy, and peak intensity for Na (589.6 nm) over a range of pressures (a) 1×10^5 Pa (b) 6.89×10^6 Pa (c) 1.38×10^7 Pa (d) 2.07×10^7 Pa (e) 2.76×10^7 Pa




Figure 4-6: Interrelationship of pressure, gate delay, energy, and signal-to-background for Na (588.995 nm) (a) 1×10^5 Pa (b) 6.89×10^6 Pa (c) 1.38×10^7 Pa (d) 2.07×10^7 Pa (e) 2.76×10^7 Pa




Figure 4-7: Interrelationship of pressure, gate delay, energy, and signal-to-background for Na (589.6 nm) (a) 1×10^5 Pa (b) 6.89×10^6 Pa (c) 1.38×10^7 Pa (d) 2.07×10^7 Pa (e) 2.76×10^7 Pa






Figure 4-8: Spectra of Na (588.995 nm and 589.6 nm) taken with a pulse energy of 40 mJ and a gate delay of 50 ns. From bottom to top, the spectra were taken at 1×10^5 Pa, 6.89×10^6 Pa, 1.38×10^7 Pa, 2.07×10^7 , and 2.76×10^7 Pa. For clarity, the spectra have been offset from each other by 8000 a.u.



(a)



(b)

Figure 4-9: Sodium calibration curves of the (a) 588.995 nm peak and the (b) 589.6 nm peak. $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa




Figure 4-10: Spectra of sodium (588.995 nm and 589.6 nm) at 2.76×10^7 Pa made over a range of NaCl concentrations. The concentrations from bottom to top are 10 ppm, 50 ppm, 100 ppm, 500 ppm, and 1000 ppm. For clarity, the spectra have been offset from each other by 1000 a.u.

4.4.2 Manganese

The effect of laser pulse energy and gate delay on spectra for manganese were studied using a concentration of 1000 ppm Mn over a range of pressures. Although a Mn triplet exists at 403 nm, peak broadening in liquids causes it to be unresolvable, and therefore we report a single 403 nm peak. The interrelationship of pressure, gate delay, energy, and intensity are shown in Figure 4-11. As pressure increases, peak intensity also rises. A similar finding was reported by Michel et al. under a single tested condition (single E and t_d) [17]. Figure 4-11 shows that irrespective of gate delay and energy the peak intensity of Mn increases with pressure. Mn also exhibits a higher peak intensity at a lower laser pulse energy at a shorter t_d . When the corresponding SBR plots are examined (Figure 4-12), SBR is shown to be smallest at the lowest pressure (1×10^5 Pa). Again, the need for a short t_d and a low E is evident.

From the optimization studies, an energy of 30 mJ with a gate delay of 50 ns was selected as a condition that would provide good detection of Mn over a broad range of pressures. The selected condition is plotted at all five pressure conditions in Figure 4-13. To determine the limit of detection of Mn, a calibration curve was constructed using 30 mJ and t_d of 50 ns (Figure 4-14). Figure 4-15 shows spectra made at these conditions at 2.76×10^7 Pa over a range of concentrations. The limit of detection was found to be 500 ppm which is higher than the concentration found in vent fluids.




Figure 4-11: Interrelationship of pressure, gate delay, energy, and intensity for Mn (403.076 nm) (a) 1×10^5 Pa (b) 6.89×10^6 Pa (c) 1.38×10^7 Pa (d) 2.07×10^7 Pa (e) 2.76×10^7 Pa






Figure 4-12: Interrelationship of pressure, t_d , E , and SBR for Mn (403.076 nm) (a) 1×10^5 Pa (b) 6.89×10^6 Pa (c) 1.38×10^7 Pa (d) 2.07×10^7 Pa (e) 2.76×10^7 Pa



(a)



(b)

Figure 4-13: Manganese spectra using a 30 mJ energy pulse and a gate delay of 50 ns. (a) Full spectral region. (b) Manganese (403 nm peak) In (a) and (b) the spectra from bottom to top are at 1×10^5 Pa, 6.89×10^6 Pa, 1.38×10^7 Pa, 2.07×10^7 Pa, 2.76×10^7 , respectively. For clarity, the spectra have been offset from each other by 2000 a.u.




Figure 4-14: Mn (403 nm) calibration curve. $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa.





Figure 4-15: Spectra of manganese at 2.76×10^7 Pa made at a range of Mn concentrations (a) Full spectral region (b) Manganese. In (a) and (b) the concentrations of the spectra from bottom to top are 100 ppm, 500 ppm, and 1000 ppm. For clarity, the spectra have been offset from each other by 100 a.u.

4.4.3 Calcium

Three calcium peaks, 393 nm (ionic), 396 nm (ionic) and 422 nm (atomic), were studied. The interrelationship of the measurement parameters of Ca were determined and are shown in Figures 4-16 to 4-21. The importance of a short t_d is evident when SBR is examined (Figures 4-19 to 4-21). Although three calcium peaks are detectable, the 422 nm peak is the strongest, and therefore the selection of an optimal condition was based on this peak. A laser energy pulse of 30 mJ with a gate delay of 50 ns was selected as the optimal condition for detection of Ca, and spectra illustrating this condition are shown in Figure 4-22. Calcium calibration curves were constructed using these conditions and are shown in Figure 4-23 for both the 393 nm and the 422 nm peaks. These suggest the limit of detection for Ca is 50 ppm using the present apparatus. Spectra for selected concentrations are illustrated in Figure 4-24.




Figure 4-16: Interrelationship of pressure, t_d , E, and intensity for Ca (393 nm) (a) 1×10^5 Pa (b) 1.38×10^7 Pa (c) 2.76×10^7 Pa




Figure 4-17: Interrelationship of pressure, t_d , energy, and intensity for Ca (396 nm)
(a) 1×10^5 Pa (b) 1.38×10^7 Pa (c) 2.76×10^7 Pa







Figure 4-18: Interrelationship of pressure, t_d , E , and intensity for Ca (422 nm) (a) 1×10^5 Pa (b) 1.38×10^7 Pa (c) 2.76×10^7 Pa



(a)




(b)




(c)


Figure 4-19: Interrelationship of pressure, t_d , E, and signal-to-background for Ca (393 nm) (a) 1×10^5 Pa (b) 1.38×10^7 Pa (c) 2.76×10^7 Pa



(a)




(b)




(c)


Figure 4-20: Interrelationship of pressure, t_d , E , and signal-to-background for Ca (396 nm) (a) 1×10^5 Pa (b) 1.38×10^7 Pa (c) 2.76×10^7 Pa



(a)




(b)




(c)

Figure 4-21: Interrelationship of pressure, t_d , E , and signal-to-background for Ca (422 nm) (a) 1×10^5 Pa (b) 1.38×10^7 Pa (c) 2.76×10^7 Pa




(a)




(b)

Figure 4-22: Calcium spectra using 30 mJ and a 50 ns gate delay. (a) Full spectrum (b) Calcium peaks (393 nm, 396 nm, and 422 nm). Both (a) and (b) spectra from bottom to top: 1×10^5 Pa, 1.38×10^7 Pa, and 2.76×10^7 Pa. For clarity, the spectra have been offset from each other by 15,000 a.u.




(a) Calcium (393 nm) calibration curve




(b) Calcium (422 nm) calibration curve

Figure 4-23: Calcium calibration curves (a) 393 nm (b) 422 nm, $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa.



(a)



(b)

Figure 4-24: Calcium spectra at 2.76×10^7 Pa, Calcium peaks are present at 393 nm, 396 nm, and 422 nm. (a) Spectra from bottom to top are 10 ppm, 50 ppm, 100 ppm, 500 ppm, 1000 ppm respectively. For clarity, the spectra have been offset from each other by 2000 a.u. (b) Spectra from bottom to top are 10 ppm, 50 ppm, and 100 ppm respectively. For clarity, the spectra have been offset from each other by 50 a.u.

4.5 Conclusions

Sodium, manganese, and calcium are all detectable in high pressure, bulk aqueous solutions using single pulse LIBS with little effect of pressure on the spectra. This comprehensive study of the interrelationship of gate delay and energy for selecting the optimal condition for detection of these analytes has shown that, irrespective of the laser pulse energy selected, the gate delay should be very short (less than 200 ns) for the detection of analytes in bulk aqueous solutions. The need for a short gate delay is independent of pressure. This study has also shown that a low energy pulse (less than ≈ 60 mJ) is optimal.

Calibration curves were made to determine limits of detection using the current system set-up and further work is needed to look at reproducibility of the actual curves. Calibration curves show that Na, Mn, and Ca can be detected at 50 ppm, 500 ppm, and 50 ppm, respectively. The calibration curves also demonstrate a minimal effect of pressure on spectra. However, the limits of detection were higher than expected. With the current LIBS set-up, the detection limits of Na and Ca are below the levels found in vent fluids. However, Mn would not be detectable. This can be attributed to the low light throughput of the $f/10$ Echelle system. To significantly improve the light throughput, it would be advisable to use a spectrometer with a smaller f number. For example, using a spectrometer with an f number of 2 could improve the throughput by a factor of approximately 25 and therefore improve the ability to detect Na, Mn, and Ca in high pressure aqueous environments. For example, with the use of a PMT as the detector, Cremers *et al.* [14] showed substantially improved detection limits for bulk liquids for Na I (589.00 nm) at a level of 0.014 ppm and Ca II (393.37 nm) at a level of 0.8 ppm. Therefore, additional work is necessary to optimize the light collection by changing the system components.

4.6 Acknowledgments

We acknowledge the National Science Foundation for support of this research under grant OCE-0527927. Additional support was received from the Deep Ocean Exploration Institute and the Ocean Ventures Fund of the Woods Hole Oceanographic Institution.

Bibliography

- [1] R. S. Harmon, F. C. DeLucia, C. E. McManus, N. J. McMillan, T. F. Jenkins, M. E. Walsh, and A. Miziolek. Laser-induced breakdown spectroscopy - an emerging chemical sensor technology for real-time field-portable, geochemical, mineralogical, and environmental applications. *Applied Geochemistry*, 21(5):730–747, May 2006.
- [2] A. Bertolini, G. Carelli, F. Francesconi, M. Francesconi, L. Marchesini, P. Marsili, F. Sorrentino, G. Cristoforetti, S. Legnaioli, V. Palleschi, L. Pardini, and A. Salvetti. Modi: a new mobile instrument for *in situ* double-pulse LIBS analysis. *Analytical and Bioanalytical Chemistry*, V385(2):240–247, May 2006.
- [3] M. Corsi, G. Cristoforetti, M. Hidalgo, S. Legnaioli, V. Palleschi, A. Salvetti, E. Tognoni, and C. Vallebona. Double pulse, calibration-free laser-induced breakdown spectroscopy: A new technique for *in situ* standard-less analysis of polluted soils. *Applied Geochemistry*, 21(5):748–755, 2006.
- [4] S. Palanco, C. Lopez-Moreno, and J. J. Laserna. Design, construction and assessment of a field-deployable laser-induced breakdown spectrometer for remote elemental sensing. *Spectrochimica Acta Part B*, 61(1):88–95, January 2006.
- [5] Z. A. Arp, D. A. Cremers, R. C. Wiens, D. M. Wayne, B. Sallé, and S. Maurice. Analysis of water ice and water ice/soil mixtures using laser-induced breakdown spectroscopy: Application to Mars polar exploration. *Applied Spectroscopy*, 58:897–909, 2004.
- [6] Z. A. Arp, D. A. Cremers, R. D. Harris, D. M. Oschwald, G. R. Parker Jr., and D. M. Wayne. Feasibility of generating a useful laser-induced breakdown spectroscopy plasma on rocks at high pressure: preliminary study for a Venus mission. *Spectrochimica Acta Part B*, 59:987–999, 2004.
- [7] G. B. Courrèges-Lacoste, B. Ahlers, and F. R. Pérez. Combined Raman spectrometer/laser-induced breakdown spectrometer for the next ESA mission to Mars. *Spectrochimica Acta Part A*, In Press, 2007.
- [8] R. Brennetot, J. L. Lacour, E. Vors, A. Rivoallan, D. Vailhen, and S. Maurice. Mars analysis by laser-induced breakdown spectroscopy (MALIS): Influence of Mars atmosphere on plasma emission and study of factors influencing plasma emission with the use of Doehlert designs. *Applied Spectroscopy*, 57(7):744–752, 2003.
- [9] A. Knight, N. Scherbarth, D. Cremers, and M. Ferris. Characterization of laser-induced breakdown spectroscopy (LIBS) for application to space exploration. *Applied Spectroscopy*, 54:331–340, 2000.
- [10] B. Sallé, J.-L. Lacour, P. Mauchien, P. Fichet, S. Maurice, and G. Manhes. Comparative study of different methodologies for quantitative rock analysis by

- laser-induced breakdown spectroscopy in a simulated Martian atmosphere. *Spectrochimica Acta Part B*, 61:301–313, 2006.
- [11] C. R. German and K. L. Von Damm. *Treatise on Geochemistry*, chapter Hydrothermal Processes, pages 181–222. Elsevier, 2003.
 - [12] K. L. Von Damm. Controls on the chemistry and temporal variability of seafloor hydrothermal fluids. *Seafloor hydrothermal systems: physical, chemical, biological, and geological interactions: Geophysical Monograph 91*, pages 222–247, 1995.
 - [13] K. L. Von Damm. Chemistry of hydrothermal vent fluids from 9° - 10° N, East Pacific Rise: ‘Time zero,’ The immediate post-eruptive period. *Journal of Geophysical Research*, 105:11203–11222, 2000.
 - [14] D. A. Cremers, L. J. Radziemski, and T. R. Loree. Spectrochemical analysis of liquids using the laser spark. *Applied Spectroscopy*, 38:721–729, 1984.
 - [15] R. Knopp, F. J. Scherbaum, and J. I. Kim. Laser induced breakdown spectroscopy (LIBS) as an analytical tool for the detection of metal ions in aqueous solutions. *Fresenius’ Journal of Analytical Chemistry*, 355:16–20, 1996.
 - [16] W. Pearman, J. Scaffidi, and S. M. Angel. Dual-pulse laser-induced breakdown spectroscopy in bulk aqueous solution with an orthogonal beam geometry. *Applied Optics*, 42:6085–6093, 2003.
 - [17] A. P. M. Michel, M. Lawrence-Snyder, S. M. Angel, and A. D. Chave. Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: evaluation of key measurement parameters. *Applied Optics*, 46, 2007.
 - [18] M. Lawrence-Snyder, J. Scaffidi, S. M. Angel, A. P.M. Michel, and A. D. Chave. Sequential-pulse laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, 61:171–176, 2007.
 - [19] M. Lawrence-Snyder, J. Scaffidi, S. M. Angel, A. P. M. Michel, and A. D. Chave. Laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, 60:786–790, 2006.
 - [20] A. De Giacomo, M. Dell’Aglia, F. Colao, R. Fantoni, and V. Lazic. Double-pulse LIBS in bulk water and on submerged bronze samples. *Applied Surface Science*, 247:157–162, 2005.
 - [21] A. E. Pichahchy, D. A. Cremers, and M. J. Ferris. Elemental analysis of metals under water using laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 52:25–39, 1997.
 - [22] C. Haisch, J. Liermann, U. Panne, and R. Niessner. Characterization of colloidal particles by laser-induced plasma spectroscopy (LIPS). *Analytica Chimica Acta*, 346:23–25, 1997.

- [23] P. K. Kennedy, D. X. Hammer, and B. A. Rockwell. Laser-induced breakdown in aqueous media. *Progress in Quantum Electronics*, 21:155–248, 1997.
- [24] R. Noll. Terms and notations for laser-induced breakdown spectroscopy. *Analytical and Bioanalytical Chemistry*, V385(2):214–218, 2006.
- [25] A. P. M. Michel and A. D. Chave. Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics. *Spectrochimica Acta Part B*, In Press.

Chapter 5

Double pulse laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: Interrelationship of gate delay, pulse energies, interpulse delay, and pressure

5.1 Abstract

Laser-induced breakdown spectroscopy (LIBS) has been identified as an analytical chemistry technique suitable for field use. Mid-ocean ridge hydrothermal vents would greatly benefit from the development of an *in situ* LIBS sensor. In this paper, double pulse laser-induced breakdown spectroscopy is used to detect five analytes (sodium, manganese, calcium, magnesium, and potassium) that are of key importance in understanding the chemistry of hydrothermal vent fluids, and of mixtures of vent fluids and seawater. The high pressure aqueous environment of the deep ocean is simulated in the laboratory and the key double pulse experimental parameters (laser pulse energies, gate delay, and interpulse delay) are studied at pressures up to 2.76×10^7 Pa. Each element is found to have a unique optimal set of parameters for detection, and the elements are not detectable outside of the set. For all pressures and energies, a short (≤ 100 ns) gate delay is necessary. As pressure increases, a shorter interpulse

delay is needed and the conditions effectively become single pulse. Calibration curves reveal the limits of detection of the elements (5000 ppm Mg, 500 ppm K, 500 ppm Ca, 1000 ppm Mn, and 50 ppm Na). When compared to our previous single pulse work for Ca, Mn, and Na, double pulse LIBS for high pressure aqueous solutions did not improve the limits of detection.

5.2 Introduction

Laser-induced breakdown spectroscopy (LIBS) has recently been identified as an analytical chemistry technique suitable for field deployment for the analysis of environmental and geochemical samples [1]. Several groups are evaluating the use of LIBS for space exploration [2–7]. Another environmental area in which there is a critical need for new sensors is the ocean. Sensor development for use with underwater vehicles is ongoing and a more requirement is for sensors to be developed for use on permanent ocean observatories.

One environment identified as potentially benefitting from application of an oceanic LIBS sensor is the deep-sea hydrothermal vent environment. Hydrothermal vents occur at mid-ocean ridges where seawater circulates through the permeable ocean crust, allowing the fluid to interact with the surrounding rock and resulting in major fluid chemical changes. At vent orifices, exit temperatures reach 200 - 405°C at ambient pressures of 8.1×10^6 Pa to 3.6×10^7 Pa corresponding to ocean depths of 800 m to 3600 m [8].

At hydrothermal vents, as the hot fluids mix with seawater, rapid chemical changes occur and some elements precipitate out (e.g., sulfate minerals [9]). Obtaining *in situ* chemical measurements of the fluid is difficult due to the corrosive nature and high temperature of the fluid. Collection of the fluid for analysis shipboard or in a laboratory introduces chemical changes as the temperature and pressure of the fluid is changed during sample recovery. Five critical elements at hydrothermal vents are sodium, calcium, manganese, magnesium, and potassium. Sodium is the dominant cation in vent fluids, and provides insight into phase separation processes [10]. Calcium is the second most dominant cation in vent fluids, and is usually found at a greater concentration in vent fluids than seawater [11]. Ca is released into vent fluids when Na is taken up in albitization reactions with the host rock [11]. Manganese exists as a trace metal in seawater but has a higher concentration in vent fluids due to leaching from the host rock [10]. Magnesium is practically nonexistent in hydrothermal vent fluids; however, if any is detected in vent fluids, contamination by

entrainment of ambient seawater is indicated [10]. Potassium is typically highly enriched in vent fluids due to leaching from basalts [10]. In vent fluids, concentrations range from approximately 250 - 23,163 ppm for Na, 0.6 - 399 ppm for Mn, -54 - 4477 ppm for Ca, -47 - 3166 ppm for K, and 0 ppm for Mg [8]. In seawater, concentrations are approximately 10,933 ppm Na, <0.001 Mn, 419 ppm Ca, 405 ppm K, and 1300 ppm Mg [8].

Development of an oceanic LIBS sensor necessitates laboratory investigations into the system parameters for the detection of analytes under high pressure bulk aqueous conditions. Although LIBS analysis of liquids is more difficult than the analysis of solid or air samples, a few studies have focused on dissolved analytes in bulk solutions [12–18]. Some issues for LIBS in aqueous solutions include a reduction in plasma light intensity and emission lifetime due to quenching, the Stark effect may cause spectral lines to be broadened, and moving breakdown can change the distance between the plasma and the collection fiber optic [12, 13, 19–21]. The breakdown threshold in water is also significantly greater than for solids [22]. Laser-induced plasmas are weak in water, as they are cooled by Bremsstrahlung and shockwave emissions, and by thermal conduction. Rapid cooling of the plasma also increases electron-ion recombination and plasma emissions last only a few hundred nanoseconds [23].

In 1984, Cremers *et al.* showed that several elements, including Na, K, Mg, and Ca, could be identified in bulk aqueous solution using double pulse LIBS. This work, carried out at atmospheric pressure, showed that double pulse LIBS improved the detection limit for metals and ions in bulk aqueous solution [12]. In double pulse LIBS for dissolved analytes, the first laser pulse produces a plasma that creates a laser-induced cavitation bubble. The second laser pulse produces a plasma within the bubble [23, 24]. This is in contrast to single pulse LIBS, where the plasma is simply formed in a liquid environment. The duration of the laser-induced plasma in the bubble is on the order of a few microseconds and the bubble lifetime is on the order of a few hundred microseconds. Therefore, it can be assumed that the plasma from the second pulse is expanding in a quasi-stationary environment induced by the first laser pulse [24, 25]. When the bubble is first formed, its pressure is greater than that of the surrounding liquid and the bubble begins to expand which leads to a pressure drop. At the point of maximum expansion, the bubble pressure is less than the pressure of the surrounding fluid and the bubble begins to collapse. During this collapsing phase, the temperature and pressure in the bubble again increases and if there is enough energy stored within the bubble, it can re-expand. Many

such oscillations of expansion and compression are possible [23]. By adjusting the interpulse delay between the laser pulses, it is possible to select the phase that the bubble is in such that bubble pressure is at a nadir. If the interpulse delay is too long or too short, the plasma will expand in a high pressure environment so that the spectral lines will be broadened due to collisions [23]. Selection of the interpulse delay time therefore is critical for bulk liquids.

Since Cremer's pioneering work, several groups have used dual pulse LIBS for the detection of analytes in bulk liquids [12, 14, 18, 25–27] and more recently for bulk solutions at high pressures [15, 16]. In Michel *et al.* [15], it was reported that analyte detection in high pressure bulk solutions was highly dependent on the interpulse delay. When a shorter interpulse delay time was used ($\ll 1 \mu\text{s}$), the signal intensity was enhanced compared to when longer delay times were employed. However, it was noted that this time may not be long enough for cavitation bubbles to fully form and expand, as occurs at low pressure. Michel *et al.* also found that the optimal energy levels needed for emission seemed to vary by analyte [15]. Lawrence-Snyder *et al.* [16] report that increasing solution pressure reduces double pulse emission enhancement so that little improvement was noted over single pulse above 1×10^7 Pa.

In Chapter 4, optimization of single pulse conditions for three analytes (Na, Mn and Ca) was reported and the limits of detection were not at the levels that we had hoped to achieve. Through optimization of the double pulse set-up, the goal is to improve the limits of detection and, in addition, establish limits of detection for other elements. Although Lawrence-Snyder *et al.* previously reported on the use of double pulses for the detection of analytes at high pressures and found no enhancements above 1×10^7 Pa [16], several differences exist between these studies. Lawrence-Snyder *et al.* report their findings using an orthogonal beam geometry which would be impractical for the development of an oceanic sensor in many applications, the work does not extend to the ambient pressures for most hydrothermal vents, and optimization of laser pulse energies was not carried out. Through the completion of a thorough optimization, the use of double pulse LIBS at pressures up to 2.76×10^7 Pa is now investigated and the limits of detection for five key elements are determined.

5.3 Experimental

Double pulse high pressure aqueous LIBS experiments were completed using the laboratory set-up detailed in Figure 5-1. Two Big Sky CFR-200 Nd:YAG lasers configured with the beams co-linearly aligned prior to exiting the aperture (Particle Image Ve-




Figure 5-1: Laboratory set-up for high pressure aqueous double pulse LIBS experiments

locimetry or PIV configuration) were operated at 1064 nm with a 5 Hz repetition rate. Each laser is equipped with a motorized variable attenuator serially controlled by a computer, enabling the laser pulse energies (E_1 and E_2) to be varied independently from 0 to 200 mJ, in increments of approximately 1 mJ. Plasma emission is collected with an Echelle spectrometer (LLA Echelle ESA 3000). The spectrometer is capable of detecting elements with wavelengths of 200 - 780 nm at a spectral resolution of 10 - 50 pm. Accurate timing of the laser firing, the interpulse delay time (ΔT , the time between the firing of laser pulse 1 and laser pulse 2) and the gate delay (t_d , the time between the firing of the second laser pulse and the turn-on of the spectrometer) was controlled by a timing box (Berkeley Nucleonics Corporation Model 565). The integration time of the spectrometer, gate width (t_b), is also controllable. The double pulse parameters are clarified in Figure 5-2.

An 8.89 cm \times 8.89 cm \times 8.89 cm titanium sample chamber that holds 27 ml of liquid connects to a high pressure metering pump (Eldex Model A-30-S) using Swagelok fittings to pressurize samples up to 4.1×10^7 Pa. The sample chamber is equipped with a sapphire window (Meller Optics, 2.54 cm diameter \times 0.64 cm thickness, AR-coated at 532 nm/1064 nm, custom part) that allows laser pulses to enter the sample chamber. A series of AR-coated optics are used to focus the laser beams into the sample chamber and to focus the plasma light onto an optical fiber for delivery to the spectrometer. The plasma light is collected collinear to the incoming laser beam to simulate the design that would be most practical for an ocean-going




Figure 5-2: Double pulse LIBS timing parameters

LIBS system. Data are collected using ESAWIN software. Laser energy is measured using a laser energy sensor (Coherent J25LP-MB) combined with an energy meter (Coherent FieldMaxII-Top).

For double pulse optimization studies, peak intensities for five analytes (Mg, K, Na, Ca, and Mn) were measured over a range of system parameters: E_1 , E_2 , t_d , and ΔT . E_1 and E_2 were each tested at 20 mJ, 60 mJ, 100 mJ, and 140 mJ. The average irradiance (I_f) at the beam waist can be estimated from

$$I_f = \frac{\pi E_L D^2}{4\tau_L f^2 \lambda^2 M^4} \quad (5.1)$$

where E_L is the laser pulse energy, D is the diameter of the illuminated aperture of the focusing lens (≈ 25 mm), and τ_L is the pulse duration at the full peak width at half of the maximum intensity (FWHM; $\tau_L = 7.5$ ns), f is the focal length of the focusing lens (35 mm), λ is the laser wavelength (1064 nm), and M^2 is the beam propagation ratio that is typically 2 - 10 for Nd:YAG lasers (we estimate a value of 6) [28]. The pulse energies of 20 mJ, 60 mJ, 100 mJ, and 140 mJ correspond to irradiance at the beam waist of $\approx 2.62 \times 10^{12}$ W/cm², 7.87×10^{12} W/cm², 1.31×10^{13} W/cm², 1.84×10^{13} W/cm², respectively.

For optimization of experimental parameters, studies were conducted at three

pressures, 1×10^5 Pa, 1.38×10^7 Pa, and 2.76×10^7 Pa, which correspond to ocean depths of approximately 0 m, 1362 m, and 2724 m respectively. For Mg and K, the gate delay was tested at 10 ns, 50 ns, 100 ns, and 500 ns and the interpulse delay was tested at 50 ns, 100 ns, 500 ns, 1000 ns, and 5000 ns. For Na, Ca, and Mn, the gate delay was held constant at 50 ns and the interpulse delay was tested at 50 ns, 500 ns, and 5000 ns. Each combination of these conditions was evaluated for a total of 320 conditions for Mg and K and 48 conditions for Mn, Na, and Ca. For all studies, the gate width was held constant at 200 ns and the amplification of the Echelle spectrometer was set to the maximum value of 4000. Five spectra were taken at each optimization condition and for calibration curves ten spectra were taken. All spectra were comprised of 100 accumulations. To account for the high resolution of the spectrometer and the peak broadening that occurs from liquids, all data were grouped into sets of 9 wavelengths. Each set of datapoints (9×5 for the optimization or 9×10 for the calibration curves) was then processed using extreme value distribution statistics described by Michel and Chave [29].

All calibration curves were made at three pressures (1×10^5 Pa, 1.38×10^7 Pa, and 2.76×10^7 Pa). Where shown, error bars represent the double-sided 95% confidence limits for the extreme value parameters defined in [29].

Solutions were made using $\text{MgCl}_2 \cdot 6\text{H}_2\text{O}$, NaCl, KCl, $\text{MnSO}_4 \cdot \text{H}_2\text{O}$, and $\text{CaCl}_2 \cdot \text{H}_2\text{O}$ dissolved in DI water. All concentrations are given in parts per million (ppm, wt./vol.). For the optimization studies the concentrations used were 5000 ppm Mg, 1000 ppm K, 100 ppm Na, 1000 ppm Ca, and 1000 ppm Mn.

5.4 Results and Discussion

For double pulse LIBS of high pressure aqueous solutions, optimizing the key parameters (E_1 , E_2 , t_d , and ΔT) individually for each element of interest is essential for identifying the conditions under which each can be detected. The optimization studies presented here show that these conditions are pressure dependent for double pulse LIBS. Outside the range of these conditions, some of the elements prove to be undetectable. Through the optimization of the parameters, a set of conditions are established that allow calibration curves to be made for determining the limits of detection for high pressure bulk aqueous solutions using the current LIBS system setup.

5.4.1 Magnesium

Results for the 518.4 nm Mg (I) peak are presented in Figures 5-3 to 5-5. From these figures, the need for a short gate delay (≤ 100 ns) is evident, irrespective of the pressure or the laser pulse energies. At all pressures, when $t_d = 500$ ns, Mg (I) is not detectable. At 1×10^5 Pa, the intensity was greatest with a longer interpulse delay (1000 ns to 5000 ns) (Figure 5-3). It was also best to use a smaller E_1 than E_2 . For example, two favorable conditions were $E_1 = 60$ mJ, $E_2 = 100$ mJ and $E_1 = 20$ mJ, $E_2 = 100$ mJ. At 2.76×10^7 Pa, the greatest intensity peak exists when ΔT had the smallest value (50 ns) (Figure 5-5). The E_1 and E_2 that gave the greatest intensity were both in the range of 60 mJ - 140 mJ. At 1.38×10^7 Pa, the optimal ΔT was intermediate between that for 1×10^5 Pa and 2.76×10^7 Pa. At 2.76×10^7 Pa, a very intense peak was obtained for the condition $E_1 = 60$ mJ, $E_2 = 60$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns. Therefore, a calibration curve was constructed using these conditions as shown in Figure 5-6. From this calibration curve and from examining spectra of the calibration data (Figure 5-7), it is clear that Mg (I) can be detected to only 5000 ppm.




Figure 5-3: Mg (I) (518.4 nm peak) optimization at 1×10^5 Pa




Figure 5-4: Mg (I) (518.4 nm peak) optimization at 1.38×10^7 Pa




Figure 5-5: Mg (I) (518.4 nm peak) optimization at 2.76×10^7 Pa




Figure 5-6: Mg (I) calibration curve of the 518.4 nm peak, $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa, ($E_1 = 60$ mJ, $E_2 = 60$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns)




Figure 5-7: Spectra of Mg (I) peak (518.4 nm) at 2.76×10^7 Pa. The concentrations from bottom to top are 1000 ppm and 5000 ppm. ($E_1 = 60$ mJ, $E_2 = 60$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns). For clarity, the spectra have been offset from each other by 500 a.u.

5.4.2 Potassium

The results of the optimization studies for the 769.9 nm K (I) peak are shown in Figures 5-8 - 5-10. Potassium was detectable over an unusually wide range of conditions. Potassium has the lowest ionization energy (4.31 eV) of the elements that were studied which contributes to ease of detection. In particular, a wide range of inter-pulse delay times was suitable at all pressures. At 1×10^5 Pa and 1.38×10^7 Pa, the use of two high energy pulses resulted in the greatest peak intensity. At the highest pressure condition (2.76×10^7 Pa), a lower energy pulse followed by a higher energy pulse was advantageous. For the calibration curve, the condition $E_1 = 100$ mJ, $E_2 = 140$ mJ, $t_d = 1000$ ns, and $\Delta T = 50$ ns was selected (Figure 5-11). The calibration curve reveals that K (I) is detectable to 500 ppm. Spectra of the calibration data confirm this and are shown in Figure (5-12).




Figure 5-8: K (I) (769.9 nm peak) optimization at 1×10^5 Pa




Figure 5-9: K (I) (769.9 nm peak) optimization at 1.38×10^7 Pa




Figure 5-10: K (I) (769.9 nm peak) optimization at 2.76×10^7 Pa




Figure 5-11: K (I) calibration curve of the 769.9 nm peak, $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa, ($E_1 = 100$ mJ, $E_2 = 140$ mJ, $t_d = 1000$ ns, and $\Delta T = 50$ ns)




Figure 5-12: Spectra of 769.9 nm K (I) peak at 2.76×10^7 Pa. Concentrations of spectra from bottom to top are 100 ppm, 500 ppm, 1000 ppm. ($E_1 = 100$ mJ, $E_2 = 140$ mJ, $t_d = 1000$ ns, and $\Delta T = 50$ ns). For clarity, the spectra have been offset from each other by 200 a.u.

5.4.3 Calcium

The results for the 422 nm Ca (I) peak are shown in Figures 5-13 to 5-15. At 1×10^5 Pa, the greatest intensity peak was detected when a long interpulse delay time was used (500 to 5000 ns) with a low energy pulse (typically 20 mJ) followed by a higher energy pulse (60 mJ - 140 mJ). At both 1.38×10^7 Pa and 2.76×10^7 Pa, the greatest intensity peak was detected when two high energy pulses (typically 100 mJ - 140 mJ) were separated by 50 ns. As a result, the condition selected for the calibration curve was $E_1 = 100$ mJ, $E_2 = 100$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns. The calcium calibration curves for both the 393 nm Ca (II) and 422 nm Ca (I) peaks are shown in Figure 5-16. The limit of detection for calcium using the conditions selected is 500 ppm. Spectra at these conditions are shown in Figure 5-17.




Figure 5-13: Ca (I) (422 nm peak) optimization at 1×10^5 Pa




Figure 5-14: Ca (I) (422 nm peak) optimization at 1.38×10^7 Pa






Figure 5-15: Ca (I) (422 nm peak) optimization at 2.76×10^7 Pa



(a) 393 nm Ca (II) peak



(b) 422 nm Ca (I) peak

Figure 5-16: Calcium calibration curves, $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa, ($E_1 = 100$ mJ, $E_2 = 100$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns)




Figure 5-17: Spectra of calcium (393 nm Ca (II), 396 nm Ca (II), and 422 nm Ca (I) peaks) at 2.76×10^7 Pa. Concentrations of spectra from bottom to top are 100 ppm, 500 ppm, and 1000 ppm. ($E_1 = 100$ mJ, $E_2 = 100$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns). For clarity, the spectra have been offset from each other by 15,000 a.u.

5.4.4 Manganese

The optimization study results for the 403 nm Mn (I) peak are shown in Figures 5-18 to 5-20. Although a Mn (I) triplet exists at 403 nm, peak broadening in liquids causes it to be unresolvable, and hence a single 403 nm peak is used for this study. At 1×10^5 Pa, a low first energy pulse (20 mJ) followed by a higher second energy pulse (60 mJ - 140 mJ) with a long interpulse delay time (5000 ns) gave the highest peak intensity (Figure 5-18). When the pressure was increased, the need for a significantly shorter interpulse delay time was evident (Figures 5-18 - 5-19). At the highest pressure condition, two high energy pulses gave the most intense peak (Figure 5-20). Using a low second energy pulse (20 mJ) was not beneficial at this pressure condition. At the intermediate pressure, 1.38×10^7 Pa, several parameter combinations can be used. Either a low energy first pulse (20 mJ) followed by a higher pulse (60 mJ - 140 mJ) or two higher energy pulses (60 mJ - 140 mJ) were suitable characteristics of both the low pressure and high pressure conditions are workable and hence this is possibly a transition pressure. For Mn (I), the conditions that were selected for the calibration curve were $E_1 = 100$ mJ, $E_2 = 60$ mJ, $t_d = 50$ ns, $\Delta T = 50$ ns, with results shown in Figure 5-21. Spectra at these conditions are shown in Figure 5-22. From both the calibration curve (Figure 5-21) and the subsequent spectra (Figure 5-22), the limit of detection is 1000 ppm.




Figure 5-18: Mn (I) (403 nm peak) optimization at 1×10^5 Pa




Figure 5-19: Mn (I) (403 nm peak) optimization at 1.38×10^7 Pa




Figure 5-20: Mn (I) (403 nm peak) optimization at 2.76×10^7 Pa




Figure 5-21: Mn (I) (403 nm peak) calibration curve. $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa, ($E_1 = 100$ mJ, $E_2 = 60$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns)




Figure 5-22: Spectra of Mn (I) peak (403 nm) at 2.76×10^7 Pa. Concentrations from bottom to top are 100 ppm, 500 ppm, and 1000 ppm. ($E_1 = 100$ mJ, $E_2 = 60$ mJ, $t_d = 50$ ns, and $\Delta T = 50$ ns). For clarity, the spectra have been offset from each other by 1000 a.u.

5.4.5 Sodium

The 588.995 nm Na (I) peak from the sodium doublet was used for optimization studies (Figures 5-23 - 5-25). The highest intensity at 1×10^5 Pa was recorded when a low energy pulse was followed by a high energy pulse and a long interpulse delay time was used (Figure 5-23). For example, the greatest intensity was recorded for $E_1 = 20$ mJ, $E_2 = 140$ mJ, and $\Delta T = 5000$ ns. At 1.38×10^7 Pa, the greatest intensity peaks were recorded when two high energy pulses (60 mJ to 140 mJ) were fired in rapid succession in either order and separated by 50 ns (Figure 5-24). Since the interpulse delay time is very small, these conditions are close to that for single pulse operation with a very high energy pulse. At 2.76×10^7 Pa, the greatest intensity peak again exists when two high energy pulses (60 mJ and 140 mJ) are rapidly fired in either order separated by 50 ns (Figure 5-25). The lowest intensity peaks were recorded at all pressures when the second energy pulse was 20 mJ, suggesting that the second pulse must be of sufficient energy or irradiance to excite or re-excite plasma emission. The sodium calibration curve was therefore made at $E_1 = 60$ mJ, $E_2 = 140$ mJ, $\Delta T = 50$ ns, and $t_d = 50$ ns and is shown in Figure 5-26. Spectra made at the high pressure (2.76×10^7 Pa) condition which clearly indicate the limit of detection of 50 ppm are shown in Figure 5-27.




Figure 5-23: Na (I) (588.995 nm) optimization at 1×10^5 Pa




Figure 5-24: Na (I) optimization at 1.38×10^7 Pa




Figure 5-25: Na (I) (588.995 nm) optimization at 2.76×10^7 Pa





Figure 5-26: Na (I) calibration curves of the (a) 588.995 nm peak and the (b) 589.6 nm peak. $\bigcirc = 1 \times 10^5$ Pa, $\square = 1.38 \times 10^7$ Pa, $\triangle = 2.76 \times 10^7$ Pa.




Figure 5-27: Spectra of the Na (I) doublet peaks (588.995 nm and 589.6 nm) at 2.76×10^7 Pa. Concentrations from bottom to top are 10 ppm, 50 ppm, 100 ppm, 500 ppm, and 1000 ppm. ($E_1 = 60$ mJ, $E_2 = 140$ mJ, $\Delta T = 50$ ns, and $t_d = 50$ ns). For clarity, the spectra have been offset from each other by 2000 a.u., except for the 1000 ppm spectrum which has been offset from the 500 ppm spectrum by 8000 a.u.

The limits of detection for sodium as well as the other analytes reported here are much higher than expected based on earlier work [12, 14]. In an effort to determine if the low throughput ($f/10$) Echelle spectrometer is the limiting factor an additional calibration curve was completed at 1×10^5 Pa at an optimal atmospheric pressure condition. Lawrence-Snyder *et al.* report with the use of a Chromex Czerny-Turner spectrometer coupled to an ICCD camera that for Na at 3.4×10^6 Pa using $E_1 = 7$ mJ, $E_2 = 48$ mJ, and $t_d = 1 \mu\text{s}$, the optimal interpulse delay time is between approximately 20 and 50 μs [16]. In our work, first, we selected a low energy pulse (20 mJ) followed by a high energy pulse (140 mJ) and the optimal interpulse delay time between them was determined. This interpulse delay was then used to create a calibration curve. In the optimization study for high pressure, the interpulse timing was only carried out to 5 μs . The first peak intensity was measured at 1 μs , then at 5 μs , then at 5 μs increments to a maximum of 170 μs . Figure 5-28 details the effect of interpulse timing on the intensity on the 588.995 nm Na (I) peak. The intensity is fairly uniform from 10 μs until 140 μs . After 140 μs , the intensity drops off. From this plot, it appears valid to select an interpulse delay time between 1 - 140 μs ; therefore, we select 70 μs to construct a calibration curve. Comparing this time scale to that shown in Lawrence-Snyder *et al.* [16], this seems consistent due to our use of a significantly higher second energy pulse. Using the 70 μs interpulse delay time, a calibration curve was constructed and is shown in Figure 5-29. Although we have now used the optimal condition for 1×10^5 Pa with a long interpulse delay time, the detection limit is again only 50 ppm. This can also be verified by looking at the spectra (Figure 5-30). This suggests that the Echelle spectrometer is the limiting detection factor.




Figure 5-28: Effect of interpulse delay on intensity on the 588 nm Na peak at 1×10^5 Pa. (1000 ppm)






Figure 5-29: Na calibration curve at 1×10^5 Pa made using $E_1 = 20$ mJ, $E_2 = 140$ mJ, and $t_d = 50$ ns. □ = 588 nm, ○ = 589 nm



(a) Concentrations from bottom to top: 50 ppm, 100 ppm, 500 ppm, and 1000 ppm. For clarity, the spectra have been offset by 5000 a.u.



(b) Concentrations from bottom to top: 50 ppm, 100 ppm. For clarity, the spectra have been offset by 500 a.u.

Figure 5-30: Na (I) spectra at 1×10^5 Pa. ($E_1 = 20$ mJ, $E_2 = 140$ mJ, and $t_d = 50$ ns)

5.5 Conclusions

Double pulse LIBS was used to detect five analytes in bulk aqueous solutions important in hydrothermal vent chemistry at pressures up to 2.76×10^7 Pa. The key double pulse parameters were optimized for each of the elements at three pressures (1×10^5 Pa, 1.38×10^7 Pa, and 2.76×10^7 Pa). The parameters needed for detection were found to be both element and pressure dependent. The use of the optimal parameters is essential because outside of this set of parameters, the elements may not be detectable. Potassium and sodium were detectable over a wide range of conditions. In general, for all elements, as pressure was increased, the use of a shorter interpulse delay was necessary and at 2.76×10^7 Pa, an interpulse delay time on the order of 50 ns should be used. For all conditions studied, a short gate delay (usually ≤ 100 ns) was required. This is similar to the results from single pulse LIBS (Chapter 4). The need for a short gate delay suggests that in bulk liquids, the plasma lifetime is short, possibly lasting only on the order of 500 ns.

Using each of the optimally established conditions, calibration curves were made at three pressures (1×10^5 Pa, 1.38×10^7 Pa, and 2.76×10^7 Pa). From these, the limits of detection for the five analytes were found to be 5000 ppm Mg, 500 ppm K, 500 ppm Ca, 1000 ppm Mn, and 50 ppm Na using the current system set-up. The limits of detection were the same for all three pressures tested. The main reason that LIBS researchers choose to use double pulse LIBS instead of single pulse LIBS is to achieve improved sensitivity and improved signal. When optimization of conditions was completed using single pulse LIBS and then subsequently calibration curves were made using the same system set-up for the present work for Mn, Ca, and Na (Chapter 4), the resulting limits of detection were found to be 500 ppm Mn, 50 ppm Ca, and 50 ppm Na. This suggests that the use of double pulse LIBS in high pressure aqueous solutions may not be advantageous. Lawrence-Snyder *et al.* reported no emission enhancements using double pulse LIBS above 1×10^7 Pa [16]. DeGiacomo *et al.* emphasize the need for selecting an appropriate interpulse delay time for underwater LIBS [23]. DeGiacomo *et al.* stress the need to find the interpulse delay time that allows for the second pulse to form a plasma in a bubble that is maximally expanded. The high pressure environment of our experiments may cause the bubble to collapse too rapidly. As a result, the highest intensity peaks that are observed occur when two pulses are close together, similar to a single pulse. Lawrence-Snyder *et al.* [30] suggest that at higher solution pressures (8×10^7 Pa), the bubble formed by the first laser pulse is confined by its surrounding pressure. As a result, the bubble never

expands to the maximum volume that is observed at lower pressures. Therefore, the emission enhancements are not observed.

Although the use of double pulse LIBS proved less favorable than expected, it should be noted that one major contributing factor was the spectrometer used in these studies. The Echelle spectrometer has a very high resolution but a very low light throughput and poor sensitivity, with an f number of 10. In an effort to maximize the light throughput, it would be advisable to use a spectrometer with a smaller f number. As an example, the use of a spectrometer with an f number of 2 could improve the throughput by a factor of 25 and thus improve the limits of detection. Furthermore, the use of a PMT as the detector, may further improve detection limits for bulk liquids as demonstrated by Cremers *et al.* who measured at atmospheric pressure Na I (589.00 nm) at 0.014 ppm, K I (766.49 nm) at 1.2 ppm, Mg II (279.55 nm) at 100 ppm, and Ca II (393.37 nm) at 0.8 ppm [12]. Further work is needed to maximize the light collection by changing system components for bulk aqueous solution experiments.

5.6 Acknowledgments

We acknowledge the National Science Foundation for support of this research under grant OCE-0527927. Additional support was received from the Deep Ocean Exploration Institute and the Ocean Ventures Fund of the Woods Hole Oceanographic Institution.

Bibliography

- [1] R. S. Harmon, F. C. DeLucia, C. E. McManus, N. J. McMillan, T. F. Jenkins, M. E. Walsh, and A. Miziolek. Laser-induced breakdown spectroscopy - an emerging chemical sensor technology for real-time field-portable, geochemical, mineralogical, and environmental applications. *Applied Geochemistry*, 21(5):730–747, May 2006.
- [2] Z. A. Arp, D. A. Cremers, R. C. Wiens, D. M. Wayne, B. Sallé, and S. Maurice. Analysis of water ice and water ice/soil mixtures using laser-induced breakdown spectroscopy: Application to Mars polar exploration. *Applied Spectroscopy*, 58:897–909, 2004.
- [3] Z. A. Arp, D. A. Cremers, R. D. Harris, D. M. Oschwald, G. R. Parker Jr., and D. M. Wayne. Feasibility of generating a useful laser-induced breakdown spectroscopy plasma on rocks at high pressure: preliminary study for a Venus mission. *Spectrochimica Acta Part B*, 59:987–999, 2004.
- [4] G. B. Courrèges-Lacoste, B. Ahlers, and F. R. Pérez. Combined Raman spectrometer/laser-induced breakdown spectrometer for the next ESA mission to Mars. *Spectrochimica Acta Part A*, In Press, 2007.
- [5] R. Brennetot, J. L. Lacour, E. Vors, A. Rivoallan, D. Vailhen, and S. Maurice. Mars analysis by laser-induced breakdown spectroscopy (MALIS): Influence of Mars atmosphere on plasma emission and study of factors influencing plasma emission with the use of Doehlert designs. *Applied Spectroscopy*, 57(7):744–752, 2003.
- [6] A. Knight, N. Scherbarth, D. Cremers, and M. Ferris. Characterization of laser-induced breakdown spectroscopy (LIBS) for application to space exploration. *Applied Spectroscopy*, 54:331–340, 2000.
- [7] B. Sallé, J.-L. Lacour, P. Mauchien, P. Fichet, S. Maurice, and G. Manhes. Comparative study of different methodologies for quantitative rock analysis by laser-induced breakdown spectroscopy in a simulated Martian atmosphere. *Spectrochimica Acta Part B*, 61:301–313, 2006.
- [8] C. R. German and K. L. Von Damm. *Treatise on Geochemistry*, chapter Hydrothermal Processes, pages 181–222. Elsevier, 2003.
- [9] J. H. Trefry, D. B. Butterfield, S. Metz, G. J. Massoth, R. P. Trocine, and R. A. Feely. Trace metals in hydrothermal solutions from Cleft segment on the southern Juan de Fuca Ridge. *Journal of Geophysical Research*, 99:4925–4935, 1994.
- [10] K. L. Von Damm. Controls on the chemistry and temporal variability of seafloor hydrothermal fluids. *Seafloor hydrothermal systems: physical, chemical, biological, and geological interactions: Geophysical Monograph 91*, pages 222–247, 1995.

- [11] K. L. Von Damm. Chemistry of hydrothermal vent fluids from 9° - 10° N, East Pacific Rise: 'Time zero,' The immediate post-eruptive period. *Journal of Geophysical Research*, 105:11203–11222, 2000.
- [12] D. A. Cremers, L. J. Radziemski, and T. R. Loree. Spectrochemical analysis of liquids using the laser spark. *Applied Spectroscopy*, 38:721–729, 1984.
- [13] R. Knopp, F. J. Scherbaum, and J. I. Kim. Laser induced breakdown spectroscopy (LIBS) as an analytical tool for the detection of metal ions in aqueous solutions. *Fresenius' Journal of Analytical Chemistry*, 355:16–20, 1996.
- [14] W. Pearman, J. Scaffidi, and S. M. Angel. Dual-pulse laser-induced breakdown spectroscopy in bulk aqueous solution with an orthogonal beam geometry. *Applied Optics*, 42:6085–6093, 2003.
- [15] A. P. M. Michel, M. Lawrence-Snyder, S. M. Angel, and A. D. Chave. Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: evaluation of key measurement parameters. *Applied Optics*, 46, 2007.
- [16] M. Lawrence-Snyder, J. Scaffidi, S. M. Angel, A. P.M. Michel, and A. D. Chave. Sequential-pulse laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, 61:171–176, 2007.
- [17] M. Lawrence-Snyder, J. Scaffidi, S. M. Angel, A. P. M. Michel, and A. D. Chave. Laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, 60:786–790, 2006.
- [18] A. De Giacomo, M. Dell'Aglia, F. Colao, R. Fantoni, and V. Lazic. Double-pulse LIBS in bulk water and on submerged bronze samples. *Applied Surface Science*, 247:157–162, 2005.
- [19] A. E. Pichahchy, D. A. Cremers, and M. J. Ferris. Elemental analysis of metals under water using laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 52:25–39, 1997.
- [20] C. Haisch, J. Liermann, U. Panne, and R. Niessner. Characterization of colloidal particles by laser-induced plasma spectroscopy (LIPS). *Analytica Chimica Acta*, 346:23–25, 1997.
- [21] P. K. Kennedy, D. X. Hammer, and B. A. Rockwell. Laser-induced breakdown in aqueous media. *Progress in Quantum Electronics*, 21:155–248, 1997.
- [22] A. De Giacomo, M. Dell'Aglia, and O. De Pascale. Single pulse-laser induced breakdown spectroscopy in aqueous solution. *Applied Physics A: Materials Science & Processing*, V79(4):1035–1038, September 2004.
- [23] A. De Giacomo, M. Dell'Aglia, O. De Pascale, and M. Capitelli. From single pulse to double pulse ns-laser induced breakdown spectroscopy under water: elemental analysis of aqueous solutions and submerged solid samples. *Spectrochimica Acta Part B*, in press, 2007.

- [24] A. Casavola, A. De Giacomo, M. Dell’Aglio, F. Taccogna, G. Colonna, O. De Pascale, and S. Longo. Experimental investigation and modelling of double pulse laser induced plasma spectroscopy under water. *Spectrochimica Acta Part B*, 2005.
- [25] A. De Giacomo, M. Dell’Aglio, F. Colao, and R. Fantoni. Double pulse laser produced plasma on metallic target in seawater: basic aspects and analytical approach. *Spectrochimica Acta B*, 59:1431–1438, 2004.
- [26] S. Koch, R. Court, W. Garen, W. Neu, and R. Reuter. Detection of manganese in solution in cavitation bubbles using laser induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 60(7-8):1230–1235, 2005.
- [27] V. Lazic, F. Colao, R. Fantoni, and V. Spizzicchino. Laser-induced breakdown spectroscopy in water: Improvement of the detection threshold by signal processing. *Spectrochimica Acta Part B*, 60:1002–1013, 2005.
- [28] R. Noll. Terms and notations for laser-induced breakdown spectroscopy. *Analytical and Bioanalytical Chemistry*, V385(2):214–218, 2006.
- [29] A. P. M. Michel and A. D. Chave. Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics. *Spectrochimica Acta Part B*, In Press.
- [30] M. Lawrence-Snyder, J. P. Scaffidi, W. F. Pearman, and S. M. Angel. Dependence of emission intensity on bubble dynamics in dual-pulse laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, submitted.

Chapter 6

Preliminary investigations on matrix effects of Na, K, and Ca for bulk liquids at oceanic pressures

6.1 Abstract

Chemical matrix effects occur when one element present in a sample affects the emission of another element also present in the sample. In this study, the effect of NaCl on the detection of K and Ca in bulk aqueous solutions at pressures up to 2.76×10^7 Pa is explored. In addition, the effect of the background matrix (chloride versus sulfate) on the detection of Na and K is examined. While the investigations into the effect of NaCl on K and Ca proved inconclusive, the background matrix has no effect on the ability to detect Na or K.

6.2 Introduction

When multiple elements are present in a sample, chemical matrix effects can occur in which the presence of one element affects the emission of another element. Changing the concentration of one element can affect the signal intensity of one or more other elements even when the concentrations of those elements are not altered. This can adversely affect the ability to make quantitative measurements and lead to complications in calibration. However, if matrix effects can be understood, then the concentration effect can be quantified [1].

In a plasma, the addition of an easily ionizable element can shift the neutral-ion

equilibrium concentration of an analyte toward the neutral species (Le Chatelier's principle). For example, addition of an easily ionizable species to a plasma can reduce the emission of an ionized species. The easily ionized species increases the electron density which decreases the concentration of the ionized species [1]. Cremers *et al.* [2] reported a matrix effect in bulk liquids which showed that the intensity ratio of Ca II/Ca I decreased with the addition of NaCl. A similar effect was reported by Michel *et al.* [3] (Chapter 2) using double pulse LIBS over a range of pressures. Michel *et al.* also reported no effect on K or Mn emission intensity with the addition of NaCl.

Several other matrix effects have been reported. For example, Charfi and Harith [4] showed that when Mg and Na were measured in pure solution versus in mixed solutions (both Mg and Na present), the limits of detection were lower for the pure solutions. Eppler *et al.* [5] investigated matrix effects of Ba and Pb in sand and soil, and saw that Ba (II) and Pb (I) emission was dependent on analyte speciation. As samples were varied from a pure sand to a pure soil composition, the Ba (II)/C (I) signal decreased. The decrease correlated with an increase in electron density, possibly due to a change in ionized species.

Matrix effects can be corrected by applying calibration curves for each element contained in the substrate of interest, analogous to estimating the partial derivatives for a multivariate process, as opposed to that for a single one as in most calibration curves. This requires many reference samples and is very time consuming. In addition, it is not feasible for *in situ* measurements with unknown, mixed component samples [6]. Several research groups have developed methods to compensate for matrix effects. St-Onge *et al.* have used an internal standardization method to reduce matrix effects [7]. Barrette and Turmel showed that matrix effects could be partially overcome by the use of a multivariable calibration curve [8]. Ciucci *et al.* developed an algorithm-based procedure for calibration-free quantitative elemental analysis of materials [9].

Understanding matrix effects is essential for quantitative LIBS. In this preliminary investigation, matrix effects of three elements (Na, K, and Ca), are explored in bulk aqueous solutions at pressures up to 2.76×10^7 Pa. In addition, the effect of a background matrix (chloride versus sulfate) on the detection of Na and K is examined.

6.3 Experimental

The laboratory set-up for simulating the use of LIBS in the deep ocean has been previously described in Chapter 4. For all experiments described here, single pulse LIBS experiments were run with the laser pulse energy kept constant at 40 mJ (which

corresponds to an irradiance of the beam at the beam waist of $\approx 5.25 \times 10^{12} \text{ W/cm}^2$ for this system). For all experiments, $t_d = 50 \text{ ns}$, and $t_b = 200 \text{ ns}$ and the Echelle spectrometer amplification was set to the maximum value of 4000. All spectra taken were composed of 100 accumulated shots. All raw spectral data were processed using generalized extreme value distribution statistics detailed in a paper by Michel and Chave [10] after binning the data over 20 wavelengths. Where shown, error bars represent 95% confidence intervals of the GEVD.

Solutions were made from NaCl, KCl, $\text{NaSO}_4 \cdot \text{H}_2\text{O}$, $\text{MnSO}_4 \cdot \text{H}_2\text{O}$, and $\text{CaCl}_2 \cdot 2\text{H}_2\text{O}$ dissolved in de-ionized water. All concentrations are given in parts per million (ppm, wt./vol.). All experiments were completed at three pressures, $1 \times 10^5 \text{ Pa}$, $1.38 \times 10^7 \text{ Pa}$, and $2.76 \times 10^7 \text{ Pa}$.

6.4 Results and Discussion


6.4.1 Matrix Effects of K and Na

Effect of Sodium Chloride on the Emission from Potassium


Four different conditions (Table 6.1) were used to study the effect of NaCl on the emission intensity of the 769.9 nm K peak. 120 spectra were taken at each of Conditions 1 and 3 to establish a low and high calibration point for K alone. Three replicate measurements each consisting of 60 spectra were taken at Conditions 2 and 4 to assess the effect of NaCl on K. Figure 6-1 shows the results. When NaCl was added to low concentration (102 ppm) K, no peak intensity change was seen at all three pressures. In contrast, in high concentration (1022 ppm) K, when NaCl was added, the K peak intensity decreased at all three pressures. This is in contrast to prior work [3] (Chapter 2) where no effect of NaCl was seen on the peak intensity of K.

Table 6.1: Conditions used to study the K-Na matrix effect


Condition	K	Cl	Na
1	102 ppm	93 ppm	0 ppm
2	102 ppm	4959 ppm	3155 ppm
3	1022 ppm	927 ppm	0 ppm
4	1022 ppm	4959 ppm	2614 ppm



(a) 1×10^5 Pa



(b) 1.38×10^7 Pa



(c) 2.76×10^7 Pa


Figure 6-1: Effect of presence of Na on peak intensity of K (769.9 nm) for two concentrations of K. \bigcirc = K, \square = K + NaCl

Effect of Potassium Chloride on Emission from Sodium


Four different conditions (Table 6.2) were used to study the effect of KCl on the detection of the 588 nm Na peak. 120 spectra were taken of Na at Conditions 1 and 3 to create low and high concentration calibration points for Na alone. Three replicate measurements each consisting of 60 spectra were taken at Conditions 2 and 4 to assess the effect of KCl on Na. The results are shown in Figure 6-2. The high variability from low to high pressure without any discernible trend suggests that more laboratory work is needed to determine if a matrix effect exists.

Table 6.2: Conditions used to study the Na-K matrix effect


Condition	K	Cl	Na
1	0 ppm	4865 ppm	3155 ppm
2	102 ppm	4959 ppm	3155 ppm
3	0 ppm	4032 ppm	2614 ppm
4	1022 ppm	4959 ppm	2614 ppm



(a) 1×10^5 Pa



(b) 1.38×10^7 Pa



(c) 2.76×10^7 Pa

Figure 6-2: Effect of presence of K on peak intensity of Na (588 nm). □ = Na, ○ = Na + 102 ppm K, △ = Na + 1022 ppm K


6.4.2 Matrix Effects of Ca and Na

Effect of Sodium Chloride on Emission from Calcium


Four different conditions (Table 6.3) were used study the effect of NaCl on the emission intensity of Ca [393 nm Ca (II) and 422 nm Ca (I)]. 120 spectra were taken at Conditions 1 and 3 to establish low and high concentration calibration points for Ca alone. Three replicate measurements each consisting of 60 spectra were taken at Conditions 2 and 4 to assess the effect of NaCl on the 393 nm Ca (II) ionic peak and the 422 nm (I) atomic peak. Figure 6-3 shows the effect of NaCl on the 393 nm Ca (II) ionic peak. For 106 ppm Ca, the peak intensity was not measurable at any pressure, both in the presence and absence of NaCl. When NaCl is added to a 1055 ppm Ca solution, the emission intensity is significantly lower. It should be noted that at 1×10^5 Pa and 1.38×10^7 Pa, the measured intensity under all conditions is relatively low. Figure 6-4 shows the results for the effect of NaCl on the 422 nm Ca (I) atomic peak. The results are inconclusive. To assess possible drift in the spectrometer between the pure Ca data and the Ca plus NaCl data, the ratio of the intensity of the 393 nm peak to the 422 nm peak is plotted in Figure 6-5. The result suggests that the variability is not due to simple drift, but to another cause of systematic error.

Table 6.3: Conditions used to study the Ca-Na Matrix Effect


Condition	Ca	Cl	Na
1	106 ppm	187 ppm	0 ppm
2	106 ppm	4959 ppm	3094 ppm
3	1055 ppm	1867 ppm	0 ppm
4	1055 ppm	4959 ppm	2005 ppm



(a) 1×10^5 Pa




(b) 1.38×10^7 Pa




(c) 2.76×10^7 Pa


Figure 6-3: Effect of presence of Na on peak intensity of Ca (II) (393 nm) for two concentrations of Ca. \bigcirc = Ca, \square = Ca + NaCl



(a) 1×10^5 Pa




(b) 1.38×10^7 Pa




(c) 2.76×10^7 Pa


Figure 6-4: Effect of presence of Na on peak intensity on Ca (I) (422 nm) for two concentrations of Ca. \bigcirc = Ca, \square = Ca + NaCl



(a) 1×10^5 Pa



(b) 1.38×10^7 Pa



(c) 2.76×10^7 Pa


Figure 6-5: Ratio of 393 nm Ca (II) peak to 422 nm Ca (I) peak \bigcirc = Ca, \square = Ca + NaCl

Effect of Calcium Chloride on the Emission of Sodium


A sodium calibration curve was made with 120 spectra taken at three concentrations (Conditions 1 - 3 shown in Table 6.4). Three replicates of 60 spectra were taken at two concentrations of Ca plus NaCl. The data shown in Figure 6-6 reveal high variability in the LIBS system measurements. In general, when Ca was present, the Na peak intensity was reduced. However, the systematic error denoted by the scatter in the data suggests that more data collection is needed to confirm the result.

Table 6.4: Conditions used to study the effect of $\text{CaCl}_2 \cdot 2\text{H}_2\text{O}$ on emission of Na


Condition	Ca	Cl	Na
1	0 ppm	3091 ppm	2005 ppm
2	0 ppm	4032 ppm	2614 ppm
3	0 ppm	4865 ppm	3155 ppm
4	1055 ppm	4959 ppm	2005 ppm
5	106 ppm	4959 ppm	3094 ppm



(a) 1×10^5 Pa



(b) 1.38×10^7 Pa



(c) 2.76×10^7 Pa

Figure 6-6: Effect of presence of Ca on peak intensity on Na (588 nm), \square = Na, \triangle = Na + 1055 ppm Ca, \bigcirc = 106 ppm Ca + Na

6.4.3 Detection of Na and K in a Chloride Versus Sulfate Matrix

Two experiments were carried out to determine if the presence of a chloride or sulfate substrate affects the detection of an element. The intensity of potassium (1002 ppm K) in two matrices, KCl and K₂SO₄, and the intensity of sodium (2356 ppm Na), in two matrices, NaCl and NaSO₄, were measured. Figures 6-7 and 6-8 contrast the ability to detect K and Na, respectively, in a chloride matrix versus a sulfate matrix. For K, no matrix effect is seen at any of the pressures (1×10^5 Pa, 1.38×10^7 Pa, 2.76×10^7 Pa) tested. For Na (Figure 6-8), there is a possible matrix effect, although additional experiments are necessary to determine this due to scatter in the data. In the development of an oceanic LIBS sensor operable in the vent environment, the sensors must be capable of detecting both K and Na in the presence of both sulfate and chloride, which these data support.






Figure 6-7: Detection of 769 nm potassium (1002 ppm) in a chloride versus sulfate matrix. ○ = KCl, △ = K₂SO₄



(a) 588 nm Na Peak



(b) 589 nm Na Peak

Figure 6-8: Detection of sodium (2356 ppm Na) in a chloride versus sulfate matrix.
 ○ = NaCl, △ = Na₂SO₄

6.5 Conclusions

Matrix effects of Na, K, and Ca were examined in a preliminary study. Although the main goal was determining how the presence of NaCl affects the peak intensities of K and Ca, the data showed inconsistencies and were therefore inconclusive. In addition, the results were not consistent with previous preliminary findings reported by Cremers *et al.* [1] and Michel *et al.* [3]. The data suggest that matrix effects are smaller than the systematic variability of the measurements, and suggest that further laboratory experimentation is needed to elucidate if significant matrix effects exist.

A comparative study to investigate the ability to detect analytes in a chloride versus a sulfate matrix showed no significant difference. The results indicate that Na and K are both detectable in chloride and sulfate matrices. The K peak intensities were not influenced by the substrate present (sulfate and chloride) at pressures up to 2.76×10^7 Pa. The Na peak intensities were possibly influenced by the background matrix (sulfate and chloride); yet, more experiments are needed to determine this conclusively.

6.6 Acknowledgments

Support for this research was provided by the National Science Foundation under grant OCE-0527927. Additional support was received from the Deep Ocean Exploration Institute and the Ocean Ventures Fund of the Woods Hole Oceanographic Institution.

Bibliography

- [1] D. A. Cremers and L. J. Radziemski. *Laser-Induced Breakdown Spectroscopy (LIBS): Fundamentals and Applications*, chapter History and Fundamentals of LIBS, pages 1–39. Cambridge University Press, 2006.
- [2] D. A. Cremers, L. J. Radziemski, and T. R. Loree. Spectrochemical analysis of liquids using the laser spark. *Applied Spectroscopy*, 38:721–729, 1984.
- [3] A. P. M. Michel, M. Lawrence-Snyder, S. M. Angel, and A. D. Chave. Laser-induced breakdown spectroscopy of bulk aqueous solutions at oceanic pressures: evaluation of key measurement parameters. *Applied Optics*, 46, 2007.
- [4] B. Charfi and M. A. Harith. Panoramic laser-induced breakdown spectrometry of water. *Spectrochimica Acta Part B*, 57:1141–1153, 2002.
- [5] A. S. Eppler, D. A. Cremers, D. D. Hickmott, M. J. Ferris, and A. C. Koskelo. Matrix effects in the detection of Pb and Ba in soils using laser-induced breakdown spectroscopy. *Applied Spectroscopy*, 50:1175–1181, 1996.
- [6] B. Sallé, J.-L. Lacour, P. Mauchien, P. Fichet, S. Maurice, and G. Manhes. Comparative study of different methodologies for quantitative rock analysis by laser-induced breakdown spectroscopy in a simulated Martian atmosphere. *Spectrochimica Acta Part B*, 61:301–313, 2006.
- [7] L. St-Onge, E. Kwong, M. Sabsabi, and E. B. Vadas. Quantitative analysis of pharmaceutical products by laser-induced breakdown spectroscopy. *Spectrochimica Acta Part B*, 57(7):1131–1140, 2002.
- [8] L. Barrette and S. Turmel. On-line iron-ore slurry monitoring for real-time process control of pellet making processes using laser-induced breakdown spectroscopy: graphitic vs. total carbon detection. *Spectrochimica Acta Part B*, 56(6):715–723, 2001.
- [9] A. Ciucci, M. Corsi, V. Palleschi, S. Rastelli, A. Salvetti, and E. Tognoni. New procedure for quantitative elemental analysis by laser-induced plasma spectroscopy. *Applied Spectroscopy*, 53:960–964, 1999.
- [10] A. P. M. Michel and A. D. Chave. Analysis of laser-induced breakdown spectroscopy (LIBS) spectra: The case for extreme value statistics. *Spectrochimica Acta Part B*, In Press.

Chapter 7

Conclusions and Future Directions

7.1 Conclusions

In this thesis, laser-induced breakdown spectroscopy was evaluated for its potential for development into a new *in situ* chemical sensor for the deep ocean. The elements of focus for this work were selected for their importance in hydrothermal vent chemistry. Although it is potentially and theoretically possible to use LIBS to detect all elements, it was found during this laboratory investigation that it is more difficult to detect elements in bulk liquids at the concentrations desired than was anticipated. However, one important finding was that if an element could be detected in a bulk liquid at atmospheric pressure, then it could also be detected in the same liquid at high pressures. This thesis shows that the LIBS technique can be used successfully to detect dissolved Li, Na, K, Mg, Ca, and Mn at pressures up to 2.76×10^7 Pa, although not at the levels observed naturally in seawater or vent fluids for all of these elements. The work in this thesis was completed using two different LIBS systems, one with a Czerny-Turner spectrometer and one with an Echelle spectrometer, and with two different pressure cells.

7.1.1 Development of a New Data Processing Scheme

A new data processing scheme for LIBS spectra was developed in this thesis and was applied to all data, except that presented in Chapter 2. In an examination of the variability of peak intensities for both single shot and ensemble-averaged LIBS spectra, LIBS data were found to have a dramatically non-normal statistical distribution. The distribution of the peak intensities was instead found to follow the generalized extreme value distribution. A preliminary investigation into the sources of the vari-

ability was carried out. Laser pulse fluctuations, while identified as a contributing source of variability, were ruled out as the primary source. Plasma images revealed large spatial and intensity differences on a shot-to-shot basis. This analysis led to the development of a data processing scheme that accurately deals with the extreme nature of laser-induced plasma formation and should be used for statistically accurate comparisons of LIBS spectra instead of simply averaging spectra, the standard method used by LIBS researchers. This scheme was found to be applicable for both solid and liquid samples and for use with data taken with both a Czerny-Turner and an Echelle spectrometer.

7.1.2 Single Pulse LIBS

Results Using a Czerny-Turner Spectrometer

Preliminary investigations were carried out using a Czerny-Turner spectrometer that showed that Li, Ca, Mn, K, and Na were detectable in high pressure bulk aqueous solutions. This work focused on the energy levels needed for the detection of these analytes and revealed that an optimal range of low laser pulse energies exists for their detection in both low and high pressure solutions. It was hypothesized that a low energy pulse could create a smaller, more tightly focused plasma that forms only at the focal spot. However, for a high energy pulse, the high energy density may cause breakdown even before the pulse reaches the focal spot causing breakdown to occur over a longer distance. The effect of pressure on emission intensity was investigated for Ca, Na, and Mn with no pressure effect seen for Ca and Na, yet, with an increase in intensity with increased pressure observed for Mn.

Results Using an Echelle Spectrometer

Using an Echelle spectrometer, Na, Mn, and Ca were again shown to be detectable with little effect of pressure on the spectra. Irrespective of the laser pulse energy selected or the solution pressure, the need for a short (less than 200 ns) gate delay was found for the detection of analytes. This study confirmed that a low energy pulse (less than ≈ 60 mJ) is optimal. Calibration curves showed that Na, Mn, and Ca are detectable at 50 ppm, 500 ppm, and 50 ppm, respectively. However, the limits of detection were higher than expected which was attributed to both the poor sensitivity and the low light throughput of the system used.

7.1.3 Double Pulse LIBS

Results Using a Czerny-Turner Spectrometer

In this preliminary work, double pulse LIBS was used to detect four analytes (Ca, Li, Na, and Mn) and the energy levels for maximum emission were found to vary by analyte. Analyte detection was found to be highly dependent on the interpulse delay. If the interpulse delay is short ($\ll 1 \mu\text{s}$), signal intensity is greatly enhanced when compared to that measured using longer delay times. Two pulses separated by a short ΔT approaches single pulse conditions. Such a small interpulse delay may not be sufficient for a cavitation bubble to fully form before the second laser pulse creates a spark. This was an early suggestion that dual pulse LIBS might not be advantageous at elevated pressure.

Results Using an Echelle Spectrometer

Double pulse LIBS was used to detect five analytes (Mg, K, Mn, Na, and Ca) at pressures up to 2.76×10^7 Pa. The key double pulse parameters were optimized for each of the elements at three pressures (1×10^5 Pa, 1.38×10^7 Pa, and 2.76×10^7 Pa). The parameters needed for detection were found to be both element and pressure dependent. The use of the optimal parameters is essential because outside this set of parameters, some of the elements were not detectable, although K and Na were detectable over a wide range of conditions. In general, for all elements, as pressure was increased, the use of a shorter interpulse delay was necessary and at 2.76×10^7 Pa an interpulse delay time on the order of 50 ns was ideal. For all conditions studied, a short gate delay (usually ≤ 100 ns) was required, suggesting that in bulk liquids, the plasma lifetime is short, possibly lasting only on the order of 500 ns.

Calibration curves were made that established the limits of detection for the five analytes: 5000 ppm Mg, 500 ppm K, 500 ppm Ca, 1000 ppm Mn, and 50 ppm Na using the current system set-up. Using single pulse LIBS, limits of detection were found to be 500 ppm Mn, 50 ppm Ca, and 50 ppm Na which suggests that the use of double pulse LIBS in high pressure aqueous solutions may not be advantageous. The high pressure environment may cause the bubble to collapse too rapidly and as a result the bubble does not expand to the maximum volume observed at lower pressures [1]. As a result, the emission enhancements seen at atmospheric pressure are not observed in the high pressure environment.

Although the use of double pulse LIBS proved less favorable than expected, it should be noted that one major contributing factor was the spectrometer used in

these studies. The Echelle spectrometer has a very high resolution but a very low light throughput and poor sensitivity, with an f number of 10. Further work is needed to maximize the light collection by changing system components for bulk aqueous solution experiments.

7.1.4 Matrix

Results Using a Czerny-Turner Spectrometer

The addition of NaCl was found to enhance the emission intensity for Ca.

Results Using an Echelle Spectrometer

The matrix effect of Na, K, and Ca was examined to determine how the presence of NaCl affects the peak intensity of K and Ca, although the data showed inconsistencies and were therefore inconclusive. Na and K were found to be detectable in both chloride and sulfate matrices. For Na the peak intensities were possibly influenced and for K the peak intensities were not influenced by the background matrix (sulfate versus chloride) at pressures up to 2.76×10^7 Pa.

7.2 Future Work

7.2.1 Laboratory Work

Significantly more laboratory work is needed before a robust ocean-going system can be built and used successfully in the field. Although this thesis showed that many elements could be detected in a high pressure aqueous environment, and several were detectable at levels found at or below the concentrations present in vent fluids, there are many issues that need to be addressed to improve upon the current LIBS system.

Due to the attenuation of light in water, 532 nm or green laser light will travel substantially further than 1064 nm laser light in the ocean. Therefore, it will be advantageous to develop a LIBS system that uses 532 nm laser pulses as the excitation source. The disadvantage to using 532 nm laser pulses as the excitation source is the need for a filter before the collection fiber optic. Since 532 nm is in the range that can be detected by the spectrometer, it is important to never collect the 532 nm laser pulse light. To prevent capturing the laser light, a 532 nm notch filter is added before the collection fiber optic to filter out the light from the laser pulses. The addition of this filter will minimally reduce the light available to the spectrometer. LIBS of

liquids using 532 nm should be feasible; therefore, the next phase of laboratory work should explore this option.

Improvements to the limits of detection could be made by using a spectrometer with a smaller f number than the Echelle spectrometer that was used for much of this thesis work. Although an Echelle spectrometer has the advantage of being able to simultaneously detect multiple elements due to its broad wavelength coverage, its high resolution yields low throughput. The Echelle spectrometer also had a small slit width which limited light input. In addition, the use of a PMT instead of an ICCD as the detector could be beneficial for increasing sensitivity. Changes therefore to the detection system may allow for the elements presented in this thesis to be detected at much lower limits of detection and for additional elements to be detected. For example, the detection of Cu, Fe, Cl, and Br was attempted as part of this research without success. It is recommended that a new spectrometer with a smaller f number be selected for continued work on this project. The Echelle spectrometer used in this work has an f number of 10 and it is strongly recommended to use a spectrometer with an f number of 4 or 2, which would have greater throughput.

Future work should include significant studies on the variability of plasma formation. This can include imaging of both laser-induced plasmas and bubbles. Plasma dynamics studies should be carried out in an effort to find ways to minimize shot-to-shot plasma variability. Although this thesis included some plasma images taken orthogonal from the incoming laser beam, plasma images could also be taken from multiple angles to gain a multi-dimensional understanding of plasma formation. Then by looking at what factors influence plasma dynamics (e.g., the shape of the plasma, the length of emission, emission intensity), more control over the plasma variability may be obtainable. The factors that influence the plasma dynamics include the laser (e.g., beam quality, wavelength), focusing optics, and sample pressure and should also be studied for their impact on variability.

Calibration curves were used in this thesis to determine the limits of detection. In future work, much work should be completed on the calibration curves to both decrease the variability of the curves and to verify repeatability of the curves. The calibration curves should also be extended to higher concentrations to verify that the curves are linear. In addition, if a new spectrometer is used, calibration curves can again be used to determine limits of detection. Once the variability issues are resolved and the repeatability of calibration curves are established, more work can focus on matrix effects. Matrix effect data showed inconsistencies that are thought to arise from the variability issues and from the spectrometer used. Selecting a new

spectrometer may improve these studies; however, the reduction of plasma variability again should first be addressed.

There are many additional challenges to studying liquids in the ocean environment that were not addressed in this thesis. For example, the effect of particles in the water on the detection of analytes should be addressed. In addition, the effect of a flowing solution on LIBS spectra and especially the shot-to-shot variability in a flowing system should be examined. Furthermore, laboratory work should emphasize elements not yet targeted. In theory, all elements are detectable using LIBS; thus, significant work is needed to determine all of the elements which are detectable in high pressure aqueous environments.

7.2.2 Design of an Ocean-Going Sensor

Although there are several issues that need to be addressed in the laboratory to deal with the variability issues and to improve upon the calibration curves and limits of detection, this thesis shows that the use of LIBS in a high pressure aqueous environment is feasible. Once these issues have been addressed, a LIBS sensor can be designed for use in the ocean. The natural extension of this research therefore is the development of the first ever sea-going LIBS system. This will entail the design of an instrument that can be used at hydrothermal vents aboard an underwater vehicle (Figure 7-1). The major components of an underwater LIBS system are a spectrometer, source laser(s), fiber optic links and probes for these units, and data acquisition/control electronics. All of these components will need to be packaged in underwater pressure housings and provided with power and data communications connections. Several critical issues related to integrating the system onto an underwater vehicle also need to be addressed, including vehicle payload, power resources, and system control. Software modifications for laser control and data collection will be a necessary component of this work. The design of an optical probe head is a critical element of a sea-going LIBS system. The fiber optic design must pass the laser pulses into the fiber, image the fiber spot onto the target, and image the plasma emission onto a fiber bundle that couples to the spectrometer. The fiber bundle must be designed to withstand both the pressure effects of the ocean and also the high energy laser pulses. In the ocean environment, there are other types of samples that ocean scientists will be interested in using an *in situ* chemical LIBS sensor for making measurements of in the ocean. For example, LIBS could be used for analyzing sediments, rocks, underwater archaeological artifacts, and the air-sea interface. The

successful design and testing of an ocean-going system will pave the way for its use by oceanographers for numerous applications.



Figure 7-1: Future oceanic LIBS system. Illustration by E. Paul Oberlander, WHOI.

Bibliography

- [1] M. Lawrence-Snyder, J. P. Scaffidi, W. F. Pearman, and S. M. Angel. Dependence of emission intensity on bubble dynamics in dual-pulse laser-induced breakdown spectroscopy of high-pressure bulk aqueous solutions. *Applied Spectroscopy*, submitted.

REPORT DOCUMENTATION PAGE	1. REPORT NO. MIT/WHOI 2007-19	2.	3. Recipient's Accession No.
4. Title and Subtitle Laboratory Evaluation of Laser-Induced Breakdown Spectroscopy (LIBS) as a New <i>in situ</i> Chemical Sensing Technique for the Deep Ocean		5. Report Date September 2007	
7. Author(s) Anna Pauline Miranda Michel		6.	
9. Performing Organization Name and Address MIT/WHOI Joint Program in Oceanography/Applied Ocean Science & Engineering		8. Performing Organization Rept. No.	
		10. Project/Task/Work Unit No. MIT/WHOI 2007-19	
		11. Contract(C) or Grant(G) No. (C) OCE0352278 (G) OCE0352242	
12. Sponsoring Organization Name and Address National Science Foundation WHOI Deep Ocean Exploration Institute WHOI Ocean Ventures Fund National Defense Science & Engineering Graduate Fellowship		13. Type of Report & Period Covered Ph.D. Thesis	
15. Supplementary Notes This thesis should be cited as: Anna Pauline Miranda Michel, 2007. Laboratory Evaluation of Laser-Induced Breakdown Spectroscopy (LIBS) as a New <i>in situ</i> Chemical Sensing Technique for the Deep Ocean. Ph.D. Thesis. MIT/WHOI, 2007-19.		14.	
16. Abstract (Limit: 200 words) Laser-induced breakdown spectroscopy (LIBS) possesses many of the characteristics required for <i>in situ</i> chemical sensing, and is a promising technique for field measurements in extreme environments. In this work, laboratory experiments validate the LIBS technique in a simulated deep ocean environment to pressures up to 2.76×10^7 Pa. A key focus of this work is the validation that select elements important for understanding hydrothermal vent fluid chemistry (Na, Ca, Mn, Mg, K, and Li) are detectable using LIBS. A data processing scheme that accurately deals with the extreme nature of laser-induced plasma formation was developed that allows for statistically accurate comparisons of spectra. The use of both single and double pulse LIBS for high pressure bulk aqueous solutions is explored and the system parameters needed for the detection of the key analytes are optimized. Using both single and double pulse LIBS, the limits of detection were found to be higher than expected as a result of the spectrometer used in this experimentation. However, the results of this validation show that LIBS possesses the characteristics to be a viable chemical sensing method for <i>in situ</i> analyte detection in high pressure environments like the deep ocean.			
17. Document Analysis			
a. Descriptors Laser-induced breakdown Spectroscopy chemical sensing hydrothermal vents			
b. Identifiers/Open-Ended Terms			
c. COSATI Field/Group			
18. Availability Statement Approved for publication; distribution unlimited.		19. Security Class (This Report) UNCLASSIFIED	21. No. of Pages 173
		20. Security Class (This Page)	22. Price